

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121720\
Data File : PP032222.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Dec 2020 12:32
Operator : DD\AJ
Sample : L5119-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
B-W

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 18 03:52:43 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121120.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 12 05:15:55 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.778	4.049	1201846	1066293	23.331	22.545
2) SA Decachlor...	10.640	9.339	773733	836651	24.673	24.948

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121720\
Data File : PP032222.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Dec 2020 12:32
Operator : DD\AJ
Sample : L5119-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
B-W

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 18 03:52:43 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121120.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 12 05:15:55 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.778 min
Delta R.T.: -0.004 min
Response: 1201846
Conc: 23.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
B-W




#1 Tetrachloro-m-xylene

R.T.: 4.049 min
Delta R.T.: -0.005 min
Response: 1066293
Conc: 22.54 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.640 min
Delta R.T.: -0.007 min
Response: 773733
Conc: 24.67 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.339 min
Delta R.T.: -0.009 min
Response: 836651
Conc: 24.95 ng/ml