

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121823\
 Data File : PP062478.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Dec 2023 12:46
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 18 13:00:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 18 13:00:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.402	3.690	53094626	89889416	25.795	25.395
2) SA Decachlor...	10.159	8.801	35638383	63066333	24.573	25.845
Target Compounds						
16) L4 AR-1242-1	5.576	4.800	13220002	22960599	267.748	267.018
17) L4 AR-1242-2	5.598	4.819	19432776	32202326	264.455	264.769
18) L4 AR-1242-3	5.661	4.999	12209810	17902919	269.570	268.354
19) L4 AR-1242-4	5.759	5.084	10117289	18758034	265.397	275.435
20) L4 AR-1242-5	6.500	5.616	10425059	21960842	262.672	272.287



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121823\
Data File : PP062478.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Dec 2023 12:46
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 18 13:00:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 18 13:00:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.402 min
Delta R.T.: 0.000 min
Response: 53094626
Conc: 25.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.000 min
Response: 89889416
Conc: 25.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.159 min
Delta R.T.: 0.000 min
Response: 35638383
Conc: 24.57 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.801 min
Delta R.T.: 0.000 min
Response: 63066333
Conc: 25.84 ng/ml



#16 AR-1242-1


R.T.: 5.576 min
Delta R.T.: 0.000 min
Response: 13220002
Conc: 267.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.800 min
Delta R.T.: 0.000 min
Response: 22960599
Conc: 267.02 ng/ml




#17 AR-1242-2

R.T.: 5.598 min
Delta R.T.: 0.000 min
Response: 19432776
Conc: 264.45 ng/ml



#17 AR-1242-2


R.T.: 4.819 min
Delta R.T.: 0.000 min
Response: 32202326
Conc: 264.77 ng/ml



#18 AR-1242-3


R.T.: 5.661 min
Delta R.T.: 0.000 min
Response: 12209810
Conc: 269.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 4.999 min
Delta R.T.: 0.000 min
Response: 17902919
Conc: 268.35 ng/ml




#19 AR-1242-4

R.T.: 5.759 min
Delta R.T.: 0.000 min
Response: 10117289
Conc: 265.40 ng/ml



#19 AR-1242-4


R.T.: 5.084 min
Delta R.T.: 0.000 min
Response: 18758034
Conc: 275.44 ng/ml



#20 AR-1242-5

R.T.: 6.500 min
Delta R.T.: 0.000 min
Response: 10425059
Conc: 262.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.616 min
Delta R.T.: 0.000 min
Response: 21960842
Conc: 272.29 ng/ml