

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121918\
 Data File : PP028527.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Dec 2018 21:18
 Operator : EI\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 11:49:56 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP120718.M
 Quant Title : 8080.M
 QLast Update : Fri Dec 07 14:01:32 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.094	12.207	185.5E6	330.9E6	710.603	738.769
Target Compounds							
1) T	Dalapon	3.342	3.012	370.2E6	363.8E6	633.377	658.237
2) T	3,5-DICHL...	10.507	10.299	324.5E6	371.7E6	652.082	686.644
3) T	4-Nitroph...	11.672	11.352	275.8E6	159.5E6	633.516	712.747
5) T	DICAMBA	12.416	12.518	1069.1E6	1207.8E6	658.041	691.640
6) T	MCPD	12.758	12.710	32145647	65964572	70.754	74.518
7) T	MCPA	13.001	13.098	50646638	99515635	69.280	73.640
8) T	DICHLORPROP	13.613	13.632	237.6E6	299.2E6	495.562	691.165 #
9) T	2,4-D	13.965	14.120	252.7E6	354.1E6	676.637	697.385
10) T	Pentachlo...	14.382	14.732	5432.3E6	4667.8E6	671.254	698.945
11) T	2,4,5-TP ...	15.246	15.302	1956.7E6	1787.5E6	666.186	699.271
12) T	2,4,5-T	15.638	15.944	1771.8E6	1490.2E6	665.993	685.775
13) T	2,4-DB	16.402	16.545	132.8E6	177.8E6	626.179	667.010
14) T	DINOSEB	17.888	16.937	1425.3E6	1221.6E6	671.301	696.941
15) T	Picloram	17.657	18.320	2472.7E6	2279.8E6	598.192	650.844
16) T	DCPA	18.252	18.198	2896.4E6	2326.0E6	674.779	710.704

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121918\
Data File : PP028527.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Dec 2018 21:18
Operator : EI\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
Client Sampled :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 11:49:56 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP120718.M
Quant Title : 8080.M
QLast Update : Fri Dec 07 14:01:32 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

