

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121922\
 Data File : PP053851.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Dec 2022 21:24
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 03:30:10 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 14 05:38:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.339	3.591	105.3E6	75482023	50.130	48.332
2) SA Decachlor...	10.099	8.626	80842764	64865827	47.617	46.440
Target Compounds						
26) L6 AR-1254-1	6.369	5.488	37662554	33364799	486.016	509.516
27) L6 AR-1254-2	6.588	5.636	57635519	30103788	495.377	516.186
28) L6 AR-1254-3	6.956	6.041	59082487	47052515	504.429	534.858
29) L6 AR-1254-4	7.243	6.271	42365486	25948693	534.390	566.380
30) L6 AR-1254-5	7.663	6.691	49618647	43369217	508.779	556.488


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP121922\
Data File : PP053851.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2022 21:24
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 03:30:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 14 05:38:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.339 min
Delta R.T.: 0.000 min
Response: 105262028
Conc: 50.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.591 min
Delta R.T.: -0.003 min
Response: 75482023
Conc: 48.33 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.099 min
Delta R.T.: -0.003 min
Response: 80842764
Conc: 47.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.626 min
Delta R.T.: -0.010 min
Response: 64865827
Conc: 46.44 ng/ml



#26 AR-1254-1


R.T.: 6.369 min
Delta R.T.: 0.000 min
Response: 37662554
Conc: 486.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.488 min
Delta R.T.: -0.004 min
Response: 33364799
Conc: 509.52 ng/ml




#27 AR-1254-2

R.T.: 6.588 min
Delta R.T.: -0.001 min
Response: 57635519
Conc: 495.38 ng/ml



#27 AR-1254-2


R.T.: 5.636 min
Delta R.T.: -0.005 min
Response: 30103788
Conc: 516.19 ng/ml



#28 AR-1254-3


R.T.: 6.956 min
Delta R.T.: -0.002 min
Response: 59082487
Conc: 504.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.041 min
Delta R.T.: -0.005 min
Response: 47052515
Conc: 534.86 ng/ml




#29 AR-1254-4

R.T.: 7.243 min
Delta R.T.: 0.000 min
Response: 42365486
Conc: 534.39 ng/ml



#29 AR-1254-4


R.T.: 6.271 min
Delta R.T.: -0.005 min
Response: 25948693
Conc: 566.38 ng/ml



#30 AR-1254-5

R.T.: 7.663 min
Delta R.T.: 0.000 min
Response: 49618647
Conc: 508.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.691 min
Delta R.T.: -0.005 min
Response: 43369217
Conc: 556.49 ng/ml