

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122121\
 Data File : PP042126.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Dec 2021 23:37
 Operator : AJ\MA
 Sample : M5144-10 10X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 22 01:39:44 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 16 06:21:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.925	4.065	28267	52046	2.191	2.122
2)	SA Decachlor...	10.921	9.423	29139	40210	2.211	2.000
Target Compounds							
28)	L6 AR-1254-3	7.755	0.000	4663	0	5.969	N.D. #
32)	L7 AR-1260-2	0.000	7.111f	0	9375	N.D.	6.172 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122121\
Data File : PP042126.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Dec 2021 23:37
Operator : AJ\MA
Sample : M5144-10 10X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 01:39:44 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121621.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 16 06:21:48 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.925 min
Delta R.T.: -0.002 min
Response: 28267
Conc: 2.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.065 min
Delta R.T.: -0.004 min
Response: 52046
Conc: 2.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.921 min
Delta R.T.: 0.006 min
Response: 29139
Conc: 2.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.423 min
Delta R.T.: -0.006 min
Response: 40210
Conc: 2.00 ng/ml



#28 AR-1254-3


R.T.: 7.755 min
Delta R.T.: 0.000 min
Response: 4663
Conc: 5.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.803 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 8.190 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 7.111 min
Delta R.T.: -0.025 min
Response: 9375
Conc: 6.17 ng/ml