

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122123\
 Data File : PP062566.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Dec 2023 22:43
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 22 02:22:28 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 22 02:21:38 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.395	3.682	166.9E6	225.1E6	73.385	72.938
2) SA Decachlor...	10.142	8.782	109.4E6	186.1E6	71.927	71.228
Target Compounds						
26) L6 AR-1254-1	6.425	5.606	60686060	113.4E6	720.795	712.776
27) L6 AR-1254-2	6.645	5.756	91858964	98675765	714.213	708.220
28) L6 AR-1254-3	7.012	6.166	96053409	159.4E6	720.452	715.610
29) L6 AR-1254-4	7.298	6.397	66542869	92721618	717.448	710.919
30) L6 AR-1254-5	7.719	6.819	73401236	139.5E6	721.905	712.099



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122123\
Data File : PP062566.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Dec 2023 22:43
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 02:22:28 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 22 02:21:38 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.395 min
Delta R.T.: 0.000 min
Response: 166921612
Conc: 73.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.682 min
Delta R.T.: 0.000 min
Response: 225082984
Conc: 72.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.142 min
Delta R.T.: 0.000 min
Response: 109393154
Conc: 71.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.782 min
Delta R.T.: 0.000 min
Response: 186117016
Conc: 71.23 ng/ml



#26 AR-1254-1


R.T.: 6.425 min
Delta R.T.: 0.001 min
Response: 60686060
Conc: 720.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#26 AR-1254-1

R.T.: 5.606 min
Delta R.T.: 0.000 min
Response: 113367628
Conc: 712.78 ng/ml




#27 AR-1254-2

R.T.: 6.645 min
Delta R.T.: 0.001 min
Response: 91858964
Conc: 714.21 ng/ml



#27 AR-1254-2


R.T.: 5.756 min
Delta R.T.: 0.000 min
Response: 98675765
Conc: 708.22 ng/ml



#28 AR-1254-3


R.T.: 7.012 min
Delta R.T.: 0.001 min
Response: 96053409
Conc: 720.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#28 AR-1254-3

R.T.: 6.166 min
Delta R.T.: 0.000 min
Response: 159444968
Conc: 715.61 ng/ml




#29 AR-1254-4

R.T.: 7.298 min
Delta R.T.: 0.000 min
Response: 66542869
Conc: 717.45 ng/ml



#29 AR-1254-4


R.T.: 6.397 min
Delta R.T.: 0.000 min
Response: 92721618
Conc: 710.92 ng/ml



#30 AR-1254-5

R.T.: 7.719 min
Delta R.T.: 0.001 min
Response: 73401236
Conc: 721.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750



#30 AR-1254-5

R.T.: 6.819 min
Delta R.T.: 0.000 min
Response: 139497993
Conc: 712.10 ng/ml