

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122123\
 Data File : PP062577.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Dec 2023 01:41
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP122123AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 22 03:31:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 22 03:29:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.395	3.682	115.9E6	157.8E6	53.291	52.886
2) SA Decachlor...	10.142	8.782	74720908	129.1E6	51.817	50.726
Target Compounds						
16) L4 AR-1242-1	5.567	4.789	29245369	38863758	521.802	505.762
17) L4 AR-1242-2	5.589	4.809	43204289	55385220	523.358	514.097
18) L4 AR-1242-3	5.651	4.988	27117697	30597739	529.842	519.369
19) L4 AR-1242-4	5.750	5.073	21693546	31627147	518.007	513.553
20) L4 AR-1242-5	6.489	5.605	23338095	37224225	544.074	515.318



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122123\
Data File : PP062577.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Dec 2023 01:41
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP122123AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 03:31:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 22 03:29:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.395 min
Delta R.T.: -0.001 min
Response: 115859192
Conc: 53.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP122123AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.682 min
Delta R.T.: 0.000 min
Response: 157823865
Conc: 52.89 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.142 min
Delta R.T.: 0.000 min
Response: 74720908
Conc: 51.82 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.782 min
Delta R.T.: 0.000 min
Response: 129134790
Conc: 50.73 ng/ml



#16 AR-1242-1


R.T.: 5.567 min
Delta R.T.: 0.000 min
Response: 29245369
Conc: 521.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP122123AR1242




#16 AR-1242-1

R.T.: 4.789 min
Delta R.T.: 0.000 min
Response: 38863758
Conc: 505.76 ng/ml




#17 AR-1242-2

R.T.: 5.589 min
Delta R.T.: 0.000 min
Response: 43204289
Conc: 523.36 ng/ml



#17 AR-1242-2


R.T.: 4.809 min
Delta R.T.: 0.000 min
Response: 55385220
Conc: 514.10 ng/ml



#18 AR-1242-3


R.T.: 5.651 min
Delta R.T.: 0.000 min
Response: 27117697
Conc: 529.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP122123AR1242




#18 AR-1242-3

R.T.: 4.988 min
Delta R.T.: 0.000 min
Response: 30597739
Conc: 519.37 ng/ml




#19 AR-1242-4

R.T.: 5.750 min
Delta R.T.: 0.000 min
Response: 21693546
Conc: 518.01 ng/ml



#19 AR-1242-4


R.T.: 5.073 min
Delta R.T.: 0.000 min
Response: 31627147
Conc: 513.55 ng/ml



#20 AR-1242-5

R.T.: 6.489 min
Delta R.T.: 0.000 min
Response: 23338095
Conc: 544.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP122123AR1242



#20 AR-1242-5

R.T.: 5.605 min
Delta R.T.: 0.000 min
Response: 37224225
Conc: 515.32 ng/ml