

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122222\
 Data File : PP053963.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Dec 2022 18:07
 Operator : YP\AJ
 Sample : N6183-03
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 22 20:44:59 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 14 05:38:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.338	3.589	43152389	30598570	20.551	19.593
2)	SA Decachlor...	10.098	8.621	25494807	21665282	15.017	15.511
Target Compounds							
3)	L1 AR-1016-1	5.508	4.712f	91536	161205	1.350	3.510 #
4)	L1 AR-1016-2	5.532	4.712	185812	161205	1.865	2.573 #
5)	L1 AR-1016-3	5.584	0.000	315777	0	4.996	N.D. #
6)	L1 AR-1016-4	5.686	0.000	43837	0	0.885	N.D. #
7)	L1 AR-1016-5	5.994	5.153	62644	116141	1.190	3.203 #
8)	L2 AR-1221-1	4.546	0.000	1240659	0	48.159	N.D. #
9)	L2 AR-1221-2	4.645	3.888	855679	764212	43.951	52.281
10)	L2 AR-1221-3	4.692f	0.000	104868	0	1.804	N.D. #
11)	L3 AR-1232-1	4.692f	0.000	104868	0	2.341	N.D. #
12)	L3 AR-1232-2	5.241	4.712	31434	161205	1.222	5.401 #
13)	L3 AR-1232-3	5.532	0.000	185812	0	3.959	N.D. #
14)	L3 AR-1232-4	5.686	0.000	43837	0	1.850	N.D. #
15)	L3 AR-1232-5	5.792	5.153f	51971	116141	2.818	7.053 #
16)	L4 AR-1242-1	5.508	4.712f	91536	161205	1.748	4.552 #
17)	L4 AR-1242-2	5.532	4.712	185812	161205	2.439	3.355 #
18)	L4 AR-1242-3	5.584	0.000	315777	0	6.445	N.D. #
19)	L4 AR-1242-4	5.686	0.000	43837	0	1.142	N.D. #
20)	L4 AR-1242-5	6.430	5.491	239451	154387	5.871	5.058
21)	L5 AR-1248-1	5.508	4.712f	91536	161205	2.172	5.667 #
22)	L5 AR-1248-2	5.792	0.000	51971	0	0.822	N.D. #
23)	L5 AR-1248-3	5.994	0.000	62644	0	0.894	N.D. #
24)	L5 AR-1248-4	6.396	5.153	82703	116141	1.173	2.414 #
25)	L5 AR-1248-5	6.430	5.513f	239451	200904	3.398	5.350 #
26)	L6 AR-1254-1	6.379	5.491	211827	154387	2.734	2.358
27)	L6 AR-1254-2	6.584	0.000	100216	0	0.861	N.D. #
28)	L6 AR-1254-3	6.960	6.058	661652	398932	5.649	4.535
29)	L6 AR-1254-4	7.246	0.000	95066	0	1.199	N.D. #
30)	L6 AR-1254-5	7.683f	6.688	1463458	766219	15.006	9.832 #
31)	L7 AR-1260-1	7.122	6.170	284886	329335	2.950	5.049 #
32)	L7 AR-1260-2	7.371	6.358	1259751	290944	11.403	3.864 #
33)	L7 AR-1260-3	7.742	6.514	374267	271241	4.950	3.742
34)	L7 AR-1260-4	7.948f	6.987	3199305	306117	34.088	5.642 #
35)	L7 AR-1260-5	8.282	7.229	850189	703759	5.362	5.885
36)	L8 AR-1262-1	7.742	6.781	374267	122168	3.127	3.425
37)	L8 AR-1262-2	8.282	6.987	850189	306117	4.434	3.938
38)	L8 AR-1262-3	8.602	7.515	835687	371177	6.031	5.891
39)	L8 AR-1262-4	8.687	7.579	1007899	506058	16.195	4.426 #
40)	L8 AR-1262-5	9.336	8.073	462227	377276	6.541	7.573
41)	L9 AR-1268-1	8.602	7.515	835687	371177	3.424	2.052 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122222\
Data File : PP053963.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Dec 2022 18:07
Operator : YP\AJ
Sample : N6183-03
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 20:44:59 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 14 05:38:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.687	7.579	1007899	506058	4.435	2.966 #
43)	L9 AR-1268-3	8.916	7.789	329665	153296	1.720	1.038 #
44)	L9 AR-1268-4	9.336	8.073	462227	377276	6.005	6.770
45)	L9 AR-1268-5	9.759	8.366	714131	453149	1.214	0.959



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122222\
Data File : PP053963.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Dec 2022 18:07
Operator : YP\AJ
Sample : N6183-03
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 20:44:59 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 14 05:38:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.338 min
Delta R.T.: -0.001 min
Response: 43152389
Conc: 20.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.589 min
Delta R.T.: -0.004 min
Response: 30598570
Conc: 19.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.098 min
Delta R.T.: -0.004 min
Response: 25494807
Conc: 15.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.621 min
Delta R.T.: -0.015 min
Response: 21665282
Conc: 15.51 ng/ml



#3 AR-1016-1


R.T.: 5.508 min
Delta R.T.: -0.004 min
Response: 91536
Conc: 1.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.712 min
Delta R.T.: 0.026 min
Response: 161205
Conc: 3.51 ng/ml




#4 AR-1016-2

R.T.: 5.532 min
Delta R.T.: -0.002 min
Response: 185812
Conc: 1.87 ng/ml



#4 AR-1016-2


R.T.: 4.712 min
Delta R.T.: 0.007 min
Response: 161205
Conc: 2.57 ng/ml



#5 AR-1016-3


R.T.: 5.584 min
Delta R.T.: -0.012 min
Response: 315777
Conc: 5.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.882 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.686 min
Delta R.T.: -0.009 min
Response: 43837
Conc: 0.88 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 4.924 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 5.994 min
Delta R.T.: 0.003 min
Response: 62644
Conc: 1.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.153 min
Delta R.T.: 0.015 min
Response: 116141
Conc: 3.20 ng/ml




#8 AR-1221-1

R.T.: 4.546 min
Delta R.T.: -0.003 min
Response: 1240659
Conc: 48.16 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.810 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.645 min
Delta R.T.: 0.010 min
Response: 855679
Conc: 43.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.888 min
Delta R.T.: -0.007 min
Response: 764212
Conc: 52.28 ng/ml




#10 AR-1221-3

R.T.: 4.692 min
Delta R.T.: -0.020 min
Response: 104868
Conc: 1.80 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 3.972 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.692 min
Delta R.T.: -0.019 min
Response: 104868
Conc: 2.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.971 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.241 min
Delta R.T.: 0.000 min
Response: 31434
Conc: 1.22 ng/ml



#12 AR-1232-2


R.T.: 4.712 min
Delta R.T.: 0.007 min
Response: 161205
Conc: 5.40 ng/ml



#13 AR-1232-3


R.T.: 5.532 min
Delta R.T.: -0.002 min
Response: 185812
Conc: 3.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.882 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.686 min
Delta R.T.: -0.009 min
Response: 43837
Conc: 1.85 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T. : 4.966 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.792 min
Delta R.T.: 0.006 min
Response: 51971
Conc: 2.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.153 min
Delta R.T.: 0.015 min
Response: 116141
Conc: 7.05 ng/ml




#16 AR-1242-1

R.T.: 5.508 min
Delta R.T.: -0.003 min
Response: 91536
Conc: 1.75 ng/ml



#16 AR-1242-1


R.T.: 4.712 min
Delta R.T.: 0.026 min
Response: 161205
Conc: 4.55 ng/ml



#17 AR-1242-2


R.T.: 5.532 min
Delta R.T.: -0.002 min
Response: 185812
Conc: 2.44 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.712 min
Delta R.T.: 0.007 min
Response: 161205
Conc: 3.36 ng/ml




#18 AR-1242-3

R.T.: 5.584 min
Delta R.T.: -0.013 min
Response: 315777
Conc: 6.45 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T.: 4.882 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.686 min
Delta R.T.: -0.009 min
Response: 43837
Conc: 1.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 4.966 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 6.430 min
Delta R.T.: -0.006 min
Response: 239451
Conc: 5.87 ng/ml



#20 AR-1242-5


R.T.: 5.491 min
Delta R.T.: 0.000 min
Response: 154387
Conc: 5.06 ng/ml



#21 AR-1248-1


R.T.: 5.508 min
Delta R.T.: -0.004 min
Response: 91536
Conc: 2.17 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.712 min
Delta R.T.: 0.026 min
Response: 161205
Conc: 5.67 ng/ml




#22 AR-1248-2

R.T.: 5.792 min
Delta R.T.: 0.006 min
Response: 51971
Conc: 0.82 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T.: 4.924 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 5.994 min
Delta R.T.: 0.003 min
Response: 62644
Conc: 0.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 4.966 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.396 min
Delta R.T.: 0.000 min
Response: 82703
Conc: 1.17 ng/ml



#24 AR-1248-4


R.T.: 5.153 min
Delta R.T.: 0.015 min
Response: 116141
Conc: 2.41 ng/ml



#25 AR-1248-5


R.T.: 6.430 min
Delta R.T.: -0.006 min
Response: 239451
Conc: 3.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.513 min
Delta R.T.: -0.021 min
Response: 200904
Conc: 5.35 ng/ml




#26 AR-1254-1

R.T.: 6.379 min
Delta R.T.: 0.009 min
Response: 211827
Conc: 2.73 ng/ml



#26 AR-1254-1


R.T.: 5.491 min
Delta R.T.: -0.001 min
Response: 154387
Conc: 2.36 ng/ml



#27 AR-1254-2


R.T.: 6.584 min
Delta R.T.: -0.005 min
Response: 100216
Conc: 0.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 5.640 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 6.960 min
Delta R.T.: 0.002 min
Response: 661652
Conc: 5.65 ng/ml



#28 AR-1254-3


R.T.: 6.058 min
Delta R.T.: 0.012 min
Response: 398932
Conc: 4.53 ng/ml



#29 AR-1254-4


R.T.: 7.246 min
Delta R.T.: 0.002 min
Response: 95066
Conc: 1.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.276 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 7.683 min
Delta R.T.: 0.019 min
Response: 1463458
Conc: 15.01 ng/ml



#30 AR-1254-5


R.T.: 6.688 min
Delta R.T.: -0.008 min
Response: 766219
Conc: 9.83 ng/ml



#31 AR-1260-1


R.T.: 7.122 min
Delta R.T.: 0.000 min
Response: 284886
Conc: 2.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.170 min
Delta R.T.: -0.009 min
Response: 329335
Conc: 5.05 ng/ml




#32 AR-1260-2

R.T.: 7.371 min
Delta R.T.: -0.010 min
Response: 1259751
Conc: 11.40 ng/ml



#32 AR-1260-2


R.T.: 6.358 min
Delta R.T.: -0.009 min
Response: 290944
Conc: 3.86 ng/ml



#33 AR-1260-3


R.T.: 7.742 min
Delta R.T.: 0.000 min
Response: 374267
Conc: 4.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.514 min
Delta R.T.: -0.008 min
Response: 271241
Conc: 3.74 ng/ml




#34 AR-1260-4

R.T.: 7.948 min
Delta R.T.: -0.020 min
Response: 3199305
Conc: 34.09 ng/ml



#34 AR-1260-4


R.T.: 6.987 min
Delta R.T.: -0.010 min
Response: 306117
Conc: 5.64 ng/ml



#35 AR-1260-5


R.T.: 8.282 min
Delta R.T.: -0.003 min
Response: 850189
Conc: 5.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.229 min
Delta R.T.: -0.010 min
Response: 703759
Conc: 5.89 ng/ml




#36 AR-1262-1

R.T.: 7.742 min
Delta R.T.: 0.000 min
Response: 374267
Conc: 3.13 ng/ml



#36 AR-1262-1


R.T.: 6.781 min
Delta R.T.: -0.008 min
Response: 122168
Conc: 3.43 ng/ml



#37 AR-1262-2


R.T.: 8.282 min
Delta R.T.: -0.003 min
Response: 850189
Conc: 4.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.987 min
Delta R.T.: -0.010 min
Response: 306117
Conc: 3.94 ng/ml




#38 AR-1262-3

R.T.: 8.602 min
Delta R.T.: 0.004 min
Response: 835687
Conc: 6.03 ng/ml



#38 AR-1262-3


R.T.: 7.515 min
Delta R.T.: -0.010 min
Response: 371177
Conc: 5.89 ng/ml



#39 AR-1262-4


R.T.: 8.687 min
Delta R.T.: 0.000 min
Response: 1007899
Conc: 16.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.579 min
Delta R.T.: -0.010 min
Response: 506058
Conc: 4.43 ng/ml




#40 AR-1262-5

R.T.: 9.336 min
Delta R.T.: -0.008 min
Response: 462227
Conc: 6.54 ng/ml



#40 AR-1262-5


R.T.: 8.073 min
Delta R.T.: -0.011 min
Response: 377276
Conc: 7.57 ng/ml



#41 AR-1268-1


R.T.: 8.602 min
Delta R.T.: 0.007 min
Response: 835687
Conc: 3.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.515 min
Delta R.T.: -0.011 min
Response: 371177
Conc: 2.05 ng/ml




#42 AR-1268-2

R.T.: 8.687 min
Delta R.T.: -0.003 min
Response: 1007899
Conc: 4.43 ng/ml



#42 AR-1268-2


R.T.: 7.579 min
Delta R.T.: -0.011 min
Response: 506058
Conc: 2.97 ng/ml



#43 AR-1268-3


R.T.: 8.916 min
Delta R.T.: -0.002 min
Response: 329665
Conc: 1.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.789 min
Delta R.T.: -0.010 min
Response: 153296
Conc: 1.04 ng/ml




#44 AR-1268-4

R.T.: 9.336 min
Delta R.T.: -0.007 min
Response: 462227
Conc: 6.00 ng/ml



#44 AR-1268-4


R.T.: 8.073 min
Delta R.T.: -0.011 min
Response: 377276
Conc: 6.77 ng/ml



#45 AR-1268-5

R.T.: 9.759 min
Delta R.T.: -0.002 min
Response: 714131
Conc: 1.21 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.366 min
Delta R.T.: -0.012 min
Response: 453149
Conc: 0.96 ng/ml