

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122320\
 Data File : PP032360.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Dec 2020 12:21
 Operator : DD\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 23 13:28:36 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 23 13:05:40 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.772	4.041	4459937	4315045	95.569	95.866
2) SA Decachlor...	10.625	9.326	2897821	3107012	96.966	96.289
Target Compounds						
3) L1 AR-1016-1	6.074	5.291	1181848	1115554	937.851	941.520
4) L1 AR-1016-2	6.097	5.311	1865348	1730841	946.682	945.118
5) L1 AR-1016-3	6.163	5.503	1182751	929627	969.833	935.936
6) L1 AR-1016-4	6.268	5.553	972001	734329	985.927	910.071
7) L1 AR-1016-5	6.582	5.782	857565	961166	947.492	927.109
31) L7 AR-1260-1	7.752	6.874	1400456	1643846	928.997	939.126
32) L7 AR-1260-2	8.017	7.071	1784276	1910019	952.022	965.129
33) L7 AR-1260-3	8.382	7.225	1616206	1869775	956.424	959.344
34) L7 AR-1260-4	8.610	7.707	1541357	1541490	962.825	939.768
35) L7 AR-1260-5	8.924	7.954	3219225	3405165	996.030	986.119



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122320\
Data File : PP032360.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Dec 2020 12:21
Operator : DD\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 23 13:28:36 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122320.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 23 13:05:40 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 4459937
Conc: 95.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#1 Tetrachloro-m-xylene

R.T.: 4.041 min
Delta R.T.: 0.000 min
Response: 4315045
Conc: 95.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.625 min
Delta R.T.: 0.000 min
Response: 2897821
Conc: 96.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.326 min
Delta R.T.: 0.000 min
Response: 3107012
Conc: 96.29 ng/ml



#3 AR-1016-1


R.T.: 6.074 min
Delta R.T.: 0.000 min
Response: 1181848
Conc: 937.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#3 AR-1016-1

R.T.: 5.291 min
Delta R.T.: 0.000 min
Response: 1115554
Conc: 941.52 ng/ml




#4 AR-1016-2

R.T.: 6.097 min
Delta R.T.: 0.000 min
Response: 1865348
Conc: 946.68 ng/ml



#4 AR-1016-2


R.T.: 5.311 min
Delta R.T.: 0.000 min
Response: 1730841
Conc: 945.12 ng/ml



#5 AR-1016-3


R.T.: 6.163 min
Delta R.T.: -0.001 min
Response: 1182751
Conc: 969.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 5.503 min
Delta R.T.: 0.000 min
Response: 929627
Conc: 935.94 ng/ml




#6 AR-1016-4

R.T.: 6.268 min
Delta R.T.: -0.001 min
Response: 972001
Conc: 985.93 ng/ml



#6 AR-1016-4


R.T.: 5.553 min
Delta R.T.: 0.000 min
Response: 734329
Conc: 910.07 ng/ml



#7 AR-1016-5


R.T.: 6.582 min
Delta R.T.: 0.000 min
Response: 857565
Conc: 947.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 5.782 min
Delta R.T.: 0.000 min
Response: 961166
Conc: 927.11 ng/ml




#31 AR-1260-1

R.T.: 7.752 min
Delta R.T.: 0.000 min
Response: 1400456
Conc: 929.00 ng/ml



#31 AR-1260-1


R.T.: 6.874 min
Delta R.T.: 0.000 min
Response: 1643846
Conc: 939.13 ng/ml



#32 AR-1260-2


R.T.: 8.017 min
Delta R.T.: 0.000 min
Response: 1784276
Conc: 952.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 7.071 min
Delta R.T.: 0.000 min
Response: 1910019
Conc: 965.13 ng/ml




#33 AR-1260-3

R.T.: 8.382 min
Delta R.T.: 0.000 min
Response: 1616206
Conc: 956.42 ng/ml



#33 AR-1260-3


R.T.: 7.225 min
Delta R.T.: 0.000 min
Response: 1869775
Conc: 959.34 ng/ml



#34 AR-1260-4


R.T.: 8.610 min
Delta R.T.: 0.000 min
Response: 1541357
Conc: 962.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 7.707 min
Delta R.T.: 0.000 min
Response: 1541490
Conc: 939.77 ng/ml



#35 AR-1260-5

R.T.: 8.924 min
Delta R.T.: 0.000 min
Response: 3219225
Conc: 996.03 ng/ml



#35 AR-1260-5

R.T.: 7.954 min
Delta R.T.: 0.000 min
Response: 3405165
Conc: 986.12 ng/ml