

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122320\
 Data File : PP032370.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Dec 2020 15:08
 Operator : DD\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 23 15:31:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 23 15:19:14 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.771	4.041	1181933	1129278	25.244	25.018
2)	SA Decachlor...	10.624	9.327	814578	853627	26.381	25.902
Target Compounds							
16)	L4 AR-1242-1	6.074	5.291	239735	239880	251.775	261.672
17)	L4 AR-1242-2	6.097	5.312	379390	364969	252.921	259.246
18)	L4 AR-1242-3	6.163	5.503	231164	196387	250.514	254.363
19)	L4 AR-1242-4	6.268	5.599	187811	191020	250.033	254.018
20)	L4 AR-1242-5	7.046	6.161	175326	196737	245.934	247.496



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122320\
Data File : PP032370.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Dec 2020 15:08
Operator : DD\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 23 15:31:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122320.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 23 15:19:14 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 1181933
Conc: 25.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 4.041 min
Delta R.T.: 0.000 min
Response: 1129278
Conc: 25.02 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.624 min
Delta R.T.: 0.000 min
Response: 814578
Conc: 26.38 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.327 min
Delta R.T.: 0.000 min
Response: 853627
Conc: 25.90 ng/ml



#16 AR-1242-1


R.T.: 6.074 min
Delta R.T.: 0.000 min
Response: 239735
Conc: 251.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 5.291 min
Delta R.T.: 0.000 min
Response: 239880
Conc: 261.67 ng/ml




#17 AR-1242-2

R.T.: 6.097 min
Delta R.T.: 0.000 min
Response: 379390
Conc: 252.92 ng/ml



#17 AR-1242-2


R.T.: 5.312 min
Delta R.T.: 0.000 min
Response: 364969
Conc: 259.25 ng/ml



#18 AR-1242-3


R.T.: 6.163 min
Delta R.T.: 0.000 min
Response: 231164
Conc: 250.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 5.503 min
Delta R.T.: 0.000 min
Response: 196387
Conc: 254.36 ng/ml




#19 AR-1242-4

R.T.: 6.268 min
Delta R.T.: 0.000 min
Response: 187811
Conc: 250.03 ng/ml



#19 AR-1242-4


R.T.: 5.599 min
Delta R.T.: 0.000 min
Response: 191020
Conc: 254.02 ng/ml



#20 AR-1242-5

R.T.: 7.046 min
Delta R.T.: 0.000 min
Response: 175326
Conc: 245.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 6.161 min
Delta R.T.: 0.000 min
Response: 196737
Conc: 247.50 ng/ml