

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122321\
 Data File : PP042188.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Dec 2021 12:53
 Operator : AJ\MA
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP122321

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 24 23:52:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 24 23:47:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.925	4.063	597344	1109454	45.495	48.576
2) SA Decachlor...	10.920	9.415	656355	847544	53.887	52.409
Target Compounds						
26) L6 AR-1254-1	7.145	6.210	228100	587828	457.547	469.969
27) L6 AR-1254-2	7.374	6.369	351135	505515	458.019	464.527
28) L6 AR-1254-3	7.754	6.794	355212	775228	460.373	466.628
29) L6 AR-1254-4	8.049	7.033	253880	467054	456.296	478.767
30) L6 AR-1254-5	8.480	7.466	288874	618721	461.315	482.840



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122321\
Data File : PP042188.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Dec 2021 12:53
Operator : AJ\MA
Sample : AR1254ICV500
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP122321

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 24 23:52:59 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 24 23:47:53 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.925 min
Delta R.T.: 0.000 min
Response: 597344
Conc: 45.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP122321




#1 Tetrachloro-m-xylene

R.T.: 4.063 min
Delta R.T.: 0.000 min
Response: 1109454
Conc: 48.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.920 min
Delta R.T.: 0.002 min
Response: 656355
Conc: 53.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.415 min
Delta R.T.: -0.003 min
Response: 847544
Conc: 52.41 ng/ml



#26 AR-1254-1


R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 228100
Conc: 457.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP122321




#26 AR-1254-1

R.T.: 6.210 min
Delta R.T.: -0.001 min
Response: 587828
Conc: 469.97 ng/ml




#27 AR-1254-2

R.T.: 7.374 min
Delta R.T.: 0.000 min
Response: 351135
Conc: 458.02 ng/ml



#27 AR-1254-2


R.T.: 6.369 min
Delta R.T.: -0.002 min
Response: 505515
Conc: 464.53 ng/ml



#28 AR-1254-3


R.T.: 7.754 min
Delta R.T.: 0.000 min
Response: 355212
Conc: 460.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP122321




#28 AR-1254-3

R.T.: 6.794 min
Delta R.T.: -0.001 min
Response: 775228
Conc: 466.63 ng/ml




#29 AR-1254-4

R.T.: 8.049 min
Delta R.T.: 0.000 min
Response: 253880
Conc: 456.30 ng/ml



#29 AR-1254-4


R.T.: 7.033 min
Delta R.T.: -0.002 min
Response: 467054
Conc: 478.77 ng/ml



#30 AR-1254-5

R.T.: 8.480 min
Delta R.T.: 0.000 min
Response: 288874
Conc: 461.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP122321



#30 AR-1254-5

R.T.: 7.466 min
Delta R.T.: -0.001 min
Response: 618721
Conc: 482.84 ng/ml