

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122322\
 Data File : PP053992.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Dec 2022 10:28
 Operator : YP\AJ
 Sample : PB149845BL
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 24 06:02:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 14 05:38:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.342	3.591	45005269	29887810	21.433	19.138
2)	SA Decachlor...	10.115	8.631	36018064	31154025	21.215	22.305
Target Compounds							
3)	L1 AR-1016-1	5.510	4.693	22900	32620	0.338	0.710 #
4)	L1 AR-1016-2	5.534	4.693	34984	32620	0.351	0.521 #
5)	L1 AR-1016-3	5.597	0.000	96949	0	1.534	N.D. #
6)	L1 AR-1016-4	5.685	0.000	46820	0	0.945	N.D. #
7)	L1 AR-1016-5	5.998	5.156f	65211	114121	1.239	3.147 #
8)	L2 AR-1221-1	4.549	0.000	1615635	0	62.714	N.D. #
9)	L2 AR-1221-2	4.649	3.891	746544	540698	38.346	36.990
10)	L2 AR-1221-3	4.714	0.000	125463	0	2.159	N.D. #
11)	L3 AR-1232-1	4.714	0.000	125463	0	2.801	N.D. #
12)	L3 AR-1232-2	5.257f	4.693	35420	32620	1.377	1.093
13)	L3 AR-1232-3	5.534	0.000	34984	0	0.745	N.D. #
14)	L3 AR-1232-4	5.685	4.978	46820	26482	1.976	1.714
15)	L3 AR-1232-5	5.796	5.156f	92799	114121	5.032	6.930 #
16)	L4 AR-1242-1	5.510	4.693	22900	32620	0.437	0.921 #
17)	L4 AR-1242-2	5.534	4.693	34984	32620	0.459	0.679 #
18)	L4 AR-1242-3	5.597	0.000	96949	0	1.979	N.D. #
19)	L4 AR-1242-4	5.685	4.978	46820	26482	1.220	0.928
20)	L4 AR-1242-5	6.422	5.512f	95088	164447	2.331	5.388 #
21)	L5 AR-1248-1	5.510	4.693	22900	32620	0.543	1.147 #
22)	L5 AR-1248-2	5.796	0.000	92799	0	1.468	N.D. #
23)	L5 AR-1248-3	5.998	4.978	65211	26482	0.931	0.631 #
24)	L5 AR-1248-4	6.410	5.156f	194469	114121	2.757	2.372
25)	L5 AR-1248-5	6.422	5.537	95088	93613	1.349	2.493 #
26)	L6 AR-1254-1	6.383	5.512f	211842	164447	2.734	2.511
27)	L6 AR-1254-2	6.586	5.651	16392	576934	0.141	9.893 #
28)	L6 AR-1254-3	0.000	6.025f	0	530601	N.D.	6.031 #
29)	L6 AR-1254-4	7.233	6.277	428284	1169590	5.402	25.529 #
30)	L6 AR-1254-5	7.668	0.000	794080	0	8.142	N.D. #
31)	L7 AR-1260-1	7.127	0.000	151320	0	1.567	N.D. #
32)	L7 AR-1260-2	7.398f	6.372	3686121	396743	33.367	5.268 #
33)	L7 AR-1260-3	7.732	0.000	148950	0	1.970	N.D. #
34)	L7 AR-1260-4	7.963	0.000	1058551	0	11.279	N.D. #
35)	L7 AR-1260-5	8.277	0.000	599454	0	3.780	N.D. #
36)	L8 AR-1262-1	7.732	0.000	148950	0	1.244	N.D. #
37)	L8 AR-1262-2	8.277	0.000	599454	0	3.126	N.D. #
38)	L8 AR-1262-3	8.602	0.000	57080	0	0.412	N.D. #
39)	L8 AR-1262-4	8.698	7.563f	84685	269936	1.361	2.361 #
40)	L8 AR-1262-5	9.336	8.084	179145	462800	2.535	9.290 #
41)	L9 AR-1268-1	8.591	0.000	101458	0	0.416	N.D. #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122322\
Data File : PP053992.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2022 10:28
Operator : YP\AJ
Sample : PB149845BL
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 24 06:02:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 14 05:38:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.698	7.563f	84685	269936	0.373	1.582 #
43)	L9 AR-1268-3	8.927	0.000	80366	0	0.419	N.D. #
44)	L9 AR-1268-4	9.336	8.084	179145	462800	2.327	8.304 #
45)	L9 AR-1268-5	9.762	8.374	179311	406041	0.305	0.859 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122322\
Data File : PP053992.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2022 10:28
Operator : YP\AJ
Sample : PB149845BL
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 24 06:02:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 14 05:38:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.342 min
Delta R.T.: 0.003 min
Response: 45005269
Conc: 21.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.591 min
Delta R.T.: -0.002 min
Response: 29887810
Conc: 19.14 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.115 min
Delta R.T.: 0.013 min
Response: 36018064
Conc: 21.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.631 min
Delta R.T.: -0.005 min
Response: 31154025
Conc: 22.30 ng/ml



#3 AR-1016-1


R.T.: 5.510 min
Delta R.T.: -0.002 min
Response: 22900
Conc: 0.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.693 min
Delta R.T.: 0.007 min
Response: 32620
Conc: 0.71 ng/ml




#4 AR-1016-2

R.T.: 5.534 min
Delta R.T.: 0.000 min
Response: 34984
Conc: 0.35 ng/ml



#4 AR-1016-2


R.T.: 4.693 min
Delta R.T.: -0.013 min
Response: 32620
Conc: 0.52 ng/ml



#5 AR-1016-3


R.T.: 5.597 min
Delta R.T.: 0.000 min
Response: 96949
Conc: 1.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.882 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.685 min
Delta R.T.: -0.010 min
Response: 46820
Conc: 0.94 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 4.924 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 5.998 min
Delta R.T.: 0.007 min
Response: 65211
Conc: 1.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.156 min
Delta R.T.: 0.017 min
Response: 114121
Conc: 3.15 ng/ml




#8 AR-1221-1

R.T.: 4.549 min
Delta R.T.: 0.000 min
Response: 1615635
Conc: 62.71 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.810 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.649 min
Delta R.T.: 0.014 min
Response: 746544
Conc: 38.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.891 min
Delta R.T.: -0.004 min
Response: 540698
Conc: 36.99 ng/ml




#10 AR-1221-3

R.T.: 4.714 min
Delta R.T.: 0.002 min
Response: 125463
Conc: 2.16 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 3.972 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.714 min
Delta R.T.: 0.003 min
Response: 125463
Conc: 2.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.971 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.257 min
Delta R.T.: 0.015 min
Response: 35420
Conc: 1.38 ng/ml



#12 AR-1232-2


R.T.: 4.693 min
Delta R.T.: -0.012 min
Response: 32620
Conc: 1.09 ng/ml



#13 AR-1232-3


R.T.: 5.534 min
Delta R.T.: 0.000 min
Response: 34984
Conc: 0.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.882 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.685 min
Delta R.T.: -0.010 min
Response: 46820
Conc: 1.98 ng/ml



#14 AR-1232-4


R.T.: 4.978 min
Delta R.T.: 0.012 min
Response: 26482
Conc: 1.71 ng/ml



#15 AR-1232-5


R.T.: 5.796 min
Delta R.T.: 0.010 min
Response: 92799
Conc: 5.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.156 min
Delta R.T.: 0.018 min
Response: 114121
Conc: 6.93 ng/ml




#16 AR-1242-1

R.T.: 5.510 min
Delta R.T.: -0.001 min
Response: 22900
Conc: 0.44 ng/ml



#16 AR-1242-1


R.T.: 4.693 min
Delta R.T.: 0.007 min
Response: 32620
Conc: 0.92 ng/ml



#17 AR-1242-2


R.T.: 5.534 min
Delta R.T.: 0.000 min
Response: 34984
Conc: 0.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.693 min
Delta R.T.: -0.012 min
Response: 32620
Conc: 0.68 ng/ml




#18 AR-1242-3

R.T.: 5.597 min
Delta R.T.: 0.000 min
Response: 96949
Conc: 1.98 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T.: 4.882 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.685 min
Delta R.T.: -0.010 min
Response: 46820
Conc: 1.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.978 min
Delta R.T.: 0.012 min
Response: 26482
Conc: 0.93 ng/ml




#20 AR-1242-5

R.T.: 6.422 min
Delta R.T.: -0.013 min
Response: 95088
Conc: 2.33 ng/ml



#20 AR-1242-5


R.T.: 5.512 min
Delta R.T.: 0.020 min
Response: 164447
Conc: 5.39 ng/ml



#21 AR-1248-1


R.T.: 5.510 min
Delta R.T.: -0.002 min
Response: 22900
Conc: 0.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.693 min
Delta R.T.: 0.007 min
Response: 32620
Conc: 1.15 ng/ml




#22 AR-1248-2

R.T.: 5.796 min
Delta R.T.: 0.010 min
Response: 92799
Conc: 1.47 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T.: 4.924 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 5.998 min
Delta R.T.: 0.007 min
Response: 65211
Conc: 0.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.978 min
Delta R.T.: 0.011 min
Response: 26482
Conc: 0.63 ng/ml




#24 AR-1248-4

R.T.: 6.410 min
Delta R.T.: 0.013 min
Response: 194469
Conc: 2.76 ng/ml



#24 AR-1248-4


R.T.: 5.156 min
Delta R.T.: 0.017 min
Response: 114121
Conc: 2.37 ng/ml



#25 AR-1248-5


R.T.: 6.422 min
Delta R.T.: -0.013 min
Response: 95088
Conc: 1.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.537 min
Delta R.T.: 0.004 min
Response: 93613
Conc: 2.49 ng/ml




#26 AR-1254-1

R.T.: 6.383 min
Delta R.T.: 0.014 min
Response: 211842
Conc: 2.73 ng/ml



#26 AR-1254-1


R.T.: 5.512 min
Delta R.T.: 0.020 min
Response: 164447
Conc: 2.51 ng/ml



#27 AR-1254-2


R.T.: 6.586 min
Delta R.T.: -0.003 min
Response: 16392
Conc: 0.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.651 min
Delta R.T.: 0.011 min
Response: 576934
Conc: 9.89 ng/ml




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.957 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 6.025 min
Delta R.T.: -0.022 min
Response: 530601
Conc: 6.03 ng/ml



#29 AR-1254-4


R.T.: 7.233 min
Delta R.T.: -0.011 min
Response: 428284
Conc: 5.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.277 min
Delta R.T.: 0.000 min
Response: 1169590
Conc: 25.53 ng/ml




#30 AR-1254-5

R.T.: 7.668 min
Delta R.T.: 0.004 min
Response: 794080
Conc: 8.14 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T. : 6.696 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 7.127 min
Delta R.T.: 0.004 min
Response: 151320
Conc: 1.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 6.179 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 7.398 min
Delta R.T.: 0.017 min
Response: 3686121
Conc: 33.37 ng/ml



#32 AR-1260-2


R.T.: 6.372 min
Delta R.T.: 0.005 min
Response: 396743
Conc: 5.27 ng/ml



#33 AR-1260-3


R.T.: 7.732 min
Delta R.T.: -0.010 min
Response: 148950
Conc: 1.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.522 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 7.963 min
Delta R.T.: -0.004 min
Response: 1058551
Conc: 11.28 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 6.997 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.277 min
Delta R.T.: -0.007 min
Response: 599454
Conc: 3.78 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 7.239 min
Response: 0
Conc: N.D.




#36 AR-1262-1

R.T.: 7.732 min
Delta R.T.: -0.009 min
Response: 148950
Conc: 1.24 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 6.789 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 8.277 min
Delta R.T.: -0.007 min
Response: 599454
Conc: 3.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 6.996 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 8.602 min
Delta R.T.: 0.004 min
Response: 57080
Conc: 0.41 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T. : 7.525 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 8.698 min
Delta R.T.: 0.011 min
Response: 84685
Conc: 1.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.563 min
Delta R.T.: -0.025 min
Response: 269936
Conc: 2.36 ng/ml




#40 AR-1262-5

R.T.: 9.336 min
Delta R.T.: -0.009 min
Response: 179145
Conc: 2.54 ng/ml



#40 AR-1262-5


R.T.: 8.084 min
Delta R.T.: 0.000 min
Response: 462800
Conc: 9.29 ng/ml



#41 AR-1268-1


R.T.: 8.591 min
Delta R.T.: -0.004 min
Response: 101458
Conc: 0.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 7.526 min
Response: 0
Conc: N.D.




#42 AR-1268-2

R.T.: 8.698 min
Delta R.T.: 0.009 min
Response: 84685
Conc: 0.37 ng/ml



#42 AR-1268-2


R.T.: 7.563 min
Delta R.T.: -0.026 min
Response: 269936
Conc: 1.58 ng/ml



#43 AR-1268-3


R.T.: 8.927 min
Delta R.T.: 0.010 min
Response: 80366
Conc: 0.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.799 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.336 min
Delta R.T.: -0.008 min
Response: 179145
Conc: 2.33 ng/ml



#44 AR-1268-4


R.T.: 8.084 min
Delta R.T.: 0.000 min
Response: 462800
Conc: 8.30 ng/ml



#45 AR-1268-5

R.T.: 9.762 min
Delta R.T.: 0.001 min
Response: 179311
Conc: 0.30 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.374 min
Delta R.T.: -0.004 min
Response: 406041
Conc: 0.86 ng/ml