

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122322\
 Data File : PP054018.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Dec 2022 19:03
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 26 05:40:18 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 14 05:38:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.339	3.589	104.9E6	73021834	49.974	46.757
2)	SA Decachlor...	10.111	8.627	73334756	64654534	43.194	46.289
Target Compounds							
16)	L4 AR-1242-1	5.512	4.680	26595841	17829323	507.740	503.461
17)	L4 AR-1242-2	5.535	4.699	37651402	24633983	494.168	512.757
18)	L4 AR-1242-3	5.597	4.876	24123752	13471179	492.373	516.715
19)	L4 AR-1242-4	5.696	4.960	19421099	14356753	505.905	503.240
20)	L4 AR-1242-5	6.437	5.486	20180342	16554069	494.784	542.343



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122322\
Data File : PP054018.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2022 19:03
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 05:40:18 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 14 05:38:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.339 min
Delta R.T.: 0.000 min
Response: 104934160
Conc: 49.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.589 min
Delta R.T.: -0.005 min
Response: 73021834
Conc: 46.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.111 min
Delta R.T.: 0.009 min
Response: 73334756
Conc: 43.19 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.627 min
Delta R.T.: -0.008 min
Response: 64654534
Conc: 46.29 ng/ml



#16 AR-1242-1


R.T.: 5.512 min
Delta R.T.: 0.000 min
Response: 26595841
Conc: 507.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.680 min
Delta R.T.: -0.006 min
Response: 17829323
Conc: 503.46 ng/ml




#17 AR-1242-2

R.T.: 5.535 min
Delta R.T.: 0.000 min
Response: 37651402
Conc: 494.17 ng/ml



#17 AR-1242-2


R.T.: 4.699 min
Delta R.T.: -0.006 min
Response: 24633983
Conc: 512.76 ng/ml



#18 AR-1242-3


R.T.: 5.597 min
Delta R.T.: 0.000 min
Response: 24123752
Conc: 492.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.876 min
Delta R.T.: -0.006 min
Response: 13471179
Conc: 516.71 ng/ml




#19 AR-1242-4

R.T.: 5.696 min
Delta R.T.: 0.001 min
Response: 19421099
Conc: 505.91 ng/ml



#19 AR-1242-4


R.T.: 4.960 min
Delta R.T.: -0.006 min
Response: 14356753
Conc: 503.24 ng/ml



#20 AR-1242-5

R.T.: 6.437 min
Delta R.T.: 0.002 min
Response: 20180342
Conc: 494.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.486 min
Delta R.T.: -0.006 min
Response: 16554069
Conc: 542.34 ng/ml