

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122420\
 Data File : PP032414.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Dec 2020 17:04
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 25 03:18:41 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 24 07:15:30 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.772	4.044	2332416	2136041	50.217	47.585
2)	SA Decachlor...	10.628	9.329	1588940	1633351	54.239	50.636
Target Compounds							
3)	L1 AR-1016-1	6.075	5.293	645175	597867	511.418	504.689
4)	L1 AR-1016-2	6.098	5.313	987318	904184	496.496	506.697
5)	L1 AR-1016-3	6.164	5.505	592679	493891	494.318	502.843
6)	L1 AR-1016-4	6.269	5.556	493790	384885	496.230	494.700
7)	L1 AR-1016-5	6.583	5.784	465595	483637	522.611	471.218
8)	L2 AR-1221-1	5.014	4.297	64526	59932	137.520	130.348
9)	L2 AR-1221-2	5.116	4.397	101652	91460	283.146	263.919
10)	L2 AR-1221-3	5.200	4.485	371377	348759	342.441	323.157
11)	L3 AR-1232-1	5.200	4.485	371377	348759	411.261	386.550
12)	L3 AR-1232-2	5.783	5.313	510134	904184	1096.505	1117.670
13)	L3 AR-1232-3	6.098	5.505	987318	493891	1164.772	1179.648
14)	L3 AR-1232-4	6.269	5.600	493790	468105	1179.397	1269.067
15)	L3 AR-1232-5	6.369	5.784	361693	483637	1337.466	1204.062
16)	L4 AR-1242-1	6.075	5.293	645175	597867	689.980	657.173
17)	L4 AR-1242-2	6.098	5.313	987318	904184	661.457	660.394
18)	L4 AR-1242-3	6.164	5.505	592679	493891	657.751	649.758
19)	L4 AR-1242-4	6.269	5.600	493790	468105	626.776	638.827
20)	L4 AR-1242-5	7.048	6.162	60377	349731	88.888	437.255 #
21)	L5 AR-1248-1	6.075	5.293	645175	597867	900.059	836.414
22)	L5 AR-1248-2	6.369	5.556	361693	384885	369.055	365.492
23)	L5 AR-1248-3	6.583	5.600	465595	468105	368.274	423.208
24)	L5 AR-1248-4	7.008	5.784	73037	483637	57.308	362.278 #
25)	L5 AR-1248-5	7.048	6.206	60377	52227	45.056	45.120
26)	L6 AR-1254-1	6.979	6.162	287072	349731	208.530	183.154
27)	L6 AR-1254-2	7.208	6.321	311969	310884	148.187	187.768 #
28)	L6 AR-1254-3	7.587	6.758f	204895	623393	94.410	235.159 #
29)	L6 AR-1254-4	7.879	6.980	134313	71812	95.506	57.920 #
30)	L6 AR-1254-5	8.307	7.406	1095615	1063748	686.459	481.443 #
31)	L7 AR-1260-1	7.754	6.876	759445	833074	508.186	473.013
32)	L7 AR-1260-2	8.019	7.073	983649	971389	536.258	488.929
33)	L7 AR-1260-3	8.384	7.228	860109	938497	497.841	476.290
34)	L7 AR-1260-4	8.612	7.709	835777	800156	524.579	492.039
35)	L7 AR-1260-5	8.926	7.957	1835385	1829593	578.570	527.381
36)	L8 AR-1262-1	8.384	7.446	860109	874518	315.253	358.941
37)	L8 AR-1262-2	8.926	7.957	1835385	1829593	435.891	436.332
38)	L8 AR-1262-3	9.238	8.241	1134427	443355	386.719	264.917 #
39)	L8 AR-1262-4	9.309	8.305	329260	1383144	127.804	445.413 #
40)	L8 AR-1262-5	9.932	8.803	530301	530380	352.667	412.510
41)	L9 AR-1268-1	9.238f	8.241	1134427	443355	218.465	90.744 #
42)	L9 AR-1268-2	9.309	8.305	329260	1383144	67.399	284.262 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122420\
 Data File : PP032414.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Dec 2020 17:04
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 25 03:18:41 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 24 07:15:30 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
44) L9	AR-1268-4	9.932	8.803	530301	530380	416.678	405.818
45) L9	AR-1268-5	10.319	9.086	143032	149390	12.218	12.031



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122420\
Data File : PP032414.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Dec 2020 17:04
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 25 03:18:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122320.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 24 07:15:30 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.772 min
Delta R.T.: 0.000 min
Response: 2332416
Conc: 50.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.044 min
Delta R.T.: 0.002 min
Response: 2136041
Conc: 47.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.628 min
Delta R.T.: 0.003 min
Response: 1588940
Conc: 54.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.329 min
Delta R.T.: 0.003 min
Response: 1633351
Conc: 50.64 ng/ml



#3 AR-1016-1


R.T.: 6.075 min
Delta R.T.: 0.001 min
Response: 645175
Conc: 511.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.293 min
Delta R.T.: 0.001 min
Response: 597867
Conc: 504.69 ng/ml




#4 AR-1016-2

R.T.: 6.098 min
Delta R.T.: 0.000 min
Response: 987318
Conc: 496.50 ng/ml



#4 AR-1016-2


R.T.: 5.313 min
Delta R.T.: 0.002 min
Response: 904184
Conc: 506.70 ng/ml



#5 AR-1016-3


R.T.: 6.164 min
Delta R.T.: 0.001 min
Response: 592679
Conc: 494.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.505 min
Delta R.T.: 0.002 min
Response: 493891
Conc: 502.84 ng/ml




#6 AR-1016-4

R.T.: 6.269 min
Delta R.T.: 0.000 min
Response: 493790
Conc: 496.23 ng/ml



#6 AR-1016-4


R.T.: 5.556 min
Delta R.T.: 0.002 min
Response: 384885
Conc: 494.70 ng/ml



#7 AR-1016-5


R.T.: 6.583 min
Delta R.T.: 0.001 min
Response: 465595
Conc: 522.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.784 min
Delta R.T.: 0.002 min
Response: 483637
Conc: 471.22 ng/ml




#8 AR-1221-1

R.T.: 5.014 min
Delta R.T.: -0.002 min
Response: 64526
Conc: 137.52 ng/ml



#8 AR-1221-1


R.T.: 4.297 min
Delta R.T.: 0.000 min
Response: 59932
Conc: 130.35 ng/ml



#9 AR-1221-2


R.T.: 5.116 min
Delta R.T.: 0.002 min
Response: 101652
Conc: 283.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#9 AR-1221-2

R.T.: 4.397 min
Delta R.T.: 0.002 min
Response: 91460
Conc: 263.92 ng/ml




#10 AR-1221-3

R.T.: 5.200 min
Delta R.T.: 0.002 min
Response: 371377
Conc: 342.44 ng/ml



#10 AR-1221-3


R.T.: 4.485 min
Delta R.T.: 0.001 min
Response: 348759
Conc: 323.16 ng/ml



#11 AR-1232-1


R.T.: 5.200 min
Delta R.T.: 0.000 min
Response: 371377
Conc: 411.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#11 AR-1232-1

R.T.: 4.485 min
Delta R.T.: 0.001 min
Response: 348759
Conc: 386.55 ng/ml




#12 AR-1232-2

R.T.: 5.783 min
Delta R.T.: 0.001 min
Response: 510134
Conc: 1096.50 ng/ml



#12 AR-1232-2


R.T.: 5.313 min
Delta R.T.: 0.000 min
Response: 904184
Conc: 1117.67 ng/ml



#13 AR-1232-3


R.T.: 6.098 min
Delta R.T.: 0.001 min
Response: 987318
Conc: 1164.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#13 AR-1232-3

R.T.: 5.505 min
Delta R.T.: 0.001 min
Response: 493891
Conc: 1179.65 ng/ml




#14 AR-1232-4

R.T.: 6.269 min
Delta R.T.: 0.000 min
Response: 493790
Conc: 1179.40 ng/ml



#14 AR-1232-4


R.T.: 5.600 min
Delta R.T.: 0.002 min
Response: 468105
Conc: 1269.07 ng/ml



#15 AR-1232-5


R.T.: 6.369 min
Delta R.T.: 0.000 min
Response: 361693
Conc: 1337.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#15 AR-1232-5

R.T.: 5.784 min
Delta R.T.: 0.001 min
Response: 483637
Conc: 1204.06 ng/ml




#16 AR-1242-1

R.T.: 6.075 min
Delta R.T.: 0.001 min
Response: 645175
Conc: 689.98 ng/ml



#16 AR-1242-1


R.T.: 5.293 min
Delta R.T.: 0.002 min
Response: 597867
Conc: 657.17 ng/ml



#17 AR-1242-2


R.T.: 6.098 min
Delta R.T.: 0.001 min
Response: 987318
Conc: 661.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#17 AR-1242-2

R.T.: 5.313 min
Delta R.T.: 0.001 min
Response: 904184
Conc: 660.39 ng/ml




#18 AR-1242-3

R.T.: 6.164 min
Delta R.T.: 0.000 min
Response: 592679
Conc: 657.75 ng/ml



#18 AR-1242-3


R.T.: 5.505 min
Delta R.T.: 0.002 min
Response: 493891
Conc: 649.76 ng/ml



#19 AR-1242-4


R.T.: 6.269 min
Delta R.T.: 0.000 min
Response: 493790
Conc: 626.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#19 AR-1242-4

R.T.: 5.600 min
Delta R.T.: 0.001 min
Response: 468105
Conc: 638.83 ng/ml




#20 AR-1242-5

R.T.: 7.048 min
Delta R.T.: 0.002 min
Response: 60377
Conc: 88.89 ng/ml



#20 AR-1242-5


R.T.: 6.162 min
Delta R.T.: 0.001 min
Response: 349731
Conc: 437.26 ng/ml



#21 AR-1248-1


R.T.: 6.075 min
Delta R.T.: 0.001 min
Response: 645175
Conc: 900.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#21 AR-1248-1

R.T.: 5.293 min
Delta R.T.: 0.002 min
Response: 597867
Conc: 836.41 ng/ml




#22 AR-1248-2

R.T.: 6.369 min
Delta R.T.: 0.001 min
Response: 361693
Conc: 369.06 ng/ml



#22 AR-1248-2


R.T.: 5.556 min
Delta R.T.: 0.002 min
Response: 384885
Conc: 365.49 ng/ml



#23 AR-1248-3


R.T.: 6.583 min
Delta R.T.: 0.000 min
Response: 465595
Conc: 368.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#23 AR-1248-3

R.T.: 5.600 min
Delta R.T.: 0.002 min
Response: 468105
Conc: 423.21 ng/ml




#24 AR-1248-4

R.T.: 7.008 min
Delta R.T.: 0.001 min
Response: 73037
Conc: 57.31 ng/ml



#24 AR-1248-4


R.T.: 5.784 min
Delta R.T.: 0.003 min
Response: 483637
Conc: 362.28 ng/ml



#25 AR-1248-5


R.T.: 7.048 min
Delta R.T.: 0.002 min
Response: 60377
Conc: 45.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#25 AR-1248-5

R.T.: 6.206 min
Delta R.T.: 0.002 min
Response: 52227
Conc: 45.12 ng/ml




#26 AR-1254-1

R.T.: 6.979 min
Delta R.T.: -0.010 min
Response: 287072
Conc: 208.53 ng/ml



#26 AR-1254-1


R.T.: 6.162 min
Delta R.T.: 0.000 min
Response: 349731
Conc: 183.15 ng/ml



#27 AR-1254-2


R.T.: 7.208 min
Delta R.T.: -0.010 min
Response: 311969
Conc: 148.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#27 AR-1254-2

R.T.: 6.321 min
Delta R.T.: 0.000 min
Response: 310884
Conc: 187.77 ng/ml




#28 AR-1254-3

R.T.: 7.587 min
Delta R.T.: -0.009 min
Response: 204895
Conc: 94.41 ng/ml



#28 AR-1254-3


R.T.: 6.758 min
Delta R.T.: 0.017 min
Response: 623393
Conc: 235.16 ng/ml



#29 AR-1254-4


R.T.: 7.879 min
Delta R.T.: -0.011 min
Response: 134313
Conc: 95.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#29 AR-1254-4

R.T.: 6.980 min
Delta R.T.: 0.000 min
Response: 71812
Conc: 57.92 ng/ml




#30 AR-1254-5

R.T.: 8.307 min
Delta R.T.: -0.010 min
Response: 1095615
Conc: 686.46 ng/ml



#30 AR-1254-5


R.T.: 7.406 min
Delta R.T.: -0.001 min
Response: 1063748
Conc: 481.44 ng/ml



#31 AR-1260-1


R.T.: 7.754 min
Delta R.T.: 0.002 min
Response: 759445
Conc: 508.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#31 AR-1260-1

R.T.: 6.876 min
Delta R.T.: 0.002 min
Response: 833074
Conc: 473.01 ng/ml




#32 AR-1260-2

R.T.: 8.019 min
Delta R.T.: 0.002 min
Response: 983649
Conc: 536.26 ng/ml



#32 AR-1260-2


R.T.: 7.073 min
Delta R.T.: 0.002 min
Response: 971389
Conc: 488.93 ng/ml



#33 AR-1260-3


R.T.: 8.384 min
Delta R.T.: 0.002 min
Response: 860109
Conc: 497.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#33 AR-1260-3

R.T.: 7.228 min
Delta R.T.: 0.002 min
Response: 938497
Conc: 476.29 ng/ml




#34 AR-1260-4

R.T.: 8.612 min
Delta R.T.: 0.002 min
Response: 835777
Conc: 524.58 ng/ml



#34 AR-1260-4


R.T.: 7.709 min
Delta R.T.: 0.002 min
Response: 800156
Conc: 492.04 ng/ml



#35 AR-1260-5


R.T.: 8.926 min
Delta R.T.: 0.002 min
Response: 1835385
Conc: 578.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#35 AR-1260-5

R.T.: 7.957 min
Delta R.T.: 0.002 min
Response: 1829593
Conc: 527.38 ng/ml




#36 AR-1262-1

R.T.: 8.384 min
Delta R.T.: -0.011 min
Response: 860109
Conc: 315.25 ng/ml



#36 AR-1262-1


R.T.: 7.446 min
Delta R.T.: -0.001 min
Response: 874518
Conc: 358.94 ng/ml



#37 AR-1262-2


R.T.: 8.926 min
Delta R.T.: -0.011 min
Response: 1835385
Conc: 435.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#37 AR-1262-2

R.T.: 7.957 min
Delta R.T.: -0.001 min
Response: 1829593
Conc: 436.33 ng/ml




#38 AR-1262-3

R.T.: 9.238 min
Delta R.T.: 0.005 min
Response: 1134427
Conc: 386.72 ng/ml



#38 AR-1262-3


R.T.: 8.241 min
Delta R.T.: -0.001 min
Response: 443355
Conc: 264.92 ng/ml



#39 AR-1262-4


R.T.: 9.309 min
Delta R.T.: -0.013 min
Response: 329260
Conc: 127.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#39 AR-1262-4

R.T.: 8.305 min
Delta R.T.: -0.001 min
Response: 1383144
Conc: 445.41 ng/ml




#40 AR-1262-5

R.T.: 9.932 min
Delta R.T.: -0.014 min
Response: 530301
Conc: 352.67 ng/ml



#40 AR-1262-5


R.T.: 8.803 min
Delta R.T.: -0.001 min
Response: 530380
Conc: 412.51 ng/ml



#41 AR-1268-1


R.T.: 9.238 min
Delta R.T.: 0.021 min
Response: 1134427
Conc: 218.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#41 AR-1268-1

R.T.: 8.241 min
Delta R.T.: 0.003 min
Response: 443355
Conc: 90.74 ng/ml




#42 AR-1268-2

R.T.: 9.309 min
Delta R.T.: 0.000 min
Response: 329260
Conc: 67.40 ng/ml



#42 AR-1268-2


R.T.: 8.305 min
Delta R.T.: 0.001 min
Response: 1383144
Conc: 284.26 ng/ml



#44 AR-1268-4


R.T.: 9.932 min
Delta R.T.: 0.004 min
Response: 530301
Conc: 416.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#44 AR-1268-4

R.T.: 8.803 min
Delta R.T.: 0.003 min
Response: 530380
Conc: 405.82 ng/ml



#45 AR-1268-5

R.T.: 10.319 min
Delta R.T.: 0.006 min
Response: 143032
Conc: 12.22 ng/ml



#45 AR-1268-5

R.T.: 9.086 min
Delta R.T.: 0.004 min
Response: 149390
Conc: 12.03 ng/ml