

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122420\
 Data File : PP032426.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Dec 2020 20:24
 Operator : DD\AJ
 Sample : L5208-07
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 DUP-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 25 03:26:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 24 07:15:30 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.771	4.044	1061709	967336	22.859	21.550
2)	SA Decachlor...	10.628	9.329	591263	597096	20.183	18.511
Target Compounds							
26)	L6 AR-1254-1	6.978	6.162	2918158	4198664	2119.754	2198.831
27)	L6 AR-1254-2	7.207	6.320	7201687	6096632	3420.842	3682.243
28)	L6 AR-1254-3	7.586	6.740	10506724	12110547	4841.202	4568.397
29)	L6 AR-1254-4	7.880	6.980	9526648	8357548	6774.124	6740.728
30)	L6 AR-1254-5	8.307	7.407	16409470	22273162	10281.367	10080.632



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122420\
Data File : PP032426.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Dec 2020 20:24
Operator : DD\AJ
Sample : L5208-07
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
DUP-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 25 03:26:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122320.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 24 07:15:30 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 1061709
Conc: 22.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
DUP-1




#1 Tetrachloro-m-xylene

R.T.: 4.044 min
Delta R.T.: 0.002 min
Response: 967336
Conc: 21.55 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.628 min
Delta R.T.: 0.003 min
Response: 591263
Conc: 20.18 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.329 min
Delta R.T.: 0.003 min
Response: 597096
Conc: 18.51 ng/ml



#26 AR-1254-1


R.T.: 6.978 min
Delta R.T.: -0.011 min
Response: 2918158
Conc: 2119.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
DUP-1




#26 AR-1254-1

R.T.: 6.162 min
Delta R.T.: -0.001 min
Response: 4198664
Conc: 2198.83 ng/ml




#27 AR-1254-2

R.T.: 7.207 min
Delta R.T.: -0.011 min
Response: 7201687
Conc: 3420.84 ng/ml



#27 AR-1254-2


R.T.: 6.320 min
Delta R.T.: -0.001 min
Response: 6096632
Conc: 3682.24 ng/ml



#28 AR-1254-3


R.T.: 7.586 min
Delta R.T.: -0.010 min
Response: 10506724
Conc: 4841.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
DUP-1




#28 AR-1254-3

R.T.: 6.740 min
Delta R.T.: -0.001 min
Response: 12110547
Conc: 4568.40 ng/ml




#29 AR-1254-4

R.T.: 7.880 min
Delta R.T.: -0.010 min
Response: 9526648
Conc: 6774.12 ng/ml



#29 AR-1254-4


R.T.: 6.980 min
Delta R.T.: 0.000 min
Response: 8357548
Conc: 6740.73 ng/ml



#30 AR-1254-5

R.T.: 8.307 min
Delta R.T.: -0.009 min
Response: 16409470
Conc: 10281.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
DUP-1



#30 AR-1254-5

R.T.: 7.407 min
Delta R.T.: -0.001 min
Response: 22273162
Conc: 10080.63 ng/ml