

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP122619.M
 Title : 8080.M
 Last Update : Fri Dec 27 06:31:24 2019
 Response Via : Initial Calibration

Calibration Files

750 =PP030550.D 1000 =PP030551.D 1500 =PP030552.D
 500 =PP030549.D 200 =PP030548.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	5.384	5.271	5.414	5.428	5.384	5.376 E5	1.15
2) T	3,5-DICHLOROBENZOI...	6.491	6.134	6.074	6.665	6.864	6.446 E5	5.26
3) T	4-Nitrophenol	4.420	4.210	4.278	4.424	3.841	4.235 E5	5.64
4) S	2,4-DCAA	3.878	3.618	3.614	4.013	4.210	3.867 E5	6.66
5) T	DICAMBA	1.588	1.500	1.502	1.608	1.630	1.566 E6	3.88
6) T	MCPP	1.104	1.023	0.999	1.149	1.257	1.106 E6	9.39
7) T	MCPA	1.570	1.436	1.406	1.661	1.951	1.605 E6	13.67
8) T	DICHLORPROP	3.907	3.595	3.525	4.036	4.252	3.863 E5	7.87
9) T	2,4-D	4.302	3.944	3.897	4.358	4.114	4.123 E5	5.01
10) T	Pentachlorophenol	6.206	5.774	5.672	6.276	6.189	6.024 E6	4.63
11) T	2,4,5-TP (SILVEX)	2.160	1.983	1.994	2.169	2.177	2.097 E6	4.73
12) T	2,4,5-T	2.087	1.913	1.930	2.075	2.007	2.002 E6	3.98
13) T	2,4-DB	2.666	2.443	2.487	2.676	2.685	2.591 E5	4.50
14) T	DINOSEB	8.860	8.333	9.250	7.766	6.054	8.052 E5	15.51
15) T	Picloram	1.992	2.029	2.126	1.844	1.563	1.911 E6	11.47
16) T	DCPA	2.658	2.427	2.497	2.614	2.554	2.550 E6	3.60

Signal #2 Calibration Files

750 =PP030550.D 1000 =PP030551.D 1500 =PP030552.D
 500 =PP030549.D 200 =PP030548.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	2.522	2.486	2.529	2.477	2.421	2.487 E5	1.73
2) T	3,5-DICHLOROBENZOI...	2.791	2.679	2.674	2.808	2.820	2.754 E5	2.60
3) T	4-Nitrophenol	1.075	1.060	1.095	0.989	0.772	0.998 E5	13.30
4) S	2,4-DCAA	2.130	2.030	2.019	2.124	2.087	2.078 E5	2.49
5) T	DICAMBA	8.286	7.961	8.021	8.242	7.849	8.072 E5	2.31
6) T	MCPP	4.754	4.601	4.747	4.534	3.457	4.419 E5	12.35
7) T	MCPA	7.159	6.867	7.046	6.893	6.063	6.806 E5	6.34
8) T	DICHLORPROP	2.053	1.927	1.934	2.066	2.080	2.012 E5	3.73
9) T	2,4-D	2.312	2.168	2.175	2.197	2.064	2.183 E5	4.05
10) T	Pentachlorophenol	3.262	3.069	3.002	3.282	3.184	3.160 E6	3.85
11) T	2,4,5-TP (SILVEX)	1.190	1.117	1.127	1.177	1.143	1.151 E6	2.74
12) T	2,4,5-T	1.006	0.954	0.981	0.963	0.852	0.951 E6	6.18
13) T	2,4-DB	1.360	1.269	1.325	1.321	1.255	1.306 E5	3.31
14) T	DINOSEB	4.357	4.166	4.758	3.672	2.742	3.939 E5	19.66
15) T	Picloram	1.116	1.160	1.245	1.006	0.770	1.059 E6	17.30
16) T	DCPA	1.493	1.384	1.410	1.486	1.458	1.446 E6	3.31

(#) = Out of Range