

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122623\
 Data File : PP062590.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Dec 2023 10:56
 Operator : YP\AJ
 Sample : 05964-09
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PC0MA4

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 00:47:07 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 22 03:29:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.397	3.683	51808972	67038370	23.830	22.464
2) SA Decachlor...	10.144	8.783	35747771	58318118	24.790	22.908

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122623\
Data File : PP062590.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2023 10:56
Operator : YP\AJ
Sample : 05964-09
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PC0MA4

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 00:47:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 22 03:29:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.397 min
Delta R.T.: 0.000 min
Response: 51808972
Conc: 23.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
PC0MA4




#1 Tetrachloro-m-xylene

R.T.: 3.683 min
Delta R.T.: 0.001 min
Response: 67038370
Conc: 22.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.144 min
Delta R.T.: 0.002 min
Response: 35747771
Conc: 24.79 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.783 min
Delta R.T.: 0.000 min
Response: 58318118
Conc: 22.91 ng/ml