

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122623\
 Data File : PP062602.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Dec 2023 14:11
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 00:51:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 22 03:29:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.397	3.684	109.9E6	147.0E6	50.534	49.244
2) SA Decachlor...	10.145	8.783	72272804	118.7E6	50.120	46.609
Target Compounds						
3) L1 AR-1016-1	5.569	4.791	34893916	44819405	512.000	477.314
4) L1 AR-1016-2	5.591	4.810	52090893	64800786	513.319	483.839
5) L1 AR-1016-3	5.654	4.990	32460737	35393207	515.124	484.535
6) L1 AR-1016-4	5.752	5.032	26540498	29251061	517.762	480.174
7) L1 AR-1016-5	6.047	5.249	27676769	38409098	522.430	477.392
31) L7 AR-1260-1	7.178	6.298	51065997	74440832	515.809	465.689
32) L7 AR-1260-2	7.436	6.488	57384118	86084854	505.761	470.185
33) L7 AR-1260-3	7.796	6.644	42869096	83805651	500.323	465.951
34) L7 AR-1260-4	8.022	7.122	47291272	68591474	488.346	466.161
35) L7 AR-1260-5	8.341	7.365	84728712	153.4E6	475.362	476.070



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122623\
Data File : PP062602.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2023 14:11
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 00:51:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 22 03:29:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.397 min
Delta R.T.: 0.000 min
Response: 109864547
Conc: 50.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.002 min
Response: 146955483
Conc: 49.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.145 min
Delta R.T.: 0.004 min
Response: 72272804
Conc: 50.12 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.783 min
Delta R.T.: 0.002 min
Response: 118652943
Conc: 46.61 ng/ml



#3 AR-1016-1


R.T.: 5.569 min
Delta R.T.: 0.002 min
Response: 34893916
Conc: 512.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.791 min
Delta R.T.: 0.001 min
Response: 44819405
Conc: 477.31 ng/ml




#4 AR-1016-2

R.T.: 5.591 min
Delta R.T.: 0.001 min
Response: 52090893
Conc: 513.32 ng/ml



#4 AR-1016-2


R.T.: 4.810 min
Delta R.T.: 0.000 min
Response: 64800786
Conc: 483.84 ng/ml



#5 AR-1016-3


R.T.: 5.654 min
Delta R.T.: 0.002 min
Response: 32460737
Conc: 515.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.990 min
Delta R.T.: 0.002 min
Response: 35393207
Conc: 484.53 ng/ml




#6 AR-1016-4

R.T.: 5.752 min
Delta R.T.: 0.001 min
Response: 26540498
Conc: 517.76 ng/ml



#6 AR-1016-4


R.T.: 5.032 min
Delta R.T.: 0.000 min
Response: 29251061
Conc: 480.17 ng/ml



#7 AR-1016-5


R.T.: 6.047 min
Delta R.T.: 0.001 min
Response: 27676769
Conc: 522.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.249 min
Delta R.T.: 0.000 min
Response: 38409098
Conc: 477.39 ng/ml




#31 AR-1260-1

R.T.: 7.178 min
Delta R.T.: 0.002 min
Response: 51065997
Conc: 515.81 ng/ml



#31 AR-1260-1


R.T.: 6.298 min
Delta R.T.: 0.000 min
Response: 74440832
Conc: 465.69 ng/ml



#32 AR-1260-2


R.T.: 7.436 min
Delta R.T.: 0.002 min
Response: 57384118
Conc: 505.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.488 min
Delta R.T.: 0.001 min
Response: 86084854
Conc: 470.19 ng/ml




#33 AR-1260-3

R.T.: 7.796 min
Delta R.T.: 0.002 min
Response: 42869096
Conc: 500.32 ng/ml



#33 AR-1260-3


R.T.: 6.644 min
Delta R.T.: 0.000 min
Response: 83805651
Conc: 465.95 ng/ml



#34 AR-1260-4


R.T.: 8.022 min
Delta R.T.: 0.002 min
Response: 47291272
Conc: 488.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.122 min
Delta R.T.: 0.000 min
Response: 68591474
Conc: 466.16 ng/ml



#35 AR-1260-5

R.T.: 8.341 min
Delta R.T.: 0.002 min
Response: 84728712
Conc: 475.36 ng/ml



#35 AR-1260-5

R.T.: 7.365 min
Delta R.T.: 0.000 min
Response: 153351834
Conc: 476.07 ng/ml