

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122623\
 Data File : PP062621.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Dec 2023 19:19
 Operator : YP\AJ
 Sample : 06007-01
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 E8965C-ERM-001

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 00:56:57 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 22 03:29:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.396	3.683	48732621	65210940	22.415	21.852
2) SA Decachlor...	10.148	8.782	27166702	43932784	18.840	17.258

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122623\
Data File : PP062621.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2023 19:19
Operator : YP\AJ
Sample : 06007-01
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
E8965C-ERM-001

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 00:56:57 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 22 03:29:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.396 min
Delta R.T.: 0.000 min
Response: 48732621
Conc: 22.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
E8965C-ERM-001




#1 Tetrachloro-m-xylene

R.T.: 3.683 min
Delta R.T.: 0.000 min
Response: 65210940
Conc: 21.85 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.148 min
Delta R.T.: 0.006 min
Response: 27166702
Conc: 18.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.782 min
Delta R.T.: 0.000 min
Response: 43932784
Conc: 17.26 ng/ml