

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122722\
 Data File : PP054078.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Dec 2022 14:52
 Operator : YP\AJ
 Sample : N6197-05 10X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LAW-2022-61A

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/28/2022
 Supervised By :Ankita Jodhani 12/28/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 28 03:06:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 14 05:38:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.338	3.588	3726208	2264824	1.775	1.450
2) SA Decachlor...	10.098	8.618	3661421	2519697	2.157m	1.804

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122722\
Data File : PP054078.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Dec 2022 14:52
Operator : YP\AJ
Sample : N6197-05 10X
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
LAW-2022-61A


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/28/2022
Supervised By :Ankita Jodhani 12/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 28 03:06:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP121322.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 14 05:38:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.338 min
Delta R.T.: -0.001 min
Response: 3726208
Conc: 1.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-2022-61A


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/28/2022
Supervised By :Ankita Jodhani 12/28/2022




#1 Tetrachloro-m-xylene

R.T.: 3.588 min
Delta R.T.: -0.006 min
Response: 2264824
Conc: 1.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.098 min
Delta R.T.: -0.004 min
Response: 3661421
Conc: 2.16 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.618 min
Delta R.T.: -0.017 min
Response: 2519697
Conc: 1.80 ng/ml