

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122820\
 Data File : PP032470.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Dec 2020 00:59
 Operator : DD\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 29 08:41:36 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 29 08:36:39 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.753	4.036	2925862	2512228	60.447	53.180
2) SA Decachlor...	10.591	9.316	2469208	2414578	51.016	55.344
Target Compounds						
21) L5 AR-1248-1	6.054	5.284	537452	473484	530.915	507.248
22) L5 AR-1248-2	6.347	5.546	697952	633877	532.624	506.485
23) L5 AR-1248-3	6.561	5.590	868182	679028	539.736	507.061
24) L5 AR-1248-4	6.986	5.774	973800	824940	526.397	508.221
25) L5 AR-1248-5	7.025	6.196	963541	812466	525.964	495.295



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122820\
Data File : PP032470.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Dec 2020 00:59
Operator : DD\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 29 08:41:36 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122820.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 29 08:36:39 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.753 min
Delta R.T.: -0.001 min
Response: 2925862
Conc: 60.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.036 min
Delta R.T.: 0.003 min
Response: 2512228
Conc: 53.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.591 min
Delta R.T.: 0.000 min
Response: 2469208
Conc: 51.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.316 min
Delta R.T.: 0.001 min
Response: 2414578
Conc: 55.34 ng/ml



#21 AR-1248-1


R.T.: 6.054 min
Delta R.T.: -0.002 min
Response: 537452
Conc: 530.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.284 min
Delta R.T.: 0.003 min
Response: 473484
Conc: 507.25 ng/ml




#22 AR-1248-2

R.T.: 6.347 min
Delta R.T.: -0.001 min
Response: 697952
Conc: 532.62 ng/ml



#22 AR-1248-2


R.T.: 5.546 min
Delta R.T.: 0.003 min
Response: 633877
Conc: 506.49 ng/ml



#23 AR-1248-3


R.T.: 6.561 min
Delta R.T.: -0.001 min
Response: 868182
Conc: 539.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.590 min
Delta R.T.: 0.003 min
Response: 679028
Conc: 507.06 ng/ml




#24 AR-1248-4

R.T.: 6.986 min
Delta R.T.: -0.001 min
Response: 973800
Conc: 526.40 ng/ml



#24 AR-1248-4


R.T.: 5.774 min
Delta R.T.: 0.003 min
Response: 824940
Conc: 508.22 ng/ml



#25 AR-1248-5

R.T.: 7.025 min
Delta R.T.: -0.001 min
Response: 963541
Conc: 525.96 ng/ml

Instrument :
ECD_P
ClientSampleId :



#25 AR-1248-5

R.T.: 6.196 min
Delta R.T.: 0.004 min
Response: 812466
Conc: 495.30 ng/ml