

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122820\
 Data File : PP032471.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Dec 2020 1:15
 Operator : DD\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 29 08:42:12 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 29 08:36:39 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.752	4.035	2997606	2563392	61.929	54.263
2) SA Decachlor...	10.591	9.316	2524467	2461045	52.157	56.409
Target Compounds						
26) L6 AR-1254-1	6.957	6.152	985581	1246580	532.849	507.902
27) L6 AR-1254-2	7.185	6.311	1528375	1073781	529.660	504.509
28) L6 AR-1254-3	7.564	6.730	1669877	1804913	532.810	504.264
29) L6 AR-1254-4	7.857	6.970	1282541	1075211	526.197	504.602
30) L6 AR-1254-5	8.284	7.398	1355998	1610791	533.698	503.592



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122820\
Data File : PP032471.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Dec 2020 1:15
Operator : DD\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 29 08:42:12 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122820.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 29 08:36:39 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.752 min
Delta R.T.: -0.002 min
Response: 2997606
Conc: 61.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.035 min
Delta R.T.: 0.002 min
Response: 2563392
Conc: 54.26 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.591 min
Delta R.T.: 0.000 min
Response: 2524467
Conc: 52.16 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.316 min
Delta R.T.: 0.002 min
Response: 2461045
Conc: 56.41 ng/ml



#26 AR-1254-1


R.T.: 6.957 min
Delta R.T.: -0.002 min
Response: 985581
Conc: 532.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#26 AR-1254-1

R.T.: 6.152 min
Delta R.T.: 0.002 min
Response: 1246580
Conc: 507.90 ng/ml




#27 AR-1254-2

R.T.: 7.185 min
Delta R.T.: -0.002 min
Response: 1528375
Conc: 529.66 ng/ml



#27 AR-1254-2


R.T.: 6.311 min
Delta R.T.: 0.003 min
Response: 1073781
Conc: 504.51 ng/ml



#28 AR-1254-3


R.T.: 7.564 min
Delta R.T.: -0.001 min
Response: 1669877
Conc: 532.81 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 6.730 min
Delta R.T.: 0.003 min
Response: 1804913
Conc: 504.26 ng/ml




#29 AR-1254-4

R.T.: 7.857 min
Delta R.T.: -0.003 min
Response: 1282541
Conc: 526.20 ng/ml



#29 AR-1254-4


R.T.: 6.970 min
Delta R.T.: 0.003 min
Response: 1075211
Conc: 504.60 ng/ml



#30 AR-1254-5

R.T.: 8.284 min
Delta R.T.: -0.002 min
Response: 1355998
Conc: 533.70 ng/ml

Instrument :
ECD_P
ClientSampleId :



#30 AR-1254-5

R.T.: 7.398 min
Delta R.T.: 0.003 min
Response: 1610791
Conc: 503.59 ng/ml