

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122821\
 Data File : PP042318.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Dec 2021 4:37
 Operator : AJ\MA
 Sample : M5244-37
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 29 11:39:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 24 23:47:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.921	4.058	285433	430537	21.739	18.851
2)	SA Decachlor...	10.919	9.409	345308	332869	28.350	20.584 #
Target Compounds							
3)	L1 AR-1016-1	6.231	5.327	139576	215635	314.007	235.232 #
4)	L1 AR-1016-2	6.255	5.348	194142	408070	293.501	323.867
5)	L1 AR-1016-3	6.321	5.541	105779	185610	259.332	264.292
6)	L1 AR-1016-4	6.426	5.592	90840	365555	278.839	673.172 #
7)	L1 AR-1016-5	6.745	5.824	99874	222769	291.010	341.666
8)	L2 AR-1221-1	5.155	4.315	6573	12020	41.481	46.412
9)	L2 AR-1221-2	5.264	4.418	8976	16062	79.609	75.666
10)	L2 AR-1221-3	5.349	4.509	36238	101181	103.591	145.473 #
11)	L3 AR-1232-1	5.349	4.509	36238	101181	125.452	173.061 #
12)	L3 AR-1232-2	5.937	5.348	113542	408070	765.914	766.879
13)	L3 AR-1232-3	6.255	5.541	194142	185610	742.028	648.468
14)	L3 AR-1232-4	6.426	5.637	90840	283303	693.071	1174.552 #
15)	L3 AR-1232-5	6.528	5.824	177170	222769	1716.930	917.469 #
16)	L4 AR-1242-1	6.231	5.327	139576	215635	428.053	316.685 #
17)	L4 AR-1242-2	6.255	5.348	194142	408070	406.777	438.152
18)	L4 AR-1242-3	6.321	5.541	105779	185610	353.176	351.231
19)	L4 AR-1242-4	6.426	5.637	90840	283303	370.767	559.930 #
20)	L4 AR-1242-5	7.210	6.207	382493	988503	1470.617	1667.520
21)	L5 AR-1248-1	6.231	5.327	139576	215635	557.513	415.069 #
22)	L5 AR-1248-2	6.528	5.592	177170	365555	501.642	519.105
23)	L5 AR-1248-3	6.745	5.637	99874	283303	237.490	381.787 #
24)	L5 AR-1248-4	7.170	5.824	326264	222769	719.577	264.610 #
25)	L5 AR-1248-5	7.210	6.250	382493	431343	857.854	520.742 #
26)	L6 AR-1254-1	7.148	6.207	404259	988503	810.905	790.310
27)	L6 AR-1254-2	7.377	6.364	705525	317622	920.283	291.869 #
28)	L6 AR-1254-3	7.760	6.791	451310	336470	584.921	202.529 #
29)	L6 AR-1254-4	8.049	7.030	191687	290094	344.518	297.370
30)	L6 AR-1254-5	8.502f	7.462	2671078	4040039	4265.559	3152.782 #
31)	L7 AR-1260-1	7.919	6.926	384308	263512	648.480	247.955 #
32)	L7 AR-1260-2	8.182	7.138	1424827	1510947	1845.357	1223.950 #
33)	L7 AR-1260-3	8.568	7.281	218759	227124	422.879	169.497 #
34)	L7 AR-1260-4	8.785	7.767	150193	177852	259.391	199.172
35)	L7 AR-1260-5	9.109	8.013	271159	423322	231.743	219.961
36)	L8 AR-1262-1	8.568	7.462	218759	4040039	285.067	6072.015 #
37)	L8 AR-1262-2	9.109	7.767	271159	177852	221.237	163.976 #
38)	L8 AR-1262-3	9.430	8.303	238929	163163	486.080	214.164 #
39)	L8 AR-1262-4	9.492f	8.366	167035	381841	453.222	250.671 #
40)	L8 AR-1262-5	10.169	8.867	68658	81116	149.902	113.447
41)	L9 AR-1268-1	9.430	8.303	238929	163163	157.589	72.104 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122821\
 Data File : PP042318.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Dec 2021 4:37
 Operator : AJ\MA
 Sample : M5244-37
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 29 11:39:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 24 23:47:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.492f	8.366	167035	381841	117.710	174.899 #
43) L9	AR-1268-3	9.744	8.565f	5585	49798	4.340	26.975 #
44) L9	AR-1268-4	10.169	8.867	68658	81116	132.493	102.589
45) L9	AR-1268-5	10.582	9.156	65899	95650	16.994	16.759



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122821\
Data File : PP042318.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Dec 2021 4:37
Operator : AJ\MA
Sample : M5244-37
Misc :
ALS Vial : 50 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 29 11:39:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 24 23:47:53 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.921 min
Delta R.T.: -0.003 min
Response: 285433
Conc: 21.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.058 min
Delta R.T.: -0.006 min
Response: 430537
Conc: 18.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.919 min
Delta R.T.: 0.000 min
Response: 345308
Conc: 28.35 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.409 min
Delta R.T.: -0.009 min
Response: 332869
Conc: 20.58 ng/ml



#3 AR-1016-1


R.T.: 6.231 min
Delta R.T.: -0.002 min
Response: 139576
Conc: 314.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.327 min
Delta R.T.: -0.006 min
Response: 215635
Conc: 235.23 ng/ml




#4 AR-1016-2

R.T.: 6.255 min
Delta R.T.: -0.002 min
Response: 194142
Conc: 293.50 ng/ml



#4 AR-1016-2


R.T.: 5.348 min
Delta R.T.: -0.005 min
Response: 408070
Conc: 323.87 ng/ml



#5 AR-1016-3


R.T.: 6.321 min
Delta R.T.: -0.001 min
Response: 105779
Conc: 259.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.541 min
Delta R.T.: -0.007 min
Response: 185610
Conc: 264.29 ng/ml




#6 AR-1016-4

R.T.: 6.426 min
Delta R.T.: -0.002 min
Response: 90840
Conc: 278.84 ng/ml



#6 AR-1016-4


R.T.: 5.592 min
Delta R.T.: -0.007 min
Response: 365555
Conc: 673.17 ng/ml



#7 AR-1016-5


R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 99874
Conc: 291.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.824 min
Delta R.T.: -0.006 min
Response: 222769
Conc: 341.67 ng/ml




#8 AR-1221-1

R.T.: 5.155 min
Delta R.T.: -0.011 min
Response: 6573
Conc: 41.48 ng/ml



#8 AR-1221-1


R.T.: 4.315 min
Delta R.T.: -0.010 min
Response: 12020
Conc: 46.41 ng/ml



#9 AR-1221-2


R.T.: 5.264 min
Delta R.T.: -0.001 min
Response: 8976
Conc: 79.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.418 min
Delta R.T.: -0.007 min
Response: 16062
Conc: 75.67 ng/ml




#10 AR-1221-3

R.T.: 5.349 min
Delta R.T.: -0.002 min
Response: 36238
Conc: 103.59 ng/ml



#10 AR-1221-3


R.T.: 4.509 min
Delta R.T.: -0.006 min
Response: 101181
Conc: 145.47 ng/ml



#11 AR-1232-1


R.T.: 5.349 min
Delta R.T.: -0.002 min
Response: 36238
Conc: 125.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.509 min
Delta R.T.: -0.006 min
Response: 101181
Conc: 173.06 ng/ml




#12 AR-1232-2

R.T.: 5.937 min
Delta R.T.: -0.003 min
Response: 113542
Conc: 765.91 ng/ml



#12 AR-1232-2


R.T.: 5.348 min
Delta R.T.: -0.006 min
Response: 408070
Conc: 766.88 ng/ml



#13 AR-1232-3


R.T.: 6.255 min
Delta R.T.: -0.002 min
Response: 194142
Conc: 742.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.541 min
Delta R.T.: -0.006 min
Response: 185610
Conc: 648.47 ng/ml




#14 AR-1232-4

R.T.: 6.426 min
Delta R.T.: -0.003 min
Response: 90840
Conc: 693.07 ng/ml



#14 AR-1232-4


R.T.: 5.637 min
Delta R.T.: -0.006 min
Response: 283303
Conc: 1174.55 ng/ml



#15 AR-1232-5


R.T.: 6.528 min
Delta R.T.: -0.002 min
Response: 177170
Conc: 1716.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.824 min
Delta R.T.: -0.006 min
Response: 222769
Conc: 917.47 ng/ml




#16 AR-1242-1

R.T.: 6.231 min
Delta R.T.: -0.001 min
Response: 139576
Conc: 428.05 ng/ml



#16 AR-1242-1


R.T.: 5.327 min
Delta R.T.: -0.005 min
Response: 215635
Conc: 316.68 ng/ml



#17 AR-1242-2


R.T.: 6.255 min
Delta R.T.: -0.001 min
Response: 194142
Conc: 406.78 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.348 min
Delta R.T.: -0.005 min
Response: 408070
Conc: 438.15 ng/ml




#18 AR-1242-3

R.T.: 6.321 min
Delta R.T.: -0.001 min
Response: 105779
Conc: 353.18 ng/ml



#18 AR-1242-3


R.T.: 5.541 min
Delta R.T.: -0.006 min
Response: 185610
Conc: 351.23 ng/ml



#19 AR-1242-4


R.T.: 6.426 min
Delta R.T.: -0.001 min
Response: 90840
Conc: 370.77 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.637 min
Delta R.T.: -0.005 min
Response: 283303
Conc: 559.93 ng/ml




#20 AR-1242-5

R.T.: 7.210 min
Delta R.T.: 0.000 min
Response: 382493
Conc: 1470.62 ng/ml



#20 AR-1242-5


R.T.: 6.207 min
Delta R.T.: -0.003 min
Response: 988503
Conc: 1667.52 ng/ml



#21 AR-1248-1


R.T.: 6.231 min
Delta R.T.: -0.002 min
Response: 139576
Conc: 557.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.327 min
Delta R.T.: -0.005 min
Response: 215635
Conc: 415.07 ng/ml




#22 AR-1248-2

R.T.: 6.528 min
Delta R.T.: -0.002 min
Response: 177170
Conc: 501.64 ng/ml



#22 AR-1248-2


R.T.: 5.592 min
Delta R.T.: -0.006 min
Response: 365555
Conc: 519.11 ng/ml



#23 AR-1248-3


R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 99874
Conc: 237.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.637 min
Delta R.T.: -0.005 min
Response: 283303
Conc: 381.79 ng/ml




#24 AR-1248-4

R.T.: 7.170 min
Delta R.T.: -0.002 min
Response: 326264
Conc: 719.58 ng/ml



#24 AR-1248-4


R.T.: 5.824 min
Delta R.T.: -0.005 min
Response: 222769
Conc: 264.61 ng/ml



#25 AR-1248-5


R.T.: 7.210 min
Delta R.T.: -0.001 min
Response: 382493
Conc: 857.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.250 min
Delta R.T.: -0.005 min
Response: 431343
Conc: 520.74 ng/ml




#26 AR-1254-1

R.T.: 7.148 min
Delta R.T.: 0.004 min
Response: 404259
Conc: 810.90 ng/ml



#26 AR-1254-1


R.T.: 6.207 min
Delta R.T.: -0.004 min
Response: 988503
Conc: 790.31 ng/ml



#27 AR-1254-2


R.T.: 7.377 min
Delta R.T.: 0.003 min
Response: 705525
Conc: 920.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.364 min
Delta R.T.: -0.006 min
Response: 317622
Conc: 291.87 ng/ml




#28 AR-1254-3

R.T.: 7.760 min
Delta R.T.: 0.007 min
Response: 451310
Conc: 584.92 ng/ml



#28 AR-1254-3


R.T.: 6.791 min
Delta R.T.: -0.005 min
Response: 336470
Conc: 202.53 ng/ml



#29 AR-1254-4


R.T.: 8.049 min
Delta R.T.: 0.000 min
Response: 191687
Conc: 344.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 7.030 min
Delta R.T.: -0.005 min
Response: 290094
Conc: 297.37 ng/ml




#30 AR-1254-5

R.T.: 8.502 min
Delta R.T.: 0.023 min
Response: 2671078
Conc: 4265.56 ng/ml



#30 AR-1254-5


R.T.: 7.462 min
Delta R.T.: -0.005 min
Response: 4040039
Conc: 3152.78 ng/ml



#31 AR-1260-1


R.T.: 7.919 min
Delta R.T.: -0.005 min
Response: 384308
Conc: 648.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.926 min
Delta R.T.: -0.007 min
Response: 263512
Conc: 247.96 ng/ml




#32 AR-1260-2

R.T.: 8.182 min
Delta R.T.: -0.007 min
Response: 1424827
Conc: 1845.36 ng/ml



#32 AR-1260-2


R.T.: 7.138 min
Delta R.T.: 0.008 min
Response: 1510947
Conc: 1223.95 ng/ml



#33 AR-1260-3


R.T.: 8.568 min
Delta R.T.: 0.010 min
Response: 218759
Conc: 422.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.281 min
Delta R.T.: -0.007 min
Response: 227124
Conc: 169.50 ng/ml




#34 AR-1260-4

R.T.: 8.785 min
Delta R.T.: -0.001 min
Response: 150193
Conc: 259.39 ng/ml



#34 AR-1260-4


R.T.: 7.767 min
Delta R.T.: -0.007 min
Response: 177852
Conc: 199.17 ng/ml



#35 AR-1260-5


R.T.: 9.109 min
Delta R.T.: 0.000 min
Response: 271159
Conc: 231.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 8.013 min
Delta R.T.: -0.008 min
Response: 423322
Conc: 219.96 ng/ml




#36 AR-1262-1

R.T.: 8.568 min
Delta R.T.: 0.011 min
Response: 218759
Conc: 285.07 ng/ml



#36 AR-1262-1


R.T.: 7.462 min
Delta R.T.: -0.003 min
Response: 4040039
Conc: 6072.02 ng/ml



#37 AR-1262-2


R.T.: 9.109 min
Delta R.T.: 0.000 min
Response: 271159
Conc: 221.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.767 min
Delta R.T.: -0.006 min
Response: 177852
Conc: 163.98 ng/ml




#38 AR-1262-3

R.T.: 9.430 min
Delta R.T.: 0.007 min
Response: 238929
Conc: 486.08 ng/ml



#38 AR-1262-3


R.T.: 8.303 min
Delta R.T.: -0.006 min
Response: 163163
Conc: 214.16 ng/ml



#39 AR-1262-4


R.T.: 9.492 min
Delta R.T.: -0.020 min
Response: 167035
Conc: 453.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.366 min
Delta R.T.: -0.007 min
Response: 381841
Conc: 250.67 ng/ml




#40 AR-1262-5

R.T.: 10.169 min
Delta R.T.: 0.002 min
Response: 68658
Conc: 149.90 ng/ml



#40 AR-1262-5


R.T.: 8.867 min
Delta R.T.: -0.005 min
Response: 81116
Conc: 113.45 ng/ml



#41 AR-1268-1


R.T.: 9.430 min
Delta R.T.: 0.009 min
Response: 238929
Conc: 157.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.303 min
Delta R.T.: -0.006 min
Response: 163163
Conc: 72.10 ng/ml




#42 AR-1268-2

R.T.: 9.492 min
Delta R.T.: -0.024 min
Response: 167035
Conc: 117.71 ng/ml



#42 AR-1268-2


R.T.: 8.366 min
Delta R.T.: -0.008 min
Response: 381841
Conc: 174.90 ng/ml



#43 AR-1268-3


R.T.: 9.744 min
Delta R.T.: 0.000 min
Response: 5585
Conc: 4.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.565 min
Delta R.T.: -0.019 min
Response: 49798
Conc: 26.98 ng/ml




#44 AR-1268-4

R.T.: 10.169 min
Delta R.T.: 0.000 min
Response: 68658
Conc: 132.49 ng/ml



#44 AR-1268-4


R.T.: 8.867 min
Delta R.T.: -0.005 min
Response: 81116
Conc: 102.59 ng/ml



#45 AR-1268-5

R.T.: 10.582 min
Delta R.T.: 0.000 min
Response: 65899
Conc: 16.99 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.156 min
Delta R.T.: -0.006 min
Response: 95650
Conc: 16.76 ng/ml