

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122921\
 Data File : PP042386.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Dec 2021 5:31
 Operator : AJ\MA
 Sample : M5244-45
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 30 12:50:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 24 23:47:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.918	4.055	213032	385566	16.225	16.882
2)	SA Decachlor...	10.915	9.400	258946	278092	21.260	17.196
Target Compounds							
3)	L1 AR-1016-1	6.228	5.343	66266	190680	149.081	208.010 #
4)	L1 AR-1016-2	6.251	5.343	89709	190680	135.621	151.334
5)	L1 AR-1016-3	6.319	5.537	61515	122949	150.814	175.069
6)	L1 AR-1016-4	6.424	5.588	60342	385933	185.225	710.698 #
7)	L1 AR-1016-5	6.741	5.819	104056	212993	303.194	326.673
8)	L2 AR-1221-1	0.000	4.308f	0	10967	N.D.	42.346 #
9)	L2 AR-1221-2	5.263	0.000	5449	0	48.329	N.D. #
10)	L2 AR-1221-3	5.367f	4.506	47781	78563	136.589	112.954
11)	L3 AR-1232-1	5.367f	4.506	47781	78563	165.413	134.375
12)	L3 AR-1232-2	5.933	5.343	33232	190680	224.173	358.342 #
13)	L3 AR-1232-3	6.251	5.537	89709	122949	342.876	429.550 #
14)	L3 AR-1232-4	6.424	5.632	60342	199528	460.388	827.228 #
15)	L3 AR-1232-5	6.525	5.819	180839	212993	1752.486	877.209 #
16)	L4 AR-1242-1	6.228	5.323	66266	107029	203.227	157.185
17)	L4 AR-1242-2	6.251	5.343	89709	190680	187.963	204.737
18)	L4 AR-1242-3	6.319	5.537	61515	122949	205.389	232.658
19)	L4 AR-1242-4	6.424	5.632	60342	199528	246.290	394.355 #
20)	L4 AR-1242-5	7.207	6.204	235582	846087	905.770	1427.276 #
21)	L5 AR-1248-1	6.228	5.323	66266	107029	264.690	206.018
22)	L5 AR-1248-2	6.525	5.588	180839	385933	512.031	548.043
23)	L5 AR-1248-3	6.741	5.632	104056	199528	247.433	268.889
24)	L5 AR-1248-4	7.146f	5.819	524233	212993	1156.201	252.998 #
25)	L5 AR-1248-5	7.207	6.244	235582	266578	528.362	321.829 #
26)	L6 AR-1254-1	7.146	6.204	524233	846087	1051.562	676.448 #
27)	L6 AR-1254-2	7.375	6.358	492746	243294	642.735	223.567 #
28)	L6 AR-1254-3	7.755	6.804	328709	152055	426.025	91.525 #
29)	L6 AR-1254-4	8.046	7.025	111624	209482	200.621	214.735
30)	L6 AR-1254-5	8.498f	7.455	596399	937174	952.415	731.356
31)	L7 AR-1260-1	7.921	6.921	125497	251411	211.764	236.568
32)	L7 AR-1260-2	8.181	7.118	623702	424097	807.784	343.542 #
33)	L7 AR-1260-3	8.556	7.275	111322	209942	215.195	156.674 #
34)	L7 AR-1260-4	8.783	7.760	136770	145013	236.210	162.396 #
35)	L7 AR-1260-5	9.106	8.007	143231	242812	122.411	126.167
36)	L8 AR-1262-1	8.556	7.455	111322	937174	145.065	1408.535 #
37)	L8 AR-1262-2	9.106	7.760	143231	145013	116.862	133.699
38)	L8 AR-1262-3	9.429	8.296	76881	51303	156.407	67.340 #
39)	L8 AR-1262-4	9.512	8.360	42267	144911	114.685	95.131
40)	L8 AR-1262-5	10.165	8.860	18451	15064	40.283	21.068 #
41)	L9 AR-1268-1	9.429	8.296	76881	51303	50.708	22.672 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122921\
 Data File : PP042386.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Dec 2021 5:31
 Operator : AJ\MA
 Sample : M5244-45
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 30 12:50:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 24 23:47:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.512	8.360	42267	144911	29.786	66.375 #
43) L9	AR-1268-3	9.743	0.000	4513	0	3.507	N.D. #
44) L9	AR-1268-4	10.165	8.860	18451	15064	35.605	19.052 #
45) L9	AR-1268-5	10.578	9.149	15811	17101	4.077	2.996 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122921\
Data File : PP042386.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Dec 2021 5:31
Operator : AJ\MA
Sample : M5244-45
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 30 12:50:38 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 24 23:47:53 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.918 min
Delta R.T.: -0.006 min
Response: 213032
Conc: 16.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.055 min
Delta R.T.: -0.008 min
Response: 385566
Conc: 16.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.915 min
Delta R.T.: -0.003 min
Response: 258946
Conc: 21.26 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.400 min
Delta R.T.: -0.010 min
Response: 278092
Conc: 17.20 ng/ml



#3 AR-1016-1


R.T.: 6.228 min
Delta R.T.: -0.005 min
Response: 66266
Conc: 149.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.343 min
Delta R.T.: 0.010 min
Response: 190680
Conc: 208.01 ng/ml




#4 AR-1016-2

R.T.: 6.251 min
Delta R.T.: -0.006 min
Response: 89709
Conc: 135.62 ng/ml



#4 AR-1016-2


R.T.: 5.343 min
Delta R.T.: -0.010 min
Response: 190680
Conc: 151.33 ng/ml



#5 AR-1016-3


R.T.: 6.319 min
Delta R.T.: -0.003 min
Response: 61515
Conc: 150.81 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.537 min
Delta R.T.: -0.011 min
Response: 122949
Conc: 175.07 ng/ml




#6 AR-1016-4

R.T.: 6.424 min
Delta R.T.: -0.005 min
Response: 60342
Conc: 185.23 ng/ml



#6 AR-1016-4


R.T.: 5.588 min
Delta R.T.: -0.011 min
Response: 385933
Conc: 710.70 ng/ml



#7 AR-1016-5


R.T.: 6.741 min
Delta R.T.: -0.004 min
Response: 104056
Conc: 303.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.819 min
Delta R.T.: -0.011 min
Response: 212993
Conc: 326.67 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 5.166 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 4.308 min
Delta R.T.: -0.017 min
Response: 10967
Conc: 42.35 ng/ml



#9 AR-1221-2


R.T.: 5.263 min
Delta R.T.: -0.002 min
Response: 5449
Conc: 48.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 4.425 min
Response: 0
Conc: N.D.




#10 AR-1221-3

R.T.: 5.367 min
Delta R.T.: 0.016 min
Response: 47781
Conc: 136.59 ng/ml



#10 AR-1221-3


R.T.: 4.506 min
Delta R.T.: -0.009 min
Response: 78563
Conc: 112.95 ng/ml



#11 AR-1232-1


R.T.: 5.367 min
Delta R.T.: 0.016 min
Response: 47781
Conc: 165.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.506 min
Delta R.T.: -0.009 min
Response: 78563
Conc: 134.37 ng/ml




#12 AR-1232-2

R.T.: 5.933 min
Delta R.T.: -0.007 min
Response: 33232
Conc: 224.17 ng/ml



#12 AR-1232-2


R.T.: 5.343 min
Delta R.T.: -0.010 min
Response: 190680
Conc: 358.34 ng/ml



#13 AR-1232-3


R.T.: 6.251 min
Delta R.T.: -0.006 min
Response: 89709
Conc: 342.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.537 min
Delta R.T.: -0.011 min
Response: 122949
Conc: 429.55 ng/ml




#14 AR-1232-4

R.T.: 6.424 min
Delta R.T.: -0.006 min
Response: 60342
Conc: 460.39 ng/ml



#14 AR-1232-4


R.T.: 5.632 min
Delta R.T.: -0.011 min
Response: 199528
Conc: 827.23 ng/ml



#15 AR-1232-5


R.T.: 6.525 min
Delta R.T.: -0.005 min
Response: 180839
Conc: 1752.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.819 min
Delta R.T.: -0.011 min
Response: 212993
Conc: 877.21 ng/ml




#16 AR-1242-1

R.T.: 6.228 min
Delta R.T.: -0.004 min
Response: 66266
Conc: 203.23 ng/ml



#16 AR-1242-1


R.T.: 5.323 min
Delta R.T.: -0.009 min
Response: 107029
Conc: 157.18 ng/ml



#17 AR-1242-2


R.T.: 6.251 min
Delta R.T.: -0.005 min
Response: 89709
Conc: 187.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.343 min
Delta R.T.: -0.009 min
Response: 190680
Conc: 204.74 ng/ml




#18 AR-1242-3

R.T.: 6.319 min
Delta R.T.: -0.003 min
Response: 61515
Conc: 205.39 ng/ml



#18 AR-1242-3


R.T.: 5.537 min
Delta R.T.: -0.010 min
Response: 122949
Conc: 232.66 ng/ml



#19 AR-1242-4


R.T.: 6.424 min
Delta R.T.: -0.004 min
Response: 60342
Conc: 246.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.632 min
Delta R.T.: -0.010 min
Response: 199528
Conc: 394.35 ng/ml




#20 AR-1242-5

R.T.: 7.207 min
Delta R.T.: -0.003 min
Response: 235582
Conc: 905.77 ng/ml



#20 AR-1242-5


R.T.: 6.204 min
Delta R.T.: -0.007 min
Response: 846087
Conc: 1427.28 ng/ml



#21 AR-1248-1


R.T.: 6.228 min
Delta R.T.: -0.005 min
Response: 66266
Conc: 264.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.323 min
Delta R.T.: -0.010 min
Response: 107029
Conc: 206.02 ng/ml




#22 AR-1248-2

R.T.: 6.525 min
Delta R.T.: -0.005 min
Response: 180839
Conc: 512.03 ng/ml



#22 AR-1248-2


R.T.: 5.588 min
Delta R.T.: -0.010 min
Response: 385933
Conc: 548.04 ng/ml



#23 AR-1248-3


R.T.: 6.741 min
Delta R.T.: -0.004 min
Response: 104056
Conc: 247.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.632 min
Delta R.T.: -0.011 min
Response: 199528
Conc: 268.89 ng/ml




#24 AR-1248-4

R.T.: 7.146 min
Delta R.T.: -0.025 min
Response: 524233
Conc: 1156.20 ng/ml



#24 AR-1248-4


R.T.: 5.819 min
Delta R.T.: -0.011 min
Response: 212993
Conc: 253.00 ng/ml



#25 AR-1248-5


R.T.: 7.207 min
Delta R.T.: -0.004 min
Response: 235582
Conc: 528.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.244 min
Delta R.T.: -0.011 min
Response: 266578
Conc: 321.83 ng/ml




#26 AR-1254-1

R.T.: 7.146 min
Delta R.T.: 0.002 min
Response: 524233
Conc: 1051.56 ng/ml



#26 AR-1254-1


R.T.: 6.204 min
Delta R.T.: -0.008 min
Response: 846087
Conc: 676.45 ng/ml



#27 AR-1254-2


R.T.: 7.375 min
Delta R.T.: 0.001 min
Response: 492746
Conc: 642.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.358 min
Delta R.T.: -0.012 min
Response: 243294
Conc: 223.57 ng/ml




#28 AR-1254-3

R.T.: 7.755 min
Delta R.T.: 0.001 min
Response: 328709
Conc: 426.02 ng/ml



#28 AR-1254-3


R.T.: 6.804 min
Delta R.T.: 0.009 min
Response: 152055
Conc: 91.53 ng/ml



#29 AR-1254-4


R.T.: 8.046 min
Delta R.T.: -0.003 min
Response: 111624
Conc: 200.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 7.025 min
Delta R.T.: -0.010 min
Response: 209482
Conc: 214.74 ng/ml




#30 AR-1254-5

R.T.: 8.498 min
Delta R.T.: 0.019 min
Response: 596399
Conc: 952.42 ng/ml



#30 AR-1254-5


R.T.: 7.455 min
Delta R.T.: -0.012 min
Response: 937174
Conc: 731.36 ng/ml



#31 AR-1260-1


R.T.: 7.921 min
Delta R.T.: -0.002 min
Response: 125497
Conc: 211.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.921 min
Delta R.T.: -0.012 min
Response: 251411
Conc: 236.57 ng/ml




#32 AR-1260-2

R.T.: 8.181 min
Delta R.T.: -0.008 min
Response: 623702
Conc: 807.78 ng/ml



#32 AR-1260-2


R.T.: 7.118 min
Delta R.T.: -0.011 min
Response: 424097
Conc: 343.54 ng/ml



#33 AR-1260-3


R.T.: 8.556 min
Delta R.T.: -0.002 min
Response: 111322
Conc: 215.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.275 min
Delta R.T.: -0.013 min
Response: 209942
Conc: 156.67 ng/ml




#34 AR-1260-4

R.T.: 8.783 min
Delta R.T.: -0.003 min
Response: 136770
Conc: 236.21 ng/ml



#34 AR-1260-4


R.T.: 7.760 min
Delta R.T.: -0.014 min
Response: 145013
Conc: 162.40 ng/ml



#35 AR-1260-5


R.T.: 9.106 min
Delta R.T.: -0.004 min
Response: 143231
Conc: 122.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 8.007 min
Delta R.T.: -0.014 min
Response: 242812
Conc: 126.17 ng/ml




#36 AR-1262-1

R.T.: 8.556 min
Delta R.T.: -0.002 min
Response: 111322
Conc: 145.06 ng/ml



#36 AR-1262-1


R.T.: 7.455 min
Delta R.T.: -0.010 min
Response: 937174
Conc: 1408.54 ng/ml



#37 AR-1262-2


R.T.: 9.106 min
Delta R.T.: -0.003 min
Response: 143231
Conc: 116.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.760 min
Delta R.T.: -0.012 min
Response: 145013
Conc: 133.70 ng/ml




#38 AR-1262-3

R.T.: 9.429 min
Delta R.T.: 0.005 min
Response: 76881
Conc: 156.41 ng/ml



#38 AR-1262-3


R.T.: 8.296 min
Delta R.T.: -0.013 min
Response: 51303
Conc: 67.34 ng/ml



#39 AR-1262-4


R.T.: 9.512 min
Delta R.T.: 0.000 min
Response: 42267
Conc: 114.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.360 min
Delta R.T.: -0.013 min
Response: 144911
Conc: 95.13 ng/ml




#40 AR-1262-5

R.T.: 10.165 min
Delta R.T.: -0.002 min
Response: 18451
Conc: 40.28 ng/ml



#40 AR-1262-5


R.T.: 8.860 min
Delta R.T.: -0.012 min
Response: 15064
Conc: 21.07 ng/ml



#41 AR-1268-1


R.T.: 9.429 min
Delta R.T.: 0.008 min
Response: 76881
Conc: 50.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.296 min
Delta R.T.: -0.013 min
Response: 51303
Conc: 22.67 ng/ml




#42 AR-1268-2

R.T.: 9.512 min
Delta R.T.: -0.004 min
Response: 42267
Conc: 29.79 ng/ml



#42 AR-1268-2


R.T.: 8.360 min
Delta R.T.: -0.014 min
Response: 144911
Conc: 66.37 ng/ml



#43 AR-1268-3


R.T.: 9.743 min
Delta R.T.: 0.000 min
Response: 4513
Conc: 3.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 8.584 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 10.165 min
Delta R.T.: -0.003 min
Response: 18451
Conc: 35.61 ng/ml



#44 AR-1268-4


R.T.: 8.860 min
Delta R.T.: -0.012 min
Response: 15064
Conc: 19.05 ng/ml



#45 AR-1268-5

R.T.: 10.578 min
Delta R.T.: -0.005 min
Response: 15811
Conc: 4.08 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.149 min
Delta R.T.: -0.014 min
Response: 17101
Conc: 3.00 ng/ml