

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122921\
 Data File : PP042346.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Dec 2021 14:22
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 30 00:16:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 24 23:47:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.920	4.057	654471	1180744	49.846	51.697
2)	SA Decachlor...	10.916	9.404	724119	876023	59.450	54.170
Target Compounds							
26)	L6 AR-1254-1	7.142	6.203	245846	608816	493.144	486.750
27)	L6 AR-1254-2	7.372	6.362	376372	531517	490.938	488.421
28)	L6 AR-1254-3	7.751	6.787	404784	806895	524.621	485.689
29)	L6 AR-1254-4	8.047	7.026	288800	482943	519.057	495.055
30)	L6 AR-1254-5	8.478	7.458	327633	634342	523.211	495.030



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122921\
Data File : PP042346.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Dec 2021 14:22
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 30 00:16:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 24 23:47:53 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.920 min
Delta R.T.: -0.004 min
Response: 654471
Conc: 49.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.057 min
Delta R.T.: -0.006 min
Response: 1180744
Conc: 51.70 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.916 min
Delta R.T.: -0.002 min
Response: 724119
Conc: 59.45 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.404 min
Delta R.T.: -0.014 min
Response: 876023
Conc: 54.17 ng/ml



#26 AR-1254-1


R.T.: 7.142 min
Delta R.T.: -0.002 min
Response: 245846
Conc: 493.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.203 min
Delta R.T.: -0.008 min
Response: 608816
Conc: 486.75 ng/ml




#27 AR-1254-2

R.T.: 7.372 min
Delta R.T.: -0.002 min
Response: 376372
Conc: 490.94 ng/ml



#27 AR-1254-2


R.T.: 6.362 min
Delta R.T.: -0.008 min
Response: 531517
Conc: 488.42 ng/ml



#28 AR-1254-3


R.T.: 7.751 min
Delta R.T.: -0.002 min
Response: 404784
Conc: 524.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.787 min
Delta R.T.: -0.008 min
Response: 806895
Conc: 485.69 ng/ml




#29 AR-1254-4

R.T.: 8.047 min
Delta R.T.: -0.003 min
Response: 288800
Conc: 519.06 ng/ml



#29 AR-1254-4


R.T.: 7.026 min
Delta R.T.: -0.009 min
Response: 482943
Conc: 495.05 ng/ml



#30 AR-1254-5

R.T.: 8.478 min
Delta R.T.: -0.001 min
Response: 327633
Conc: 523.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.458 min
Delta R.T.: -0.009 min
Response: 634342
Conc: 495.03 ng/ml