

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122922\
 Data File : PP054144.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Dec 2022 17:46
 Operator : YP\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 30 02:20:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 30 02:19:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.340	3.589	8597794	7012345	4.914	4.535
2) SA Decachlor...	10.101	8.615	6972072	7470726	5.309	5.238
Target Compounds						
3) L1 AR-1016-1	5.511	4.677	3108578	2449380	51.478	50.383
4) L1 AR-1016-2	5.534	4.696	4486324	3347230	50.958	48.687
5) L1 AR-1016-3	5.596	4.872	2853791	1755362	50.976	46.723
6) L1 AR-1016-4	5.695	4.915	2147379	1654815	47.751	49.990
7) L1 AR-1016-5	5.991	5.129	2888683	2033522	58.750	48.942
31) L7 AR-1260-1	7.122	6.166	4866730	4162766	56.091	52.338
32) L7 AR-1260-2	7.380	6.355	6226928	4806563	60.252	52.941
33) L7 AR-1260-3	7.741	6.509	5822468	4381090	76.414	50.219 #
34) L7 AR-1260-4	7.967	6.983	4546267	3150785	57.921	48.467
35) L7 AR-1260-5	8.284	7.225	6802561	6540301	52.094	48.692



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122922\
Data File : PP054144.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Dec 2022 17:46
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 30 02:20:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 30 02:19:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.340 min
Delta R.T.: 0.000 min
Response: 8597794
Conc: 4.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.589 min
Delta R.T.: 0.000 min
Response: 7012345
Conc: 4.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.101 min
Delta R.T.: 0.000 min
Response: 6972072
Conc: 5.31 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.615 min
Delta R.T.: 0.000 min
Response: 7470726
Conc: 5.24 ng/ml



#3 AR-1016-1


R.T.: 5.511 min
Delta R.T.: 0.000 min
Response: 3108578
Conc: 51.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.677 min
Delta R.T.: 0.000 min
Response: 2449380
Conc: 50.38 ng/ml




#4 AR-1016-2

R.T.: 5.534 min
Delta R.T.: 0.000 min
Response: 4486324
Conc: 50.96 ng/ml



#4 AR-1016-2


R.T.: 4.696 min
Delta R.T.: 0.000 min
Response: 3347230
Conc: 48.69 ng/ml



#5 AR-1016-3


R.T.: 5.596 min
Delta R.T.: 0.000 min
Response: 2853791
Conc: 50.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.872 min
Delta R.T.: 0.000 min
Response: 1755362
Conc: 46.72 ng/ml




#6 AR-1016-4

R.T.: 5.695 min
Delta R.T.: 0.000 min
Response: 2147379
Conc: 47.75 ng/ml



#6 AR-1016-4


R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 1654815
Conc: 49.99 ng/ml



#7 AR-1016-5


R.T.: 5.991 min
Delta R.T.: 0.000 min
Response: 2888683
Conc: 58.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.129 min
Delta R.T.: 0.000 min
Response: 2033522
Conc: 48.94 ng/ml




#31 AR-1260-1

R.T.: 7.122 min
Delta R.T.: 0.000 min
Response: 4866730
Conc: 56.09 ng/ml



#31 AR-1260-1


R.T.: 6.166 min
Delta R.T.: 0.000 min
Response: 4162766
Conc: 52.34 ng/ml



#32 AR-1260-2


R.T.: 7.380 min
Delta R.T.: 0.000 min
Response: 6226928
Conc: 60.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#32 AR-1260-2

R.T.: 6.355 min
Delta R.T.: 0.000 min
Response: 4806563
Conc: 52.94 ng/ml




#33 AR-1260-3

R.T.: 7.741 min
Delta R.T.: 0.000 min
Response: 5822468
Conc: 76.41 ng/ml



#33 AR-1260-3


R.T.: 6.509 min
Delta R.T.: 0.000 min
Response: 4381090
Conc: 50.22 ng/ml



#34 AR-1260-4


R.T.: 7.967 min
Delta R.T.: 0.000 min
Response: 4546267
Conc: 57.92 ng/ml

Instrument :
ECD_P
ClientSampleId :




#34 AR-1260-4

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 3150785
Conc: 48.47 ng/ml



#35 AR-1260-5

R.T.: 8.284 min
Delta R.T.: 0.000 min
Response: 6802561
Conc: 52.09 ng/ml



#35 AR-1260-5

R.T.: 7.225 min
Delta R.T.: 0.000 min
Response: 6540301
Conc: 48.69 ng/ml