

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122922\
 Data File : PP054164.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Dec 2022 23:15
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 30 04:56:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 30 04:54:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.340	3.589	135.6E6	119.8E6	74.124	74.753
2) SA Decachlor...	10.099	8.614	155.7E6	179.6E6	72.181	72.526
Target Compounds						
41) L9 AR-1268-1	8.593	7.510	168.2E6	142.6E6	731.225	730.148
42) L9 AR-1268-2	8.687	7.575	150.5E6	130.3E6	726.657	729.977
43) L9 AR-1268-3	8.916	7.783	96408778	122.3E6	706.745	731.395
44) L9 AR-1268-4	9.341	8.069	42815265	45552472	715.552	721.219
45) L9 AR-1268-5	9.758	8.360	319.3E6	375.9E6	734.776	735.595



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122922\
Data File : PP054164.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Dec 2022 23:15
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 30 04:56:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 30 04:54:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.340 min
Delta R.T.: 0.000 min
Response: 135644024
Conc: 74.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.589 min
Delta R.T.: 0.000 min
Response: 119780514
Conc: 74.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.099 min
Delta R.T.: 0.000 min
Response: 155687345
Conc: 72.18 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.614 min
Delta R.T.: 0.000 min
Response: 179621668
Conc: 72.53 ng/ml



#41 AR-1268-1


R.T.: 8.593 min
Delta R.T.: -0.001 min
Response: 168248280
Conc: 731.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.510 min
Delta R.T.: 0.000 min
Response: 142595836
Conc: 730.15 ng/ml




#42 AR-1268-2

R.T.: 8.687 min
Delta R.T.: -0.002 min
Response: 150452344
Conc: 726.66 ng/ml



#42 AR-1268-2


R.T.: 7.575 min
Delta R.T.: 0.000 min
Response: 130254948
Conc: 729.98 ng/ml



#43 AR-1268-3


R.T.: 8.916 min
Delta R.T.: -0.002 min
Response: 96408778
Conc: 706.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.783 min
Delta R.T.: 0.000 min
Response: 122341403
Conc: 731.39 ng/ml




#44 AR-1268-4

R.T.: 9.341 min
Delta R.T.: 0.000 min
Response: 42815265
Conc: 715.55 ng/ml



#44 AR-1268-4


R.T.: 8.069 min
Delta R.T.: 0.000 min
Response: 45552472
Conc: 721.22 ng/ml



#45 AR-1268-5

R.T.: 9.758 min
Delta R.T.: -0.002 min
Response: 319269029
Conc: 734.78 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.360 min
Delta R.T.: 0.000 min
Response: 375889785
Conc: 735.60 ng/ml