

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122922\
 Data File : PP054166.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Dec 2022 23:47
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 30 04:57:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 30 04:54:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.339	3.588	46577053	40057402	25.452	24.999
2) SA Decachlor...	10.098	8.613	57961374	64500656	26.873	26.043
Target Compounds						
41) L9 AR-1268-1	8.593	7.510	61216817	50098233	266.055	256.523
42) L9 AR-1268-2	8.686	7.574	51653943	45500767	249.479	254.996
43) L9 AR-1268-3	8.916	7.782	36092170	42904754	264.581	256.498
44) L9 AR-1268-4	9.340	8.069	15834280	16966387	264.631	268.624
45) L9 AR-1268-5	9.757	8.358	112.0E6	128.6E6	257.647	251.743

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP122922\
Data File : PP054166.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Dec 2022 23:47
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 30 04:57:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 30 04:54:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

