

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP123021\
 Data File : PP042435.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Dec 2021 21:31
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 31 01:31:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 24 23:47:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.915	4.053	603148	1114844	45.937	48.812
2)	SA Decachlor...	10.906	9.393f	691122	816546	56.741	50.493
Target Compounds							
3)	L1 AR-1016-1	6.224	5.339	211073	598323	474.856	652.701 #
4)	L1 AR-1016-2	6.247	5.339	315613	598323	477.138	474.862
5)	L1 AR-1016-3	6.314	5.533	198765	349281	487.301	497.345
6)	L1 AR-1016-4	6.420	5.584	166771	272450	511.915	501.719
7)	L1 AR-1016-5	6.736	5.816	132592	318036	386.343	487.779 #
8)	L2 AR-1221-1	5.153	4.313	23745	39525	149.852	152.615
9)	L2 AR-1221-2	5.257	4.413	32323	63318	286.673	298.287
10)	L2 AR-1221-3	5.342	4.503	119966	246040	342.940	353.743
11)	L3 AR-1232-1	5.342	4.503	119966	246040	415.312	420.829
12)	L3 AR-1232-2	5.930	5.339	164245	598323	1107.935	1124.419
13)	L3 AR-1232-3	6.247	5.533	315613	349281	1206.299	1220.288
14)	L3 AR-1232-4	6.420	5.629	166771	334649	1272.393	1387.427
15)	L3 AR-1232-5	6.521	5.816f	137321	318036	1330.759	1309.824
16)	L4 AR-1242-1	6.224	5.339	211073	598323	647.322	878.708 #
17)	L4 AR-1242-2	6.247	5.339	315613	598323	661.288	642.431
18)	L4 AR-1242-3	6.314	5.533	198765	349281	663.641	660.946
19)	L4 AR-1242-4	6.420	5.629	166771	334649	680.682	661.412
20)	L4 AR-1242-5	7.203	6.197	22934	243433	88.177	410.650 #
21)	L5 AR-1248-1	6.224	5.339	211073	598323	843.096	1151.695 #
22)	L5 AR-1248-2	6.521	5.584	137321	272450	388.813	386.892
23)	L5 AR-1248-3	6.736	5.629	132592	334649	315.290	450.981 #
24)	L5 AR-1248-4	7.162	5.816	24405	318036	53.825	377.770 #
25)	L5 AR-1248-5	7.203	6.240	22934	38708	51.436	46.730
26)	L6 AR-1254-1	7.135	6.197	115027	243433	230.732	194.625
27)	L6 AR-1254-2	7.365	6.355f	131702	221572	171.791	203.606
28)	L6 AR-1254-3	7.745	6.796	77621	415640	100.601	250.184 #
29)	L6 AR-1254-4	8.041	7.051f	53752	117406	96.608	120.350
30)	L6 AR-1254-5	8.469	7.449f	413267	661946	659.963	516.572
31)	L7 AR-1260-1	7.915	6.916f	307788	535097	519.359	503.506
32)	L7 AR-1260-2	8.181	7.112f	351522	620516	455.272	502.652
33)	L7 AR-1260-3	8.550	7.270f	285905	560492	552.678	418.280
34)	L7 AR-1260-4	8.779	7.793f	338932	56157	585.353	62.889 #
35)	L7 AR-1260-5	9.100	8.002f	628136	1024876	536.830	532.534
36)	L8 AR-1262-1	8.550	7.449f	285905	661946	372.566	994.879 #
37)	L8 AR-1262-2	9.100	7.755f	628136	448260	512.493	413.286
38)	L8 AR-1262-3	9.424	8.290f	380915	113454	774.938	148.917 #
39)	L8 AR-1262-4	9.502	8.354f	149561	745237	405.810	489.232
40)	L8 AR-1262-5	10.157	8.898f	187970	-95519	410.396	N.D. #
41)	L9 AR-1268-1	9.424	8.290f	380915	113454	251.238	50.137 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP123021\
 Data File : PP042435.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Dec 2021 21:31
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 31 01:31:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 24 23:47:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.502	8.354f	149561	745237	105.396	341.349 #
43) L9	AR-1268-3	9.734	8.565f	13883	12303	10.787	6.664 #
44) L9	AR-1268-4	10.157	8.898f	187970	-95519	362.734	N.D. #
45) L9	AR-1268-5	10.568f	9.142f	58430	65900	15.068	11.546



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP123021\
Data File : PP042435.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Dec 2021 21:31
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 31 01:31:01 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 24 23:47:53 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.915 min
Delta R.T.: -0.009 min
Response: 603148
Conc: 45.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.053 min
Delta R.T.: -0.011 min
Response: 1114844
Conc: 48.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.906 min
Delta R.T.: -0.012 min
Response: 691122
Conc: 56.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.393 min
Delta R.T.: -0.022 min
Response: 816546
Conc: 50.49 ng/ml



#3 AR-1016-1


R.T.: 6.224 min
Delta R.T.: -0.010 min
Response: 211073
Conc: 474.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.339 min
Delta R.T.: 0.006 min
Response: 598323
Conc: 652.70 ng/ml




#4 AR-1016-2

R.T.: 6.247 min
Delta R.T.: -0.009 min
Response: 315613
Conc: 477.14 ng/ml



#4 AR-1016-2


R.T.: 5.339 min
Delta R.T.: -0.014 min
Response: 598323
Conc: 474.86 ng/ml



#5 AR-1016-3


R.T.: 6.314 min
Delta R.T.: -0.008 min
Response: 198765
Conc: 487.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.533 min
Delta R.T.: -0.014 min
Response: 349281
Conc: 497.35 ng/ml




#6 AR-1016-4

R.T.: 6.420 min
Delta R.T.: -0.009 min
Response: 166771
Conc: 511.91 ng/ml



#6 AR-1016-4


R.T.: 5.584 min
Delta R.T.: -0.015 min
Response: 272450
Conc: 501.72 ng/ml



#7 AR-1016-5


R.T.: 6.736 min
Delta R.T.: -0.009 min
Response: 132592
Conc: 386.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.816 min
Delta R.T.: -0.015 min
Response: 318036
Conc: 487.78 ng/ml




#8 AR-1221-1

R.T.: 5.153 min
Delta R.T.: -0.013 min
Response: 23745
Conc: 149.85 ng/ml



#8 AR-1221-1


R.T.: 4.313 min
Delta R.T.: -0.012 min
Response: 39525
Conc: 152.62 ng/ml



#9 AR-1221-2


R.T.: 5.257 min
Delta R.T.: -0.008 min
Response: 32323
Conc: 286.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.413 min
Delta R.T.: -0.012 min
Response: 63318
Conc: 298.29 ng/ml




#10 AR-1221-3

R.T.: 5.342 min
Delta R.T.: -0.009 min
Response: 119966
Conc: 342.94 ng/ml



#10 AR-1221-3


R.T.: 4.503 min
Delta R.T.: -0.012 min
Response: 246040
Conc: 353.74 ng/ml



#11 AR-1232-1


R.T.: 5.342 min
Delta R.T.: -0.009 min
Response: 119966
Conc: 415.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.503 min
Delta R.T.: -0.012 min
Response: 246040
Conc: 420.83 ng/ml




#12 AR-1232-2

R.T.: 5.930 min
Delta R.T.: -0.010 min
Response: 164245
Conc: 1107.94 ng/ml



#12 AR-1232-2


R.T.: 5.339 min
Delta R.T.: -0.014 min
Response: 598323
Conc: 1124.42 ng/ml



#13 AR-1232-3


R.T.: 6.247 min
Delta R.T.: -0.009 min
Response: 315613
Conc: 1206.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.533 min
Delta R.T.: -0.014 min
Response: 349281
Conc: 1220.29 ng/ml




#14 AR-1232-4

R.T.: 6.420 min
Delta R.T.: -0.010 min
Response: 166771
Conc: 1272.39 ng/ml



#14 AR-1232-4


R.T.: 5.629 min
Delta R.T.: -0.015 min
Response: 334649
Conc: 1387.43 ng/ml



#15 AR-1232-5


R.T.: 6.521 min
Delta R.T.: -0.009 min
Response: 137321
Conc: 1330.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.816 min
Delta R.T.: -0.015 min
Response: 318036
Conc: 1309.82 ng/ml




#16 AR-1242-1

R.T.: 6.224 min
Delta R.T.: -0.008 min
Response: 211073
Conc: 647.32 ng/ml



#16 AR-1242-1


R.T.: 5.339 min
Delta R.T.: 0.008 min
Response: 598323
Conc: 878.71 ng/ml



#17 AR-1242-2


R.T.: 6.247 min
Delta R.T.: -0.008 min
Response: 315613
Conc: 661.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.339 min
Delta R.T.: -0.013 min
Response: 598323
Conc: 642.43 ng/ml




#18 AR-1242-3

R.T.: 6.314 min
Delta R.T.: -0.008 min
Response: 198765
Conc: 663.64 ng/ml



#18 AR-1242-3


R.T.: 5.533 min
Delta R.T.: -0.013 min
Response: 349281
Conc: 660.95 ng/ml



#19 AR-1242-4


R.T.: 6.420 min
Delta R.T.: -0.008 min
Response: 166771
Conc: 680.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.629 min
Delta R.T.: -0.014 min
Response: 334649
Conc: 661.41 ng/ml




#20 AR-1242-5

R.T.: 7.203 min
Delta R.T.: -0.007 min
Response: 22934
Conc: 88.18 ng/ml



#20 AR-1242-5


R.T.: 6.197 min
Delta R.T.: -0.014 min
Response: 243433
Conc: 410.65 ng/ml



#21 AR-1248-1


R.T.: 6.224 min
Delta R.T.: -0.009 min
Response: 211073
Conc: 843.10 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.339 min
Delta R.T.: 0.007 min
Response: 598323
Conc: 1151.70 ng/ml




#22 AR-1248-2

R.T.: 6.521 min
Delta R.T.: -0.009 min
Response: 137321
Conc: 388.81 ng/ml



#22 AR-1248-2


R.T.: 5.584 min
Delta R.T.: -0.014 min
Response: 272450
Conc: 386.89 ng/ml



#23 AR-1248-3


R.T.: 6.736 min
Delta R.T.: -0.009 min
Response: 132592
Conc: 315.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.629 min
Delta R.T.: -0.014 min
Response: 334649
Conc: 450.98 ng/ml




#24 AR-1248-4

R.T.: 7.162 min
Delta R.T.: -0.009 min
Response: 24405
Conc: 53.82 ng/ml



#24 AR-1248-4


R.T.: 5.816 min
Delta R.T.: -0.014 min
Response: 318036
Conc: 377.77 ng/ml



#25 AR-1248-5


R.T.: 7.203 min
Delta R.T.: -0.008 min
Response: 22934
Conc: 51.44 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.240 min
Delta R.T.: -0.015 min
Response: 38708
Conc: 46.73 ng/ml




#26 AR-1254-1

R.T.: 7.135 min
Delta R.T.: -0.009 min
Response: 115027
Conc: 230.73 ng/ml



#26 AR-1254-1


R.T.: 6.197 min
Delta R.T.: -0.015 min
Response: 243433
Conc: 194.63 ng/ml



#27 AR-1254-2


R.T.: 7.365 min
Delta R.T.: -0.009 min
Response: 131702
Conc: 171.79 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.355 min
Delta R.T.: -0.016 min
Response: 221572
Conc: 203.61 ng/ml




#28 AR-1254-3

R.T.: 7.745 min
Delta R.T.: -0.008 min
Response: 77621
Conc: 100.60 ng/ml



#28 AR-1254-3


R.T.: 6.796 min
Delta R.T.: 0.000 min
Response: 415640
Conc: 250.18 ng/ml



#29 AR-1254-4


R.T.: 8.041 min
Delta R.T.: -0.008 min
Response: 53752
Conc: 96.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 7.051 min
Delta R.T.: 0.016 min
Response: 117406
Conc: 120.35 ng/ml




#30 AR-1254-5

R.T.: 8.469 min
Delta R.T.: -0.010 min
Response: 413267
Conc: 659.96 ng/ml



#30 AR-1254-5


R.T.: 7.449 min
Delta R.T.: -0.018 min
Response: 661946
Conc: 516.57 ng/ml



#31 AR-1260-1


R.T.: 7.915 min
Delta R.T.: -0.008 min
Response: 307788
Conc: 519.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.916 min
Delta R.T.: -0.017 min
Response: 535097
Conc: 503.51 ng/ml




#32 AR-1260-2

R.T.: 8.181 min
Delta R.T.: -0.008 min
Response: 351522
Conc: 455.27 ng/ml



#32 AR-1260-2


R.T.: 7.112 min
Delta R.T.: -0.017 min
Response: 620516
Conc: 502.65 ng/ml



#33 AR-1260-3


R.T.: 8.550 min
Delta R.T.: -0.008 min
Response: 285905
Conc: 552.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.270 min
Delta R.T.: -0.018 min
Response: 560492
Conc: 418.28 ng/ml




#34 AR-1260-4

R.T.: 8.779 min
Delta R.T.: -0.007 min
Response: 338932
Conc: 585.35 ng/ml



#34 AR-1260-4


R.T.: 7.793 min
Delta R.T.: 0.019 min
Response: 56157
Conc: 62.89 ng/ml



#35 AR-1260-5


R.T.: 9.100 min
Delta R.T.: -0.009 min
Response: 628136
Conc: 536.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 8.002 min
Delta R.T.: -0.019 min
Response: 1024876
Conc: 532.53 ng/ml




#36 AR-1262-1

R.T.: 8.550 min
Delta R.T.: -0.007 min
Response: 285905
Conc: 372.57 ng/ml



#36 AR-1262-1


R.T.: 7.449 min
Delta R.T.: -0.016 min
Response: 661946
Conc: 994.88 ng/ml



#37 AR-1262-2


R.T.: 9.100 min
Delta R.T.: -0.008 min
Response: 628136
Conc: 512.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.755 min
Delta R.T.: -0.018 min
Response: 448260
Conc: 413.29 ng/ml




#38 AR-1262-3

R.T.: 9.424 min
Delta R.T.: 0.000 min
Response: 380915
Conc: 774.94 ng/ml



#38 AR-1262-3


R.T.: 8.290 min
Delta R.T.: -0.019 min
Response: 113454
Conc: 148.92 ng/ml



#39 AR-1262-4


R.T.: 9.502 min
Delta R.T.: -0.009 min
Response: 149561
Conc: 405.81 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.354 min
Delta R.T.: -0.019 min
Response: 745237
Conc: 489.23 ng/ml




#40 AR-1262-5

R.T.: 10.157 min
Delta R.T.: -0.009 min
Response: 187970
Conc: 410.40 ng/ml



#40 AR-1262-5


R.T.: 8.898 min
Delta R.T.: 0.026 min
Response: -95519
Conc: N.D.



#41 AR-1268-1


R.T.: 9.424 min
Delta R.T.: 0.003 min
Response: 380915
Conc: 251.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.290 min
Delta R.T.: -0.018 min
Response: 113454
Conc: 50.14 ng/ml




#42 AR-1268-2

R.T.: 9.502 min
Delta R.T.: -0.013 min
Response: 149561
Conc: 105.40 ng/ml



#42 AR-1268-2


R.T.: 8.354 min
Delta R.T.: -0.020 min
Response: 745237
Conc: 341.35 ng/ml



#43 AR-1268-3


R.T.: 9.734 min
Delta R.T.: -0.009 min
Response: 13883
Conc: 10.79 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.565 min
Delta R.T.: -0.019 min
Response: 12303
Conc: 6.66 ng/ml




#44 AR-1268-4

R.T.: 10.157 min
Delta R.T.: -0.011 min
Response: 187970
Conc: 362.73 ng/ml



#44 AR-1268-4


R.T.: 8.898 min
Delta R.T.: 0.026 min
Response: -95519
Conc: N.D.



#45 AR-1268-5

R.T.: 10.568 min
Delta R.T.: -0.015 min
Response: 58430
Conc: 15.07 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.142 min
Delta R.T.: -0.021 min
Response: 65900
Conc: 11.55 ng/ml