

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP123021\
 Data File : PP042424.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Dec 2021 17:47
 Operator : AJ\MA
 Sample : M5257-01DL 40X
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 D3-12292021DL

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/03/2022
 Supervised By :mohammad ahmed 01/03/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 01 05:31:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 24 23:47:53 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds								
26)	L6	AR-1254-1	7.140	6.198	299227	662858	600.220	529.956
27)	L6	AR-1254-2	7.369	6.357	477187	634502	622.441	583.056
28)	L6	AR-1254-3	7.749	6.782	474940	973548	615.547	586.001
29)	L6	AR-1254-4	8.044	7.021	332909	535788	598.335	549.225
30)	L6	AR-1254-5	8.474	7.453	512625	884114	818.633m	689.948

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP123021\
Data File : PP042424.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Dec 2021 17:47
Operator : AJ\MA
Sample : M5257-01DL 40X
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
D3-12292021DL


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/03/2022
Supervised By :mohammad ahmed 01/03/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 01 05:31:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 24 23:47:53 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 4.924 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
D3-12292021DL


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 4.063 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 10.918 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.415 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 7.140 min
Delta R.T.: -0.004 min
Response: 299227
Conc: 600.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
D3-12292021DL


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




#26 AR-1254-1

R.T.: 6.198 min
Delta R.T.: -0.013 min
Response: 662858
Conc: 529.96 ng/ml




#27 AR-1254-2

R.T.: 7.369 min
Delta R.T.: -0.005 min
Response: 477187
Conc: 622.44 ng/ml



#27 AR-1254-2

R.T.: 6.357 min
Delta R.T.: -0.013 min
Response: 634502
Conc: 583.06 ng/ml




#28 AR-1254-3

R.T.: 7.749 min
Delta R.T.: -0.005 min
Response: 474940
Conc: 615.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
D3-12292021DL


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/03/2022
Supervised By :mohammad ahmed 01/03/2022




#28 AR-1254-3

R.T.: 6.782 min
Delta R.T.: -0.014 min
Response: 973548
Conc: 586.00 ng/ml




#29 AR-1254-4

R.T.: 8.044 min
Delta R.T.: -0.005 min
Response: 332909
Conc: 598.33 ng/ml



#29 AR-1254-4

R.T.: 7.021 min
Delta R.T.: -0.014 min
Response: 535788
Conc: 549.23 ng/ml




#30 AR-1254-5

R.T.: 8.474 min
Delta R.T.: -0.005 min
Response: 512625
Conc: 818.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
D3-12292021DL

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/03/2022
Supervised By :mohammad ahmed 01/03/2022



#30 AR-1254-5

R.T.: 7.453 min
Delta R.T.: -0.014 min
Response: 884114
Conc: 689.95 ng/ml