

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP123022\
 Data File : PP054182.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Dec 2022 12:40
 Operator : YP\AJ
 Sample : PB150001BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB150001BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 31 01:18:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 30 05:08:27 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.340	3.589	39616628	29891964	22.932	19.867
2) SA Decachlor...	10.102	8.614	30282277	32438809	23.384	22.769
Target Compounds						
3) L1 AR-1016-1	5.511	4.677	27437538	20104988	455.324	417.287
4) L1 AR-1016-2	5.534	4.696	39566484	27614737	453.439	408.704
5) L1 AR-1016-3	5.596	4.873	25170806	15281485	453.539	417.798
6) L1 AR-1016-4	5.695	4.915	20136057	13630072	456.552	418.336
7) L1 AR-1016-5	5.991	5.129	21008053	16884809	430.087	413.655
31) L7 AR-1260-1	7.123	6.167	41996164	32181658	478.800	402.747
32) L7 AR-1260-2	7.381	6.355	36350044	36307960	345.694	397.355
33) L7 AR-1260-3	7.741	6.509	24084176	35052124	321.022	403.210 #
34) L7 AR-1260-4	7.968	6.983	29602721	25999716	370.541	406.816
35) L7 AR-1260-5	8.284	7.225	50171528	54144871	386.270	406.797



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP123022\
Data File : PP054182.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Dec 2022 12:40
Operator : YP\AJ
Sample : PB150001BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB150001BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 31 01:18:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 30 05:08:27 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.340 min
Delta R.T.: 0.000 min
Response: 39616628
Conc: 22.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB150001BS




#1 Tetrachloro-m-xylene

R.T.: 3.589 min
Delta R.T.: 0.000 min
Response: 29891964
Conc: 19.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.102 min
Delta R.T.: 0.001 min
Response: 30282277
Conc: 23.38 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.614 min
Delta R.T.: -0.001 min
Response: 32438809
Conc: 22.77 ng/ml



#3 AR-1016-1


R.T.: 5.511 min
Delta R.T.: 0.000 min
Response: 27437538
Conc: 455.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB150001BS




#3 AR-1016-1

R.T.: 4.677 min
Delta R.T.: 0.000 min
Response: 20104988
Conc: 417.29 ng/ml




#4 AR-1016-2

R.T.: 5.534 min
Delta R.T.: 0.000 min
Response: 39566484
Conc: 453.44 ng/ml



#4 AR-1016-2


R.T.: 4.696 min
Delta R.T.: 0.000 min
Response: 27614737
Conc: 408.70 ng/ml



#5 AR-1016-3


R.T.: 5.596 min
Delta R.T.: 0.000 min
Response: 25170806
Conc: 453.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB150001BS




#5 AR-1016-3

R.T.: 4.873 min
Delta R.T.: 0.000 min
Response: 15281485
Conc: 417.80 ng/ml




#6 AR-1016-4

R.T.: 5.695 min
Delta R.T.: 0.000 min
Response: 20136057
Conc: 456.55 ng/ml



#6 AR-1016-4


R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 13630072
Conc: 418.34 ng/ml



#7 AR-1016-5


R.T.: 5.991 min
Delta R.T.: 0.000 min
Response: 21008053
Conc: 430.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB150001BS




#7 AR-1016-5

R.T.: 5.129 min
Delta R.T.: 0.000 min
Response: 16884809
Conc: 413.66 ng/ml




#31 AR-1260-1

R.T.: 7.123 min
Delta R.T.: 0.000 min
Response: 41996164
Conc: 478.80 ng/ml



#31 AR-1260-1


R.T.: 6.167 min
Delta R.T.: 0.000 min
Response: 32181658
Conc: 402.75 ng/ml



#32 AR-1260-2


R.T.: 7.381 min
Delta R.T.: 0.001 min
Response: 36350044
Conc: 345.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB150001BS




#32 AR-1260-2

R.T.: 6.355 min
Delta R.T.: 0.000 min
Response: 36307960
Conc: 397.35 ng/ml




#33 AR-1260-3

R.T.: 7.741 min
Delta R.T.: 0.000 min
Response: 24084176
Conc: 321.02 ng/ml



#33 AR-1260-3


R.T.: 6.509 min
Delta R.T.: 0.000 min
Response: 35052124
Conc: 403.21 ng/ml



#34 AR-1260-4


R.T.: 7.968 min
Delta R.T.: 0.000 min
Response: 29602721
Conc: 370.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB150001BS




#34 AR-1260-4

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 25999716
Conc: 406.82 ng/ml



#35 AR-1260-5

R.T.: 8.284 min
Delta R.T.: 0.000 min
Response: 50171528
Conc: 386.27 ng/ml



#35 AR-1260-5

R.T.: 7.225 min
Delta R.T.: 0.000 min
Response: 54144871
Conc: 406.80 ng/ml