

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012819\
 Data File : PP028995.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Jan 2019 10:41
 Operator : EI\MA
 Sample : HSTDICC1000
 Misc : hexane
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 11:16:13 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012819.M
 Quant Title : 8080.M
 QLast Update : Mon Jan 28 11:14:57 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.053	12.175	250.0E6	402.7E6	971.818	966.060
Target Compounds							
1) T	Dalapon	3.313	2.994	506.6E6	506.4E6	883.581	879.061
2) T	3,5-DICHL...	10.443	10.258	526.7E6	500.3E6	896.769	894.340
3) T	4-Nitroph...	11.609	11.320	559.2E6	213.9E6	900.070	881.309
5) T	DICAMBA	12.368	12.484	1248.6E6	1450.2E6	919.211	906.691
6) T	MCPD	12.714	12.676	39001323	92629826	88.361	92.494
7) T	MCPA	12.960	13.070	52137868	140.9E6	95.090	91.330
8) T	DICHLORPROP	13.555	13.593	277.9E6	351.9E6	910.016	910.451
9) T	2,4-D	13.909	14.084	338.4E6	474.5E6	935.325	907.222
10) T	Pentachlo...	14.279	14.670	6306.8E6	5619.3E6	910.932	891.200
11) T	2,4,5-TP ...	15.172	15.255	2389.5E6	2389.7E6	928.993	910.913
12) T	2,4,5-T	15.566	15.915	2267.2E6	2044.2E6	917.825	909.315
13) T	2,4-DB	16.340	16.505	173.9E6	256.9E6	1009.983	986.987
14) T	DINOSEB	17.777	16.885	1959.9E6	1627.9E6	861.176	901.371
15) T	Picloram	17.609	18.294	3524.2E6	3393.8E6	1015.005	945.993
16) T	DCPA	18.170	18.154	3331.8E6	2850.8E6	935.313	916.823

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012819\
Data File : PP028995.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Jan 2019 10:41
Operator : EI\MA
Sample : HSTDICC1000
Misc : hexane
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 28 11:16:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012819.M
Quant Title : 8080.M
QLast Update : Mon Jan 28 11:14:57 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

