

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012819\
 Data File : PP029011.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Jan 2019 17:29
 Operator : EI\MA
 Sample : HSTDCCC750
 Misc : hexane
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 1/29/2019 11:28:42 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 04:49:35 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012819.M
 Quant Title : 8080.M
 QLast Update : Mon Jan 28 11:24:17 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.041	12.171	190.4E6	310.8E6	752.171	754.657
Target Compounds							
1) T	Dalapon	3.308	2.993	389.9E6	382.9E6	682.216	662.753
2) T	3,5-DICHL...	10.434	10.255	404.5E6	386.3E6	694.268	696.465
3) T	4-Nitroph...	11.599	11.316	429.3E6	159.1E6	714.742	665.266
5) T	DICAMBA	12.357	12.480	967.9E6	1127.1E6	729.644	723.477
6) T	MCPD	12.699	12.671	26098115	67099871	62.787	70.167
7) T	MCPA	12.944	13.063	50371648	102.1E6	90.066m	68.777
8) T	DICHLORPROP	13.544	13.590	216.7E6	275.5E6	723.450	720.305
9) T	2,4-D	13.898	14.081	256.6E6	366.5E6	748.716	714.455
10) T	Pentachlo...	14.274	14.669	4998.8E6	4502.9E6	742.854	729.023
11) T	2,4,5-TP ...	15.162	15.251	1874.2E6	1880.5E6	757.819	738.346
12) T	2,4,5-T	15.556	15.910	1776.7E6	1616.5E6	740.232	740.525
13) T	2,4-DB	16.329	16.502	117.6E6	195.2E6	737.287	776.585
14) T	DINOSEB	17.768	16.882	1519.4E6	1273.6E6	663.745	723.379
15) T	Picloram	17.598	18.292	2694.7E6	2679.5E6	824.349	778.836
16) T	DCPA	18.160	18.151	2590.2E6	2251.2E6	751.231	743.404

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012819\
Data File : PP029011.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Jan 2019 17:29
Operator : EI\MA
Sample : HSTDCCC750
Misc : hexane
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
1/29/2019 11:28:42 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 04:49:35 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012819.M
Quant Title : 8080.M
QLast Update : Mon Jan 28 11:24:17 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

