

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP012819.M
 Title : 8080.M
 Last Update : Mon Jan 28 11:24:17 2019
 Response Via : Initial Calibration

Calibration Files

750 =PP028994.D 1000 =PP028995.D 1500 =PP028996.D
 500 =PP028993.D 200 =PP028992.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.733	5.567	5.435	6.038	5.801	5.715	E5	4.04
2) T	3,5-DICHLOROBENZOI...	5.873	5.664	5.493	6.163	5.936	5.826	E5	4.42
3) T	4-Nitrophenol	6.213	6.145	6.135	6.203	5.337	6.006	E5	6.26
4) S	2,4-DCAA	2.572	2.500	2.433	2.524	2.631	2.532	E5	2.96
5) T	DICAMBA	1.358	1.328	1.305	1.313	1.328	1.327	E6	1.54
6) T	MCPP	4.414	4.149	4.521	4.279	3.421	4.157	E5	10.46
7) T	MCPA	5.483	5.606	5.696	5.763	5.415	5.593	E5	2.59
8) T	DICHLORPROP	3.054	2.957	2.874	2.996	3.094	2.995	E5	2.86
9) T	2,4-D	3.618	3.600	3.575	3.394	2.951	3.428	E5	8.20
10) T	Pentachlorophenol	6.923	6.639	6.276	6.869	6.939	6.729	E6	4.17
11) T	2,4,5-TP (SILVEX)	2.572	2.515	2.439	2.475	2.364	2.473	E6	3.18
12) T	2,4,5-T	2.470	2.387	2.306	2.438	2.400	2.400	E6	2.58
13) T	2,4-DB	1.722	1.831	1.832	1.443	1.147	1.595	E5	18.60
14) T	DINOSEB	2.276	2.085	1.972	2.365	2.748	2.289	E6	13.09
15) T	Picloram	3.472	3.710	3.776	3.026	2.361	3.269	E6	17.94
16) T	DCPA	3.562	3.471	3.353	3.460	3.394	3.448	E6	2.32

Signal #2 Calibration Files

750 =PP028994.D 1000 =PP028995.D 1500 =PP028996.D
 500 =PP028993.D 200 =PP028992.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.761	5.565	5.390	6.146	6.025	5.778	E5	5.43
2) T	3,5-DICHLOROBENZOI...	5.594	5.379	5.179	5.642	5.936	5.546	E5	5.15
3) T	4-Nitrophenol	2.427	2.351	2.280	2.557	2.347	2.392	E5	4.42
4) S	2,4-DCAA	4.168	4.027	3.897	4.115	4.383	4.118	E5	4.37
5) T	DICAMBA	1.599	1.543	1.486	1.571	1.591	1.558	E6	2.95
6) T	MCPP	1.001	0.985	0.965	0.935	0.895	0.956	E6	4.44
7) T	MCPA	1.543	1.516	1.495	1.457	1.414	1.485	E6	3.41
8) T	DICHLORPROP	3.865	3.743	3.635	3.815	4.062	3.824	E5	4.15
9) T	2,4-D	5.231	5.048	4.899	5.150	5.322	5.130	E5	3.20
10) T	Pentachlorophenol	6.305	5.915	5.417	6.460	6.785	6.177	E6	8.54
11) T	2,4,5-TP (SILVEX)	2.623	2.515	2.387	2.607	2.602	2.547	E6	3.88
12) T	2,4,5-T	2.248	2.152	2.062	2.207	2.246	2.183	E6	3.58
13) T	2,4-DB	2.603	2.704	2.650	2.429	2.178	2.513	E5	8.51
14) T	DINOSEB	1.806	1.732	1.645	1.793	1.827	1.761	E6	4.19
15) T	Picloram	3.588	3.572	3.455	3.392	3.195	3.440	E6	4.64
16) T	DCPA	3.109	2.970	2.797	3.132	3.133	3.028	E6	4.82

(#) = Out of Range