

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030823\
 Data File : PP056288.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Mar 2023 22:41
 Operator : YP\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 07 23:16:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 07 22:59:55 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.317	3.559	7105426	6326749	4.558	4.586
2) SA Decachlor...	10.061	8.559	9629421	5501762	4.880	5.066
Target Compounds						
21) L5 AR-1248-1	5.488	4.640	1810986	1296614	50.857	49.940
22) L5 AR-1248-2	5.762	4.877	3064223	1973002	54.884	50.721
23) L5 AR-1248-3	5.968	4.918	3262452	2037558	53.615	49.965
24) L5 AR-1248-4	6.373	5.089	3666991	2339088	53.087	49.949
25) L5 AR-1248-5	6.412	5.482	3617565	2054330	53.242	49.745



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030823\
Data File : PP056288.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Mar 2023 22:41
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 07 23:16:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 07 22:59:55 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.317 min
Delta R.T.: 0.000 min
Response: 7105426
Conc: 4.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.559 min
Delta R.T.: 0.000 min
Response: 6326749
Conc: 4.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.061 min
Delta R.T.: 0.000 min
Response: 9629421
Conc: 4.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.559 min
Delta R.T.: 0.000 min
Response: 5501762
Conc: 5.07 ng/ml



#21 AR-1248-1


R.T.: 5.488 min
Delta R.T.: 0.000 min
Response: 1810986
Conc: 50.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050




#21 AR-1248-1

R.T.: 4.640 min
Delta R.T.: 0.000 min
Response: 1296614
Conc: 49.94 ng/ml




#22 AR-1248-2

R.T.: 5.762 min
Delta R.T.: 0.000 min
Response: 3064223
Conc: 54.88 ng/ml



#22 AR-1248-2


R.T.: 4.877 min
Delta R.T.: 0.000 min
Response: 1973002
Conc: 50.72 ng/ml



#23 AR-1248-3


R.T.: 5.968 min
Delta R.T.: 0.000 min
Response: 3262452
Conc: 53.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050




#23 AR-1248-3

R.T.: 4.918 min
Delta R.T.: 0.000 min
Response: 2037558
Conc: 49.97 ng/ml




#24 AR-1248-4

R.T.: 6.373 min
Delta R.T.: 0.000 min
Response: 3666991
Conc: 53.09 ng/ml



#24 AR-1248-4


R.T.: 5.089 min
Delta R.T.: 0.000 min
Response: 2339088
Conc: 49.95 ng/ml



#25 AR-1248-5

R.T.: 6.412 min
Delta R.T.: 0.000 min
Response: 3617565
Conc: 53.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050



#25 AR-1248-5

R.T.: 5.482 min
Delta R.T.: 0.000 min
Response: 2054330
Conc: 49.75 ng/ml