

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030823\
 Data File : PP056290.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Mar 2023 23:14
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 08 00:33:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 08 00:22:13 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.317	3.559	125.8E6	109.9E6	75.371	75.269
2)	SA Decachlor...	10.059	8.558	148.2E6	80745793	75.027	75.161
Target Compounds							
26)	L6 AR-1254-1	6.346	5.441	54141256	52241690	755.181	747.041
27)	L6 AR-1254-2	6.566	5.590	83745061	44401283	753.736	748.265
28)	L6 AR-1254-3	6.933	5.993	90028892	69563075	751.037	752.964
29)	L6 AR-1254-4	7.220	6.223	66448322	40099566	752.818	750.375
30)	L6 AR-1254-5	7.641	6.641	75930143	58646071	753.458	754.619



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP030823\
Data File : PP056290.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Mar 2023 23:14
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 08 00:33:32 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030823.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 08 00:22:13 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.317 min
Delta R.T.: 0.000 min
Response: 125798957
Conc: 75.37 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.559 min
Delta R.T.: 0.000 min
Response: 109882503
Conc: 75.27 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.059 min
Delta R.T.: 0.000 min
Response: 148225101
Conc: 75.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.558 min
Delta R.T.: 0.000 min
Response: 80745793
Conc: 75.16 ng/ml



#26 AR-1254-1


R.T.: 6.346 min
Delta R.T.: 0.000 min
Response: 54141256
Conc: 755.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#26 AR-1254-1

R.T.: 5.441 min
Delta R.T.: 0.000 min
Response: 52241690
Conc: 747.04 ng/ml




#27 AR-1254-2

R.T.: 6.566 min
Delta R.T.: 0.000 min
Response: 83745061
Conc: 753.74 ng/ml



#27 AR-1254-2


R.T.: 5.590 min
Delta R.T.: 0.000 min
Response: 44401283
Conc: 748.26 ng/ml



#28 AR-1254-3


R.T.: 6.933 min
Delta R.T.: 0.000 min
Response: 90028892
Conc: 751.04 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 5.993 min
Delta R.T.: 0.000 min
Response: 69563075
Conc: 752.96 ng/ml




#29 AR-1254-4

R.T.: 7.220 min
Delta R.T.: 0.000 min
Response: 66448322
Conc: 752.82 ng/ml



#29 AR-1254-4


R.T.: 6.223 min
Delta R.T.: 0.000 min
Response: 40099566
Conc: 750.38 ng/ml



#30 AR-1254-5

R.T.: 7.641 min
Delta R.T.: 0.000 min
Response: 75930143
Conc: 753.46 ng/ml

Instrument :
ECD_P
ClientSampleId :



#30 AR-1254-5

R.T.: 6.641 min
Delta R.T.: 0.000 min
Response: 58646071
Conc: 754.62 ng/ml