

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP040219.M
 Title : 8080.M
 Last Update : Wed Apr 03 07:41:20 2019
 Response Via : Initial Calibration

Calibration Files

750 =PP029681.D 1000 =PP029682.D 1500 =PP029683.D
 500 =PP029680.D 200 =PP029679.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	6.629	6.212	6.165	6.789	7.341	6.627	E4 7.24
2) T	3,5-DICHLOROBENZOI...	7.216	6.771	6.655	7.435	7.992	7.214	E4 7.48
3) T	4-Nitrophenol	7.745	7.348	7.287	7.973	8.881	7.847	E4 8.21
4) S	2,4-DCAA	3.228	3.102	2.971	3.272	3.620	3.239	E4 7.52
5) T	DICAMBA	1.649	1.558	1.563	1.695	1.864	1.666	E5 7.50
6) T	MCPP	3.064	3.136	3.743	1.807	0.456	2.441	E4 53.85
7) T	MCPA	5.100	5.425	6.419	3.900	0.985	4.366	E4 47.94
8) T	DICHLORPROP	4.428	4.205	4.100	4.576	5.191	4.500	E4 9.53
9) T	2,4-D	4.738	4.571	4.504	4.890	5.506	4.842	E4 8.27
10) T	Pentachlorophenol	8.077	7.839	8.052	8.082	8.418	8.094	E5 2.57
11) T	2,4,5-TP (SILVEX)	2.912	2.809	2.844	2.961	3.182	2.942	E5 4.99
12) T	2,4,5-T	2.653	2.567	2.612	2.710	2.964	2.701	E5 5.78
13) T	2,4-DB	2.477	2.410	2.401	2.575	2.877	2.548	E4 7.72
14) T	DINOSEB	1.948	1.890	1.918	1.990	2.113	1.972	E5 4.43
15) T	Picloram	3.174	3.103	3.173	3.236	3.389	3.215	E5 3.36
16) T	DCPA	3.671	3.614	3.660	3.758	4.053	3.751	E5 4.71

Signal #2 Calibration Files

750 =PP029681.D 1000 =PP029682.D 1500 =PP029683.D
 500 =PP029680.D 200 =PP029679.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	5.244	4.935	4.854	5.448	6.094	5.315	E4 9.35
2) T	3,5-DICHLOROBENZOI...	5.158	4.837	4.717	5.301	5.745	5.152	E4 7.90
3) T	4-Nitrophenol	2.348	2.248	2.156	2.377	2.385	2.303	E4 4.27
4) S	2,4-DCAA	4.118	3.866	3.794	4.277	4.445	4.100	E4 6.67
5) T	DICAMBA	1.411	1.323	1.321	1.424	1.587	1.413	E5 7.66
6) T	MCPP	3.911	4.358	5.316	2.733	0.474	3.358	E4 55.41
7) T	MCPA	9.115	9.317	9.939	7.518	4.212	8.020	E4 28.78
8) T	DICHLORPROP	3.949	3.687	3.615	3.999	4.319	3.914	E4 7.15
9) T	2,4-D	4.724	4.426	4.351	4.771	5.092	4.673	E4 6.35
10) T	Pentachlorophenol	5.572	5.315	5.368	5.497	5.753	5.501	E5 3.16
11) T	2,4,5-TP (SILVEX)	2.123	2.008	2.014	2.120	2.276	2.108	E5 5.16
12) T	2,4,5-T	1.778	1.720	1.715	1.817	1.947	1.796	E5 5.28
13) T	2,4-DB	2.572	2.457	2.424	2.613	2.688	2.551	E4 4.29
14) T	DINOSEB	1.513	1.447	1.447	1.517	1.627	1.510	E5 4.87
15) T	Picloram	2.240	2.201	2.240	2.271	2.356	2.262	E5 2.57
16) T	DCPA	2.729	2.622	2.628	2.737	2.930	2.729	E5 4.57

(#) = Out of Range