

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040519\
 Data File : PP029769.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Apr 2019 5:12
 Operator : EI\MA
 Sample : K2245-04
 Misc : hexane
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 06 05:27:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040519.M
 Quant Title : 8080.M
 QLast Update : Fri Apr 05 15:13:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.168	12.359	24939835	29104147	751.462	631.681
Target Compounds							
1) T	Dalapon	3.388	0.000	202751	0	3.195	N.D. #
5) T	DICAMBA	12.477	0.000	496630	0	3.031	N.D. #
8) T	DICHLORPROP	13.631	13.813	1729708	1010916	40.090	22.965 #
9) T	2,4-D	14.042	0.000	3665379	0	84.794	N.D. #
11) T	2,4,5-TP ...	15.357	15.486	6176288	31238	23.108	0.141 #
12) T	2,4,5-T	0.000	16.108	0	5252887	N.D.	30.259 #
13) T	2,4-DB	16.529	0.000	17628057	0	741.622	N.D. #
14) T	DINOSEB	17.968	0.000	3985723	0	25.513	N.D. #
15) T	Picloram	17.756	18.527	4283549	247978	16.045	1.193 #
16) T	DCPA	18.290	0.000	5414783	0	16.514	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040519\
Data File : PP029769.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Apr 2019 5:12
Operator : EI\MA
Sample : K2245-04
Misc : hexane
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 06 05:27:37 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP040519.M
Quant Title : 8080.M
QLast Update : Fri Apr 05 15:13:06 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.388 min
Delta R.T.: 0.018 min
Response: 202751
Conc: 3.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Dalapon

R.T.: 0.000 min
Exp R.T. : 3.076 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 12.168 min
Delta R.T.: 0.000 min
Response: 24939835
Conc: 751.46 ng/ml



#4 2,4-DCAA


R.T.: 12.359 min
Delta R.T.: 0.000 min
Response: 29104147
Conc: 631.68 ng/ml



#5 DICAMBA


R.T.: 12.477 min
Delta R.T.: -0.017 min
Response: 496630
Conc: 3.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 DICAMBA

R.T.: 0.000 min
Exp R.T. : 12.675 min
Response: 0
Conc: N.D.




#8 DICHLORPROP

R.T.: 13.631 min
Delta R.T.: -0.058 min
Response: 1729708
Conc: 40.09 ng/ml



#8 DICHLORPROP


R.T.: 13.813 min
Delta R.T.: 0.024 min
Response: 1010916
Conc: 22.97 ng/ml



#9 2,4-D


R.T.: 14.042 min
Delta R.T.: 0.003 min
Response: 3665379
Conc: 84.79 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 14.278 min
Response: 0
Conc: N.D.




#11 2,4,5-TP (SILVEX)

R.T.: 15.357 min
Delta R.T.: 0.033 min
Response: 6176288
Conc: 23.11 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 15.486 min
Delta R.T.: 0.021 min
Response: 31238
Conc: 0.14 ng/ml



#12 2,4,5-T


R.T.: 0.000 min
Exp R.T.: 15.715 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#12 2,4,5-T

R.T.: 16.108 min
Delta R.T.: -0.014 min
Response: 5252887
Conc: 30.26 ng/ml




#13 2,4-DB

R.T.: 16.529 min
Delta R.T.: 0.053 min
Response: 17628057
Conc: 741.62 ng/ml



#13 2,4-DB


R.T.: 0.000 min
Exp R.T.: 16.715 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 17.968 min
Delta R.T.: 0.012 min
Response: 3985723
Conc: 25.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#14 DINOSEB

R.T.: 0.000 min
Exp R.T. : 17.112 min
Response: 0
Conc: N.D.




#15 Picloram

R.T.: 17.756 min
Delta R.T.: 0.012 min
Response: 4283549
Conc: 16.05 ng/ml



#15 Picloram


R.T.: 18.527 min
Delta R.T.: 0.004 min
Response: 247978
Conc: 1.19 ng/ml



#16 DCPA

R.T.: 18.290 min
Delta R.T.: -0.050 min
Response: 5414783
Conc: 16.51 ng/ml

Instrument :
ECD_P
ClientSampleId :



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 18.393 min
Response: 0
Conc: N.D.