

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP040519.M
 Title : 8080.M
 Last Update : Fri Apr 05 15:13:06 2019
 Response Via : Initial Calibration

Calibration Files

750 =PP029736.D 1000 =PP029737.D 1500 =PP029738.D
 500 =PP029735.D 200 =PP029734.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	6.460	6.100	6.062	6.674	6.434	6.346	E4	4.09
2) T	3,5-DICHLOROBENZOI...	7.498	7.030	6.801	7.802	7.843	7.395	E4	6.28
3) T	4-Nitrophenol	8.034	7.717	7.697	8.399	8.395	8.048	E4	4.29
4) S	2,4-DCAA	3.353	3.133	3.050	3.526	3.532	3.319	E4	6.68
5) T	DICAMBA	1.655	1.567	1.547	1.716	1.707	1.638	E5	4.78
6) T	MCPP	4.729	4.695	4.844	4.176	2.224	4.134	E4	26.56
7) T	MCPA	6.824	6.896	7.187	6.107	3.369	6.076	E4	25.75
8) T	DICHLORPROP	4.286	4.045	3.904	4.545	4.792	4.315	E4	8.38
9) T	2,4-D	4.328	4.162	4.121	4.527	4.476	4.323	E4	4.19
10) T	Pentachlorophenol	7.733	7.478	7.631	7.793	7.289	7.585	E5	2.69
11) T	2,4,5-TP (SILVEX)	2.658	2.579	2.638	2.731	2.757	2.673	E5	2.69
12) T	2,4,5-T	2.353	2.290	2.345	2.419	2.465	2.374	E5	2.87
13) T	2,4-DB	2.319	2.151	2.234	2.471	2.709	2.377	E4	9.27
14) T	DINOSEB	1.574	1.536	1.590	1.583	1.529	1.562	E5	1.79
15) T	Picloram	2.662	2.567	2.649	2.759	2.711	2.670	E5	2.70
16) T	DCPA	3.229	3.167	3.260	3.301	3.438	3.279	E5	3.09

Signal #2 Calibration Files

750 =PP029736.D 1000 =PP029737.D 1500 =PP029738.D
 500 =PP029735.D 200 =PP029734.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.518	5.158	5.148	5.813	5.912	5.510	E4	6.47
2) T	3,5-DICHLOROBENZOI...	5.728	5.279	5.221	6.115	6.170	5.702	E4	7.84
3) T	4-Nitrophenol	2.650	2.590	2.562	2.732	2.673	2.641	E4	2.56
4) S	2,4-DCAA	4.656	4.346	4.320	4.909	4.805	4.607	E4	5.78
5) T	DICAMBA	1.562	1.467	1.479	1.623	1.637	1.554	E5	5.08
6) T	MCPP	5.639	5.890	6.767	4.806	1.505	4.921	E4	41.33
7) T	MCPA	1.123	1.129	1.204	1.052	0.634	1.029	E5	22.05
8) T	DICHLORPROP	4.422	4.153	4.072	4.696	4.667	4.402	E4	6.50
9) T	2,4-D	5.005	4.742	4.698	5.228	5.097	4.954	E4	4.61
10) T	Pentachlorophenol	6.055	5.748	5.886	6.178	5.932	5.960	E5	2.75
11) T	2,4,5-TP (SILVEX)	2.216	2.109	2.141	2.303	2.305	2.215	E5	4.07
12) T	2,4,5-T	1.704	1.638	1.673	1.805	1.860	1.736	E5	5.37
13) T	2,4-DB	2.441	2.377	2.390	2.546	2.470	2.445	E4	2.78
14) T	DINOSEB	1.356	1.304	1.362	1.380	1.321	1.345	E5	2.32
15) T	Picloram	2.079	2.009	2.098	2.149	2.055	2.078	E5	2.49
16) T	DCPA	2.588	2.496	2.589	2.694	2.744	2.622	E5	3.73

(#) = Out of Range