

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP041119.M
 Title : 8080.M
 Last Update : Thu Apr 11 04:26:48 2019
 Response Via : Initial Calibration

Calibration Files

750 =PP029816.D 1000 =PP029817.D 1500 =PP029818.D
 500 =PP029815.D 200 =PP029814.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.498	5.491	5.248	5.200	4.735	5.235	E4	5.93
2) T	3,5-DICHLOROBENZOI...	6.140	5.949	5.632	5.734	5.049	5.701	E4	7.26
3) T	4-Nitrophenol	6.702	6.552	6.320	6.402	5.502	6.296	E4	7.42
4) S	2,4-DCAA	2.582	2.514	2.410	2.371	2.135	2.402	E4	7.13
5) T	DICAMBA	1.306	1.280	1.232	1.236	1.079	1.226	E5	7.17
6) T	MCPP						0.000		-1.00
7) T	MCPA						0.000		-1.00
8) T	DICHLORPROP	3.237	3.171	3.022	3.145	2.785	3.072	E4	5.81
9) T	2,4-D	3.259	3.241	3.145	3.040	2.688	3.075	E4	7.58
10) T	Pentachlorophenol	5.625	5.565	5.550	5.246	4.531	5.303	E5	8.60
11) T	2,4,5-TP (SILVEX)	1.976	1.938	1.893	1.893	1.677	1.876	E5	6.20
12) T	2,4,5-T	1.825	1.792	1.757	1.766	1.541	1.736	E5	6.47
13) T	2,4-DB	1.586	1.583	1.527	1.495	1.262	1.491	E4	8.94
14) T	DINOSEB	1.259	1.249	1.200	1.172	1.048	1.185	E5	7.15
15) T	Picloram	2.247	2.277	2.146	2.171	1.879	2.144	E5	7.33
16) T	DCPA	2.502	2.480	2.374	2.448	2.174	2.396	E5	5.55

Signal #2 Calibration Files

750 =PP029816.D 1000 =PP029817.D 1500 =PP029818.D
 500 =PP029815.D 200 =PP029814.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.497	5.334	5.094	5.390	5.123	5.288	E4	3.29
2) T	3,5-DICHLOROBENZOI...	5.215	5.047	4.873	5.123	4.582	4.968	E4	5.03
3) T	4-Nitrophenol	2.317	2.269	2.221	2.225	2.004	2.207	E4	5.44
4) S	2,4-DCAA	3.977	3.860	3.794	3.897	3.426	3.791	E4	5.65
5) T	DICAMBA	1.304	1.285	1.288	1.256	1.184	1.263	E5	3.77
6) T	MCPP						0.000		-1.00
7) T	MCPA						0.000		-1.00
8) T	DICHLORPROP	3.800	3.692	3.646	3.701	3.182	3.604	E4	6.73
9) T	2,4-D	4.367	4.273	4.257	4.271	3.683	4.170	E4	6.62
10) T	Pentachlorophenol	5.088	5.060	5.177	4.828	4.290	4.889	E5	7.34
11) T	2,4,5-TP (SILVEX)	1.914	1.903	1.945	1.847	1.707	1.863	E5	5.06
12) T	2,4,5-T	1.510	1.500	1.505	1.475	1.325	1.463	E5	5.35
13) T	2,4-DB	2.229	2.207	2.204	2.171	1.816	2.125	E4	8.21
14) T	DINOSEB	1.294	1.281	1.301	1.228	1.148	1.250	E5	5.11
15) T	Picloram	1.926	1.952	1.935	1.812	1.688	1.863	E5	6.03
16) T	DCPA	2.158	2.144	2.208	2.116	1.884	2.102	E5	6.01

(#) = Out of Range