

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042322\
 Data File : PP046539.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2022 07:25
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 21:17:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 19:00:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.865	3.129	149.4E6	214.3E6	61.885	65.339
2) SA Decachlor...	9.948	8.452	82071813	137.0E6	55.774	57.435
Target Compounds						
26) L6 AR-1254-1	6.033	5.109	28516637	111.4E6	490.941	577.516
27) L6 AR-1254-2	6.270	5.269	59587049	96601927	483.489	580.111
28) L6 AR-1254-3	6.670	5.698	58755058	137.6E6	444.041	552.008
29) L6 AR-1254-4	6.982	5.947	46179559	93383765	464.625	544.276
30) L6 AR-1254-5	7.442	6.396	45587688	117.3E6	518.320	537.651



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042322\
Data File : PP046539.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2022 07:25
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 21:17:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 22 19:00:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.865 min
Delta R.T.: -0.002 min
Response: 149386613
Conc: 61.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.129 min
Delta R.T.: 0.000 min
Response: 214337289
Conc: 65.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.948 min
Delta R.T.: -0.007 min
Response: 82071813
Conc: 55.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.452 min
Delta R.T.: -0.004 min
Response: 137001258
Conc: 57.44 ng/ml



#26 AR-1254-1


R.T.: 6.033 min
Delta R.T.: -0.004 min
Response: 28516637
Conc: 490.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.109 min
Delta R.T.: -0.002 min
Response: 111439516
Conc: 577.52 ng/ml




#27 AR-1254-2

R.T.: 6.270 min
Delta R.T.: -0.001 min
Response: 59587049
Conc: 483.49 ng/ml



#27 AR-1254-2


R.T.: 5.269 min
Delta R.T.: -0.003 min
Response: 96601927
Conc: 580.11 ng/ml



#28 AR-1254-3


R.T.: 6.670 min
Delta R.T.: -0.003 min
Response: 58755058
Conc: 444.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.698 min
Delta R.T.: -0.002 min
Response: 137642457
Conc: 552.01 ng/ml




#29 AR-1254-4

R.T.: 6.982 min
Delta R.T.: -0.003 min
Response: 46179559
Conc: 464.62 ng/ml



#29 AR-1254-4


R.T.: 5.947 min
Delta R.T.: -0.003 min
Response: 93383765
Conc: 544.28 ng/ml



#30 AR-1254-5

R.T.: 7.442 min
Delta R.T.: -0.003 min
Response: 45587688
Conc: 518.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.396 min
Delta R.T.: -0.003 min
Response: 117316994
Conc: 537.65 ng/ml