

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042322\
 Data File : PP046568.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2022 19:17
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 25 01:42:24 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 19:00:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.864	3.128	145.2E6	205.5E6	60.156	62.658
2) SA Decachlor...	9.948	8.450	86952084	143.8E6	59.090	60.273
Target Compounds						
21) L5 AR-1248-1	5.110	4.260	23977796	41670491	573.699	606.417
22) L5 AR-1248-2	5.405	4.511	34134403	55799205	577.534	596.833
23) L5 AR-1248-3	5.623	4.553	39937879	58437212	571.755	598.605
24) L5 AR-1248-4	6.061	4.734	43087043	65016173	561.897	552.306
25) L5 AR-1248-5	6.102	5.151	41625075	74923517	535.173	614.829


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042322\
Data File : PP046568.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2022 19:17
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 25 01:42:24 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 22 19:00:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.864 min
Delta R.T.: -0.003 min
Response: 145212823
Conc: 60.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.128 min
Delta R.T.: -0.002 min
Response: 205542763
Conc: 62.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.948 min
Delta R.T.: -0.008 min
Response: 86952084
Conc: 59.09 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.450 min
Delta R.T.: -0.006 min
Response: 143769212
Conc: 60.27 ng/ml



#21 AR-1248-1


R.T.: 5.110 min
Delta R.T.: -0.002 min
Response: 23977796
Conc: 573.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.260 min
Delta R.T.: -0.003 min
Response: 41670491
Conc: 606.42 ng/ml




#22 AR-1248-2

R.T.: 5.405 min
Delta R.T.: -0.004 min
Response: 34134403
Conc: 577.53 ng/ml



#22 AR-1248-2


R.T.: 4.511 min
Delta R.T.: -0.003 min
Response: 55799205
Conc: 596.83 ng/ml



#23 AR-1248-3


R.T.: 5.623 min
Delta R.T.: -0.002 min
Response: 39937879
Conc: 571.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.553 min
Delta R.T.: -0.004 min
Response: 58437212
Conc: 598.60 ng/ml




#24 AR-1248-4

R.T.: 6.061 min
Delta R.T.: -0.004 min
Response: 43087043
Conc: 561.90 ng/ml



#24 AR-1248-4


R.T.: 4.734 min
Delta R.T.: -0.002 min
Response: 65016173
Conc: 552.31 ng/ml



#25 AR-1248-5

R.T.: 6.102 min
Delta R.T.: -0.004 min
Response: 41625075
Conc: 535.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.151 min
Delta R.T.: -0.004 min
Response: 74923517
Conc: 614.83 ng/ml