

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042322\
 Data File : PP046583.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Apr 2022 01:32
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 25 02:40:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 22 19:00:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.865	3.128	146.8E6	209.4E6	60.815	63.832
2) SA Decachlor...	9.948	8.449	88706865	155.9E6	60.283	65.363
Target Compounds						
21) L5 AR-1248-1	5.111	4.260	24438655	42564688	584.726	619.430
22) L5 AR-1248-2	5.406	4.511	34084620	57556430	576.692	615.628
23) L5 AR-1248-3	5.624	4.553	40872439	60441547	585.135	619.136
24) L5 AR-1248-4	6.062	4.735	43983134	66716355	573.583	566.749
25) L5 AR-1248-5	6.103	5.151	42687360	77642271	548.831	637.140



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042322\
Data File : PP046583.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Apr 2022 01:32
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 25 02:40:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 22 19:00:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.865 min
Delta R.T.: -0.002 min
Response: 146802930
Conc: 60.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.128 min
Delta R.T.: -0.001 min
Response: 209391425
Conc: 63.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.948 min
Delta R.T.: -0.007 min
Response: 88706865
Conc: 60.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.449 min
Delta R.T.: -0.007 min
Response: 155910876
Conc: 65.36 ng/ml



#21 AR-1248-1


R.T.: 5.111 min
Delta R.T.: 0.000 min
Response: 24438655
Conc: 584.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.260 min
Delta R.T.: -0.003 min
Response: 42564688
Conc: 619.43 ng/ml




#22 AR-1248-2

R.T.: 5.406 min
Delta R.T.: -0.002 min
Response: 34084620
Conc: 576.69 ng/ml



#22 AR-1248-2


R.T.: 4.511 min
Delta R.T.: -0.003 min
Response: 57556430
Conc: 615.63 ng/ml



#23 AR-1248-3


R.T.: 5.624 min
Delta R.T.: 0.000 min
Response: 40872439
Conc: 585.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.553 min
Delta R.T.: -0.004 min
Response: 60441547
Conc: 619.14 ng/ml




#24 AR-1248-4

R.T.: 6.062 min
Delta R.T.: -0.003 min
Response: 43983134
Conc: 573.58 ng/ml



#24 AR-1248-4


R.T.: 4.735 min
Delta R.T.: -0.002 min
Response: 66716355
Conc: 566.75 ng/ml



#25 AR-1248-5

R.T.: 6.103 min
Delta R.T.: -0.003 min
Response: 42687360
Conc: 548.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.151 min
Delta R.T.: -0.004 min
Response: 77642271
Conc: 637.14 ng/ml