

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050424\
 Data File : PP064760.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2024 10:27
 Operator : YP\AJ
 Sample : P2384-10 10X
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 CABLE-10

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 05 22:41:41 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP043024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 01 05:01:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.377	3.453	2953072	4296428	1.678	1.822
2) SA Decachlor...	10.066	8.351	2208615	3878638	2.482	2.079
Target Compounds						
34) L7 AR-1260-4	8.009f	0.000	398458	0	6.762	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050424\
Data File : PP064760.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2024 10:27
Operator : YP\AJ
Sample : P2384-10 10X
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
CABLE-10

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 05 22:41:41 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP043024.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 01 05:01:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.377 min
Delta R.T.: 0.005 min
Response: 2953072
Conc: 1.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
CABLE-10




#1 Tetrachloro-m-xylene

R.T.: 3.453 min
Delta R.T.: 0.010 min
Response: 4296428
Conc: 1.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.066 min
Delta R.T.: -0.012 min
Response: 2208615
Conc: 2.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.351 min
Delta R.T.: -0.011 min
Response: 3878638
Conc: 2.08 ng/ml



#34 AR-1260-4

R.T.: 8.009 min
Delta R.T.: 0.023 min
Response: 398458
Conc: 6.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
CABLE-10



#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 6.780 min
Response: 0
Conc: N.D.