

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050424\
 Data File : PP064797.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2024 22:35
 Operator : YP\AJ
 Sample : P2385-08
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 FLOOR-03

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/06/2024
 Supervised By :Ankita Jodhani 05/06/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 05 22:58:37 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP043024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 01 05:01:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.377	3.454	30742637	47275213	17.467	20.051
2)	SA Decachlor...	10.065	8.349	24148536	39548224	27.142m	21.201m
Target Compounds							
26)	L6 AR-1254-1	6.392	5.298	3441630	8266723	57.642m	70.108
27)	L6 AR-1254-2	6.610	5.444	6135524	7916459	69.909m	75.083
28)	L6 AR-1254-3	6.975	5.840	7144695	13544288	84.222m	84.368
29)	L6 AR-1254-4	7.261	6.067	4758951	7958293	106.863	110.680
30)	L6 AR-1254-5	7.678	6.478	5680114	13813578	102.998m	105.958

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050424\
Data File : PP064797.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2024 22:35
Operator : YP\AJ
Sample : P2385-08
Misc :
ALS Vial : 58 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
FLOOR-03


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/06/2024
Supervised By :Ankita Jodhani 05/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 05 22:58:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP043024.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 01 05:01:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.377 min
Delta R.T.: 0.005 min
Response: 30742637
Conc: 17.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
FLOOR-03


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/06/2024
Supervised By :Ankita Jodhani 05/06/2024




#1 Tetrachloro-m-xylene

R.T.: 3.454 min
Delta R.T.: 0.011 min
Response: 47275213
Conc: 20.05 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.065 min
Delta R.T.: -0.014 min
Response: 24148536
Conc: 27.14 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.349 min
Delta R.T.: -0.013 min
Response: 39548224
Conc: 21.20 ng/ml m




#26 AR-1254-1

R.T.: 6.392 min
Delta R.T.: -0.006 min
Response: 3441630
Conc: 57.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
FLOOR-03


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/06/2024
Supervised By :Ankita Jodhani 05/06/2024




#26 AR-1254-1

R.T.: 5.298 min
Delta R.T.: -0.006 min
Response: 8266723
Conc: 70.11 ng/ml




#27 AR-1254-2

R.T.: 6.610 min
Delta R.T.: -0.006 min
Response: 6135524
Conc: 69.91 ng/ml m



#27 AR-1254-2

R.T.: 5.444 min
Delta R.T.: -0.006 min
Response: 7916459
Conc: 75.08 ng/ml




#28 AR-1254-3

R.T.: 6.975 min
Delta R.T.: -0.006 min
Response: 7144695
Conc: 84.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
FLOOR-03


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/06/2024
Supervised By :Ankita Jodhani 05/06/2024




#28 AR-1254-3

R.T.: 5.840 min
Delta R.T.: -0.007 min
Response: 13544288
Conc: 84.37 ng/ml




#29 AR-1254-4

R.T.: 7.261 min
Delta R.T.: -0.005 min
Response: 4758951
Conc: 106.86 ng/ml



#29 AR-1254-4

R.T.: 6.067 min
Delta R.T.: -0.007 min
Response: 7958293
Conc: 110.68 ng/ml




#30 AR-1254-5

R.T.: 7.678 min
Delta R.T.: -0.006 min
Response: 5680114
Conc: 103.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
FLOOR-03

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/06/2024
Supervised By :Ankita Jodhani 05/06/2024



#30 AR-1254-5

R.T.: 6.478 min
Delta R.T.: -0.009 min
Response: 13813578
Conc: 105.96 ng/ml