

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050522\
 Data File : PP047068.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 May 2022 21:22
 Operator : YP\AJ
 Sample : N2691-01
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 DEEP-WC-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/09/2022
 Supervised By :mohammad ahmed 05/09/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 01:25:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 05 11:06:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.855	3.122	31152905	52993374	12.905m	16.155m#
2) SA Decachlor...	9.924	8.432	29252926	45802175	19.880	19.202

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050522\
Data File : PP047068.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 May 2022 21:22
Operator : YP\AJ
Sample : N2691-01
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
DEEP-WC-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/09/2022
Supervised By :mohammad ahmed 05/09/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 01:25:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 05 11:06:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.855 min
Delta R.T.: -0.010 min
Response: 31152905
Conc: 12.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
DEEP-WC-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/09/2022
Supervised By :mohammad ahmed 05/09/2022




#1 Tetrachloro-m-xylene

R.T.: 3.122 min
Delta R.T.: -0.003 min
Response: 52993374
Conc: 16.15 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.924 min
Delta R.T.: -0.008 min
Response: 29252926
Conc: 19.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.432 min
Delta R.T.: -0.004 min
Response: 45802175
Conc: 19.20 ng/ml