

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052623\
 Data File : PQ061282.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2023 11:57
 Operator : YP\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 21:25:30 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 25 13:25:08 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.435f	2.781	1404661	134404	0.280	0.047 #
Target Compounds							
3)	L1 AR-1016-1	4.529	3.745	382401	27408	2.155	0.244 #
4)	L1 AR-1016-2	4.529	3.782	382401	118248	1.456	0.726 #
5)	L1 AR-1016-3	4.594	3.944	847730	993507	5.164	11.644 #
6)	L1 AR-1016-4	4.711f	4.005	432832	2158607	3.244	28.929 #
7)	L1 AR-1016-5	4.953	4.189	86692	3677612	0.650	40.144 #
8)	L2 AR-1221-1	3.589	2.960	196617	23230	2.996	0.589 #
9)	L2 AR-1221-2	3.679	3.059	717229	154743	14.623	5.147 #
10)	L2 AR-1221-3	3.745	3.105	636218	100952	4.310	1.097 #
11)	L3 AR-1232-1	3.745	3.105	636218	100952	5.718	1.483 #
12)	L3 AR-1232-2	4.244	3.782	88739	118248	1.359	1.538
13)	L3 AR-1232-3	4.529	3.944	382401	993507	3.128	24.793 #
14)	L3 AR-1232-4	4.711f	4.025	432832	6885377	6.867	195.589 #
15)	L3 AR-1232-5	4.771	4.189	626602	3677612	13.910	88.824 #
16)	L4 AR-1242-1	4.529	3.745	382401	27408	2.709	0.309 #
17)	L4 AR-1242-2	4.529	3.782	382401	118248	1.858	0.938 #
18)	L4 AR-1242-3	4.594	3.944	847730	993507	6.518	14.684 #
19)	L4 AR-1242-4	4.711f	4.025	432832	6885377	3.961	104.768 #
20)	L4 AR-1242-5	0.000	4.500	0	6359323	N.D.	69.378 #
21)	L5 AR-1248-1	4.529f	3.745	382401	27408	3.124	0.354 #
22)	L5 AR-1248-2	4.771	4.005	626602	2158607	3.803	18.707 #
23)	L5 AR-1248-3	4.953	4.025	86692	6885377	0.432	62.251 #
24)	L5 AR-1248-4	5.328f	4.189	76352009	3677612	336.196	26.867 #
25)	L5 AR-1248-5	0.000	4.568	0	1548524	N.D.	11.162 #
26)	L6 AR-1254-1	5.328	4.500	76352009	6359323	379.015	35.140 #
27)	L6 AR-1254-2	5.537	4.660	1462154	1267596	4.655	7.692 #
28)	L6 AR-1254-3	5.879	5.038	10834796	9898949	33.266	38.384
29)	L6 AR-1254-4	0.000	5.270	0	869345	N.D.	5.319 #
30)	L6 AR-1254-5	6.616	5.680	104449	492317	0.379	1.932 #
31)	L7 AR-1260-1	0.000	5.176	0	4312369	N.D.	23.155 #
32)	L7 AR-1260-2	6.338	5.365	30548316	1848408	101.033	8.113 #
33)	L7 AR-1260-3	6.692	5.496	3333184	3608672	17.693	17.001
34)	L7 AR-1260-4	0.000	5.972	0	1247109	N.D.	7.899 #
35)	L7 AR-1260-5	7.247f	6.215	120735	1351157	0.276	3.719 #
36)	L8 AR-1262-1	6.692	5.724	3333184	224819	10.578	0.904 #
37)	L8 AR-1262-2	7.247f	6.215	120735	1351157	0.221	2.962 #
38)	L8 AR-1262-3	7.469	6.506	157930	702046	0.433	3.757 #
39)	L8 AR-1262-4	0.000	6.562	0	2052606	N.D.	5.908 #
40)	L8 AR-1262-5	0.000	7.039	0	270473	N.D.	1.606 #
41)	L9 AR-1268-1	7.469	6.506	157930	702046	0.257	1.319 #
42)	L9 AR-1268-2	0.000	6.562	0	2052606	N.D.	4.246 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052623\
 Data File : PQ061282.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2023 11:57
 Operator : YP\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 21:25:30 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 25 13:25:08 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
44) L9	AR-1268-4	0.000	7.039	0	270473	N.D.	1.461 #
45) L9	AR-1268-5	8.353	7.319	529904	1361953	0.389	1.011 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052623\
Data File : PQ061282.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2023 11:57
Operator : YP\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 26 21:25:30 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052523.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 25 13:25:08 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.435 min
Delta R.T.: 0.031 min
Response: 1404661
Conc: 0.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.781 min
Delta R.T.: 0.021 min
Response: 134404
Conc: 0.05 ng/ml




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.585 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 7.534 min
Response: 0
Conc: N.D.



#3 AR-1016-1


R.T.: 4.529 min
Delta R.T.: 0.025 min
Response: 382401
Conc: 2.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 3.745 min
Delta R.T.: -0.018 min
Response: 27408
Conc: 0.24 ng/ml




#4 AR-1016-2

R.T.: 4.529 min
Delta R.T.: 0.006 min
Response: 382401
Conc: 1.46 ng/ml



#4 AR-1016-2


R.T.: 3.782 min
Delta R.T.: 0.004 min
Response: 118248
Conc: 0.73 ng/ml



#5 AR-1016-3


R.T.: 4.594 min
Delta R.T.: 0.013 min
Response: 847730
Conc: 5.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 3.944 min
Delta R.T.: 0.006 min
Response: 993507
Conc: 11.64 ng/ml




#6 AR-1016-4

R.T.: 4.711 min
Delta R.T.: 0.038 min
Response: 432832
Conc: 3.24 ng/ml



#6 AR-1016-4


R.T.: 4.005 min
Delta R.T.: 0.017 min
Response: 2158607
Conc: 28.93 ng/ml



#7 AR-1016-5


R.T.: 4.953 min
Delta R.T.: -0.007 min
Response: 86692
Conc: 0.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.189 min
Delta R.T.: 0.006 min
Response: 3677612
Conc: 40.14 ng/ml




#8 AR-1221-1

R.T.: 3.589 min
Delta R.T.: -0.016 min
Response: 196617
Conc: 3.00 ng/ml



#8 AR-1221-1


R.T.: 2.960 min
Delta R.T.: -0.003 min
Response: 23230
Conc: 0.59 ng/ml



#9 AR-1221-2


R.T.: 3.679 min
Delta R.T.: -0.005 min
Response: 717229
Conc: 14.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.059 min
Delta R.T.: 0.020 min
Response: 154743
Conc: 5.15 ng/ml




#10 AR-1221-3

R.T.: 3.745 min
Delta R.T.: -0.007 min
Response: 636218
Conc: 4.31 ng/ml



#10 AR-1221-3


R.T.: 3.105 min
Delta R.T.: -0.002 min
Response: 100952
Conc: 1.10 ng/ml



#11 AR-1232-1


R.T.: 3.745 min
Delta R.T.: -0.007 min
Response: 636218
Conc: 5.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.105 min
Delta R.T.: -0.001 min
Response: 100952
Conc: 1.48 ng/ml




#12 AR-1232-2

R.T.: 4.244 min
Delta R.T.: -0.002 min
Response: 88739
Conc: 1.36 ng/ml



#12 AR-1232-2


R.T.: 3.782 min
Delta R.T.: 0.004 min
Response: 118248
Conc: 1.54 ng/ml



#13 AR-1232-3


R.T.: 4.529 min
Delta R.T.: 0.006 min
Response: 382401
Conc: 3.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 3.944 min
Delta R.T.: 0.006 min
Response: 993507
Conc: 24.79 ng/ml




#14 AR-1232-4

R.T.: 4.711 min
Delta R.T.: 0.037 min
Response: 432832
Conc: 6.87 ng/ml



#14 AR-1232-4


R.T.: 4.025 min
Delta R.T.: 0.000 min
Response: 6885377
Conc: 195.59 ng/ml



#15 AR-1232-5


R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 626602
Conc: 13.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.189 min
Delta R.T.: 0.006 min
Response: 3677612
Conc: 88.82 ng/ml




#16 AR-1242-1

R.T.: 4.529 min
Delta R.T.: 0.025 min
Response: 382401
Conc: 2.71 ng/ml



#16 AR-1242-1


R.T.: 3.745 min
Delta R.T.: -0.018 min
Response: 27408
Conc: 0.31 ng/ml



#17 AR-1242-2


R.T.: 4.529 min
Delta R.T.: 0.006 min
Response: 382401
Conc: 1.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 3.782 min
Delta R.T.: 0.004 min
Response: 118248
Conc: 0.94 ng/ml




#18 AR-1242-3

R.T.: 4.594 min
Delta R.T.: 0.013 min
Response: 847730
Conc: 6.52 ng/ml



#18 AR-1242-3


R.T.: 3.944 min
Delta R.T.: 0.006 min
Response: 993507
Conc: 14.68 ng/ml



#19 AR-1242-4


R.T.: 4.711 min
Delta R.T.: 0.038 min
Response: 432832
Conc: 3.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.025 min
Delta R.T.: 0.000 min
Response: 6885377
Conc: 104.77 ng/ml




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T. : 5.388 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 4.500 min
Delta R.T.: -0.019 min
Response: 6359323
Conc: 69.38 ng/ml



#21 AR-1248-1


R.T.: 4.529 min
Delta R.T.: 0.025 min
Response: 382401
Conc: 3.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 3.745 min
Delta R.T.: -0.019 min
Response: 27408
Conc: 0.35 ng/ml




#22 AR-1248-2

R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 626602
Conc: 3.80 ng/ml



#22 AR-1248-2


R.T.: 4.005 min
Delta R.T.: 0.017 min
Response: 2158607
Conc: 18.71 ng/ml



#23 AR-1248-3


R.T.: 4.953 min
Delta R.T.: -0.007 min
Response: 86692
Conc: 0.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.025 min
Delta R.T.: 0.000 min
Response: 6885377
Conc: 62.25 ng/ml




#24 AR-1248-4

R.T.: 5.328 min
Delta R.T.: -0.026 min
Response: 76352009
Conc: 336.20 ng/ml



#24 AR-1248-4


R.T.: 4.189 min
Delta R.T.: 0.006 min
Response: 3677612
Conc: 26.87 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 5.388 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 4.568 min
Delta R.T.: 0.011 min
Response: 1548524
Conc: 11.16 ng/ml




#26 AR-1254-1

R.T.: 5.328 min
Delta R.T.: 0.002 min
Response: 76352009
Conc: 379.02 ng/ml



#26 AR-1254-1


R.T.: 4.500 min
Delta R.T.: -0.019 min
Response: 6359323
Conc: 35.14 ng/ml



#27 AR-1254-2


R.T.: 5.537 min
Delta R.T.: -0.007 min
Response: 1462154
Conc: 4.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 4.660 min
Delta R.T.: -0.006 min
Response: 1267596
Conc: 7.69 ng/ml




#28 AR-1254-3

R.T.: 5.879 min
Delta R.T.: -0.018 min
Response: 10834796
Conc: 33.27 ng/ml



#28 AR-1254-3


R.T.: 5.038 min
Delta R.T.: -0.010 min
Response: 9898949
Conc: 38.38 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 6.182 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 5.270 min
Delta R.T.: -0.004 min
Response: 869345
Conc: 5.32 ng/ml




#30 AR-1254-5

R.T.: 6.616 min
Delta R.T.: 0.021 min
Response: 104449
Conc: 0.38 ng/ml



#30 AR-1254-5


R.T.: 5.680 min
Delta R.T.: 0.000 min
Response: 492317
Conc: 1.93 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T.: 6.062 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.176 min
Delta R.T.: 0.000 min
Response: 4312369
Conc: 23.16 ng/ml




#32 AR-1260-2

R.T.: 6.338 min
Delta R.T.: 0.016 min
Response: 30548316
Conc: 101.03 ng/ml



#32 AR-1260-2


R.T.: 5.365 min
Delta R.T.: -0.002 min
Response: 1848408
Conc: 8.11 ng/ml



#33 AR-1260-3


R.T.: 6.692 min
Delta R.T.: 0.016 min
Response: 3333184
Conc: 17.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 5.496 min
Delta R.T.: -0.012 min
Response: 3608672
Conc: 17.00 ng/ml




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 6.894 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 5.972 min
Delta R.T.: 0.002 min
Response: 1247109
Conc: 7.90 ng/ml



#35 AR-1260-5


R.T.: 7.247 min
Delta R.T.: 0.042 min
Response: 120735
Conc: 0.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 6.215 min
Delta R.T.: -0.002 min
Response: 1351157
Conc: 3.72 ng/ml




#36 AR-1262-1

R.T.: 6.692 min
Delta R.T.: 0.015 min
Response: 3333184
Conc: 10.58 ng/ml



#36 AR-1262-1


R.T.: 5.724 min
Delta R.T.: 0.001 min
Response: 224819
Conc: 0.90 ng/ml



#37 AR-1262-2


R.T.: 7.247 min
Delta R.T.: 0.042 min
Response: 120735
Conc: 0.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.215 min
Delta R.T.: -0.002 min
Response: 1351157
Conc: 2.96 ng/ml




#38 AR-1262-3

R.T.: 7.469 min
Delta R.T.: -0.011 min
Response: 157930
Conc: 0.43 ng/ml



#38 AR-1262-3


R.T.: 6.506 min
Delta R.T.: 0.011 min
Response: 702046
Conc: 3.76 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 7.555 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 6.562 min
Delta R.T.: 0.007 min
Response: 2052606
Conc: 5.91 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T.: 8.059 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 7.039 min
Delta R.T.: -0.011 min
Response: 270473
Conc: 1.61 ng/ml



#41 AR-1268-1


R.T.: 7.469 min
Delta R.T.: -0.008 min
Response: 157930
Conc: 0.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 6.506 min
Delta R.T.: 0.011 min
Response: 702046
Conc: 1.32 ng/ml




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 7.557 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 6.562 min
Delta R.T.: 0.005 min
Response: 2052606
Conc: 4.25 ng/ml



#43 AR-1268-3


R.T.: 7.765 min
Delta R.T.: 0.027 min
Response: -512425
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 6.777 min
Delta R.T.: 0.020 min
Response: 210505
Conc: 0.50 ng/ml




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T. : 8.059 min
Response: 0
Conc: N.D.



#44 AR-1268-4


R.T.: 7.039 min
Delta R.T.: -0.010 min
Response: 270473
Conc: 1.46 ng/ml



#45 AR-1268-5

R.T.: 8.353 min
Delta R.T.: 0.001 min
Response: 529904
Conc: 0.39 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 7.319 min
Delta R.T.: 0.001 min
Response: 1361953
Conc: 1.01 ng/ml