

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060222\
 Data File : PP048343.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2022 20:24
 Operator : YP\AJ
 Sample : N2987-16
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 P027-SS004-0612-01

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/03/2022
 Supervised By :Ankita Jodhani 06/03/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 02:34:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 27 23:53:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.034	3.214	63786098	51077104	27.238	25.917
2) SA Decachlor...	10.384	8.643	36868612	30692200	18.969	18.987
Target Compounds						
31) L7 AR-1260-1	7.092	5.975	2029446	771648	21.743	9.014 #
32) L7 AR-1260-2	7.379	6.181	3365535	2132457	30.107	20.688 #
33) L7 AR-1260-3	7.775	6.348	3337780	1654084	38.967m	16.704 #
34) L7 AR-1260-4	8.027	6.864	3140555	2371283	30.268m	28.484
35) L7 AR-1260-5	8.391	7.128	13383439	7138687	64.856m	35.979 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060222\
Data File : PP048343.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2022 20:24
Operator : YP\AJ
Sample : N2987-16
Misc :
ALS Vial : 42 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
P027-SS004-0612-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/03/2022
Supervised By :Ankita Jodhani 06/03/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 02:34:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 27 23:53:39 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.034 min
Delta R.T.: -0.003 min
Response: 63786098
Conc: 27.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
P027-SS004-0612-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/03/2022
Supervised By :Ankita Jodhani 06/03/2022




#1 Tetrachloro-m-xylene

R.T.: 3.214 min
Delta R.T.: -0.009 min
Response: 51077104
Conc: 25.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.384 min
Delta R.T.: -0.008 min
Response: 36868612
Conc: 18.97 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.643 min
Delta R.T.: -0.017 min
Response: 30692200
Conc: 18.99 ng/ml




#31 AR-1260-1

R.T.: 7.092 min
Delta R.T.: -0.004 min
Response: 2029446
Conc: 21.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
P027-SS004-0612-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/03/2022
Supervised By :Ankita Jodhani 06/03/2022




#31 AR-1260-1

R.T.: 5.975 min
Delta R.T.: -0.010 min
Response: 771648
Conc: 9.01 ng/ml




#32 AR-1260-2

R.T.: 7.379 min
Delta R.T.: -0.005 min
Response: 3365535
Conc: 30.11 ng/ml



#32 AR-1260-2

R.T.: 6.181 min
Delta R.T.: -0.009 min
Response: 2132457
Conc: 20.69 ng/ml




#33 AR-1260-3

R.T.: 7.775 min
Delta R.T.: -0.006 min
Response: 3337780
Conc: 38.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
P027-SS004-0612-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/03/2022
Supervised By :Ankita Jodhani 06/03/2022




#33 AR-1260-3

R.T.: 6.348 min
Delta R.T.: -0.007 min
Response: 1654084
Conc: 16.70 ng/ml




#34 AR-1260-4

R.T.: 8.027 min
Delta R.T.: -0.006 min
Response: 3140555
Conc: 30.27 ng/ml m



#34 AR-1260-4

R.T.: 6.864 min
Delta R.T.: -0.011 min
Response: 2371283
Conc: 28.48 ng/ml




#35 AR-1260-5

R.T.: 8.391 min
Delta R.T.: -0.005 min
Response: 13383439
Conc: 64.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
P027-SS004-0612-01

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/03/2022
Supervised By :Ankita Jodhani 06/03/2022



#35 AR-1260-5

R.T.: 7.128 min
Delta R.T.: -0.011 min
Response: 7138687
Conc: 35.98 ng/ml