

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060222\
 Data File : PP048362.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2022 01:58
 Operator : YP\AJ
 Sample : N2989-14
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 P015-SS008-2430-01

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/03/2022
 Supervised By :Ankita Jodhani 06/03/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 06:19:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 27 23:53:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.034	3.214	53999986	43581457	23.059	22.114
2) SA Decachlor...	10.384	8.642	30126972	25514660	15.500	15.784

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP060222\
Data File : PP048362.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2022 01:58
Operator : YP\AJ
Sample : N2989-14
Misc :
ALS Vial : 59 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
P015-SS008-2430-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/03/2022
Supervised By :Ankita Jodhani 06/03/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 06:19:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 27 23:53:39 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.034 min
Delta R.T.: -0.003 min
Response: 53999986
Conc: 23.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
P015-SS008-2430-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/03/2022
Supervised By :Ankita Jodhani 06/03/2022




#1 Tetrachloro-m-xylene

R.T.: 3.214 min
Delta R.T.: -0.009 min
Response: 43581457
Conc: 22.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.384 min
Delta R.T.: -0.008 min
Response: 30126972
Conc: 15.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.642 min
Delta R.T.: -0.018 min
Response: 25514660
Conc: 15.78 ng/ml