

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070123\
 Data File : PP058684.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jul 2023 13:52
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 07:04:07 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Jul 02 07:01:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.433	3.666	84655099	124.8E6	50.000	50.000
2) SA Decachlor...	10.233	8.718	61419601	98120549	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.468	5.571	28259951	61459170	500.000	500.000
27) L6 AR-1254-2	6.688	5.720	42205618	55059905	500.000	500.000
28) L6 AR-1254-3	7.056	6.126	43212575	86864112	500.000	500.000
29) L6 AR-1254-4	7.341	6.356	29751186	49820855	500.000	500.000
30) L6 AR-1254-5	7.762	6.775	34628991	79436230	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070123\
Data File : PP058684.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jul 2023 13:52
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 07:04:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Jul 02 07:01:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.433 min
Delta R.T.: 0.000 min
Response: 84655099
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.666 min
Delta R.T.: 0.000 min
Response: 124825810
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: 0.000 min
Response: 61419601
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.718 min
Delta R.T.: 0.000 min
Response: 98120549
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 6.468 min
Delta R.T.: 0.000 min
Response: 28259951
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#26 AR-1254-1

R.T.: 5.571 min
Delta R.T.: 0.000 min
Response: 61459170
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 6.688 min
Delta R.T.: 0.000 min
Response: 42205618
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 5.720 min
Delta R.T.: 0.000 min
Response: 55059905
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 7.056 min
Delta R.T.: 0.000 min
Response: 43212575
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 6.126 min
Delta R.T.: 0.000 min
Response: 86864112
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 7.341 min
Delta R.T.: 0.000 min
Response: 29751186
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 6.356 min
Delta R.T.: 0.000 min
Response: 49820855
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 7.762 min
Delta R.T.: 0.000 min
Response: 34628991
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :



#30 AR-1254-5

R.T.: 6.775 min
Delta R.T.: 0.000 min
Response: 79436230
Conc: 500.00 ng/ml